



TOTAL SOLIDS - SM2540B

SUPERVISOR: apatelANALYST: AparnaDate: 11/06/2017DATE & TIME IN: 11/03/2017 17:00OVEN TEMP: 98 °CRun Number: LB91368DATE & TIME OUT: 11/06/2017 08:15OVEN TEMP: 105 °CBalanceID: WC SC-3OvenID: WC Oven-2

Dish #	Lab ID	Client ID	Empty Dish Weight (g)	Final Empty Dish Weight (g)	Sample Vol (ml)	1st Dish+Sample weight after Drying @103-@105°C (g)	2nd Dish+Sample weight after Drying @103-@105°C (g)	Final Dish+Sample weight after Drying @103-@105°C (g)	Weight (g)	Result (mg/L)
1	LB91368BL	LB91368BL	96.1443	96.1443	100	96.1443	96.1443	96.1443	0	0
2	I6329-01	B-EMH-F	95.5029	95.5029	100	95.8801	95.8801	95.8801	0.3772	3772
3	I6329-01DUP	B-EMH-FDUP	83.55	83.55	100	83.9422	83.9422	83.9422	0.3922	3922

A = Sample Volume (ml)

B = Empty Dish Weight (g)

C = Final Empty Dish + Sample weight after drying @105°C(g)

D = Weight (g)

Weight (g) = C - B

Result mg/L = $\frac{D}{A} \times 1000 \times 1000$

WORKLIST(Hardcopy Internal Chain)

LB91368 AS

WorkList Name : TS110317

WorkList ID : 105126

Date : 11/3/2017 3:54:49 PM

LB91368

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
11/08/2017	Water	I6329-01	TS	Cool 4 deg C	DIST01	H53	B-EMH-F	11/03/2017	SM2540 B

Date/Time 11/03/17, 16:45
 Received by: AS
 Relinquished by: cp

Date/Time 11/03/17, 17:10
 Received by: cp
 Relinquished by: AS