

7470A


LB91369

INSTRUMENT ID: CV1

Linear

 μ Abs.:

32781



A= 0.0000e+000

B= 3.0385e-004

C= -2.6518e-003

Rho= 0.9999293

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	-0.003	-0.003	-1	0.000	-1				
0.2	0.200	0.212	0.012	708	0.0 %	708				
2.5	2.500	2.489	-0.011	8201	0.0 %	8201				
5.0	5.000	4.956	-0.044	16321	0.0 %	16321				
7.5	7.500	7.587	0.087	24979	0.0 %	24979				
10.0	10.000	9.958	-0.042	32781	0.0 %	32781				

LB91369 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 06 Nov 2017 09:34:36

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	-1	-	0.0000	7470A	PPB	06 Nov 2017 09:51:35	S	Std
0.2 - 1	50.2	708	-	0.2000	7470A	PPB	06 Nov 2017 09:53:43	S	Std
2.5 - 1	52.5	8201	-	2.5000	7470A	PPB	06 Nov 2017 09:55:52	S	Std
5.0 - 1	55	16321	-	5.0000	7470A	PPB	06 Nov 2017 09:58:00	S	Std
7.5 - 1	57.5	24979	-	7.5000	7470A	PPB	06 Nov 2017 10:04:47	S	Std
10.0 - 1	58	32781	-	10.0000	7470A	PPB	06 Nov 2017 10:09:30	S	Std
ICV46 - 1	ICV46	13375	4.0613	-	7470A	PPB	06 Nov 2017 10:12:03	U	SMPL
ICB46 - 1	ICB46	-164	-0.0525	-	7470A	PPB	06 Nov 2017 10:14:11	U	SMPL
CCV01 - 1	CCV01	16543	5.0239	-	7470A	PPB	06 Nov 2017 10:16:18	U	SMPL
CCB01 - 1	CCB01	-118	-0.0385	-	7470A	PPB	06 Nov 2017 10:18:25	U	SMPL
CRA - 1	CRA	677	0.2031	-	7470A	PPB	06 Nov 2017 10:20:32	U	SMPL
HighStd - 1	HighStd	33101	10.0550	-	7470A	PPB	06 Nov 2017 10:22:39	U	SMPL
ChkStd - 1	ChkStd	24529	7.4504	-	7470A	PPB	06 Nov 2017 10:24:46	U	SMPL
PB103976BL - 1	PBW	-25	-0.0102	-	7470A	PPB	06 Nov 2017 10:29:48	U	SMPL
PB103976BS - 1	LCSW	12706	3.8580	-	7470A	PPB	06 Nov 2017 10:31:55	U	SMPL
PB103976TB - 1	PB103937TB	-91	-0.0303	-	7470A	PPB	06 Nov 2017 10:34:02	U	SMPL
I6289-02 - 1	SU-03-110217	3	-0.0017	-	7470A	PPB	06 Nov 2017 10:36:11	U	SMPL
I6294-02 - 1	N48971	-34	-0.0130	-	7470A	PPB	06 Nov 2017 10:38:19	U	SMPL
I6296-02 - 1	SV2-1854L	-50	-0.0178	-	7470A	PPB	06 Nov 2017 10:40:28	U	SMPL
I6297-02 - 1	CHRT-24646	-12	-0.0063	-	7470A	PPB	06 Nov 2017 10:42:35	U	SMPL
CCV02 - 1	CCV02	16802	5.1026	-	7470A	PPB	06 Nov 2017 10:44:41	U	SMPL
CCB02 - 1	CCB02	-103	-0.0339	-	7470A	PPB	06 Nov 2017 10:46:48	U	SMPL
I6299-02 - 1	OK-02-110217	15	0.0019	-	7470A	PPB	06 Nov 2017 10:48:55	U	SMPL
I6299-02DUP - 1	OK-02-110217DUP	10	0.0004	-	7470A	PPB	06 Nov 2017 10:51:02	U	SMPL
I6299-02MS - 1	OK-02-110217MS	12556	3.8124	-	7470A	PPB	06 Nov 2017 10:53:09	U	SMPL
I6299-02MSD - 1	OK-02-110217MSD	12394	3.7632	-	7470A	PPB	06 Nov 2017 10:55:16	U	SMPL
PB103978BL - 1	PBW	-105	-0.0346	-	7470A	PPB	06 Nov 2017 10:57:24	U	SMPL
PB103978BS - 1	LCSW	12995	3.9458	-	7470A	PPB	06 Nov 2017 10:59:32	U	SMPL
I6294-01 - 1	N48971	21003	6.3790	-	7470A	PPB	06 Nov 2017 11:06:20	U	SMPL
I6294-01DUP - 1	N48971DUP	20679	6.2806	-	7470A	PPB	06 Nov 2017 11:08:28	U	SMPL
I6294-01MS - 1	N48971MS	30983	9.4114	-	7470A	PPB	06 Nov 2017 11:13:12	U	SMPL
I6294-01MSD - 1	N48971MSD	31116	9.4518	-	7470A	PPB	06 Nov 2017 11:15:18	U	SMPL
CCV03 - 1	CCV03	16328	4.9586	-	7470A	PPB	06 Nov 2017 11:17:27	U	SMPL
CCB03 - 1	CCB03	-136	-0.0440	-	7470A	PPB	06 Nov 2017 11:19:35	U	SMPL
I6296-01 - 1	SV2-1854L	40	0.0095	-	7470A	PPB	06 Nov 2017 11:21:43	U	SMPL
I6299-02LX5 - 1		-30	-0.0118	-	7470A	PPB	06 Nov 2017 11:23:50	U	SMPL
I6299-02A - 1		13598	4.1291	-	7470A	PPB	06 Nov 2017 11:25:57	U	SMPL
I6294-01LX5 - 1		3758	1.1392	-	7470A	PPB	06 Nov 2017 11:28:04	U	SMPL
I6294-01A - 1		33036	10.0352	-	7470A	PPB	06 Nov 2017 11:30:11	U	SMPL
CCV04 - 1	CCV04	16720	5.0777	-	7470A	PPB	06 Nov 2017 11:32:19	U	SMPL
CCB04 - 1	CCB04	-171	-0.0546	-	7470A	PPB	06 Nov 2017 11:34:29	U	SMPL