



Analytical Summary Report

Analysis Method: 9045C

Analyst By : Jignesh

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB91394

Temp.: 23.1 °C

BalanceID: WC SC-4

SLOPE.: 98.4

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2270
buffer solution pH 7 yellow	W2159
PH 10.01 BUFFER,COLOR CD 475ML	W2154
buffer solution pH 7 yellow	W2219
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.01 Control Limits[+/- 0.1].

True Value of CCV2 = 12.01 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			4.01	11/06/2017	18:30
2	CAL2	1	Water			7.00	11/06/2017	18:31
3	CAL3	1	Water			10.03	11/06/2017	18:32
4	ICV	1	Water			7.00	11/06/2017	18:34
5	CCV1	1	Water			2.01	11/06/2017	18:36
6	I6298-02	1	Solid	20.02	20	9.90	11/06/2017	18:37
7	I6300-02	1	Solid	20.03	20	8.22	11/06/2017	18:39
8	I6311-02	1	Solid	20.04	20	8.52	11/06/2017	18:40
9	I6359-04	1	Solid	20.05	20	6.21	11/06/2017	18:43
10	I6359-04DUP	1	Solid	20.03	20	6.22	11/06/2017	18:45
11	CCV2	1	Water			12.01	11/06/2017	18:46

2391394

WORKLIST(Hardcopy Internal Chain)

WorkList Name : CORROSIVITY I6359

WorkList ID : 105190

Date : 11/6/2017 6:10:13 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
11/09/2017	Solid	I6300-02	Corrosivity	Cool 4 deg C	PSEG05	H52	OR-02-110617	11/06/2017	9045C
11/09/2017	Solid	I6311-02	Corrosivity	Cool 4 deg C	PSEG05	H52	TR-04-110617	11/06/2017	9045C
11/09/2017	Solid	I6359-04	Corrosivity	Cool 4 deg C	PSEG01	H52	TP-1	11/06/2017	9045C

Date/Time 11-06-17 6:10 PM
 Received by: JP
 Relinquished by: CP

Date/Time 11-06-17 7:15 PM
 Received by: CP
 Relinquished by: JP