



Analytical Summary Report

Analysis Method: 9041A

Analyst By : Jignesh

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB91395

Temp.: 23.1 °C

BalanceID: WC SC-4

SLOPE.: 98.4

Calibration Standards	Chemtech Log#
3.0-5.5 pH paper refill	W1934
6.0-8.0 pH paper refill	W2106
Test Paper, PH Short Range 9.0/10.0	W1939
6.0-8.0 pH paper refill	W2106
0.0-3.0 pH paper refill	W1931
10.0-12.0 pH paper refill	W1940

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.01 Control Limits[+/- 0.1].

True Value of CCV2 = 12.01 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			4.0	11/06/2017	18:30
2	CAL2	1	Water			7.0	11/06/2017	18:31
3	CAL3	1	Water			10.5	11/06/2017	18:32
4	ICV	1	Water			7.0	11/06/2017	18:34
5	CCV1	1	Water			2.0	11/06/2017	18:36
6	I6298-03	1	Solid	20.02	20	9.8	11/06/2017	18:37
7	CCV2	1	Water			12.0	11/06/2017	18:38