



Analytical Summary Report

Analysis Method: 9040C

Analyst By : Jignesh

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB91396

Temp.: 23.1 °C

BalanceID: WC SC-4

SLOPE.: 98.4

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2270
buffer solution pH 7 yellow	W2159
PH 10.01 BUFFER,COLOR CD 475ML	W2154
buffer solution pH 7 yellow	W2219
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.01 Control Limits[+/- 0.1].

True Value of CCV2 = 12.01 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			4.01	11/06/2017	18:30
2	CAL2	1	Water			7.00	11/06/2017	18:31
3	CAL3	1	Water			10.03	11/06/2017	18:32
4	ICV	1	Water			7.00	11/06/2017	18:34
5	CCV1	1	Water			2.01	11/06/2017	18:36
6	I6298-04	1	Solid	20.02	20	9.90	11/06/2017	18:37
7	CCV2	1	Water			12.01	11/06/2017	18:38