



Analytical Summary Report

Analysis Method: 9040C

Analyst By : JIGNESH

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB91397

Temp.: 19.2 °C

SLOPE.: 98.1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2270
buffer solution pH 7 yellow	W2159
PH 10.01 BUFFER,COLOR CD 475ML	W2154
buffer solution pH 7 yellow	W2219
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.01 Control Limits[+/- 0.1].

True Value of CCV2 = 12.01 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			4.01	11/06/2017	18:15
2	CAL2	1	Water			6.99	11/06/2017	18:16
3	CAL3	1	Water			10.02	11/06/2017	18:17
4	ICV	1	Water			7.00	11/06/2017	18:19
5	CCV1	1	Water			2.01	11/06/2017	18:20
6	I6351-02	1	Water			7.75	11/06/2017	18:21
7	I6351-02DUP	1	Water			7.76	11/06/2017	18:22
8	I6358-02	1	Water			6.62	11/06/2017	18:24
9	CCV2	1	Water			12.02	11/06/2017	18:25

2391397

WORKLIST(Hardcopy Internal Chain)

WorkList Name : CORROSIONITY I6351 WorkList ID : 105191 Date : 11/6/2017 6:10:55 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
11/07/2017	Water	I6351-02-A	Corrosivity	Cool 4 deg C	PSEG01	H52	DRUMS-1-7	11/06/2017	9040C
11/06/2017	Water	I6358-02-X	Corrosivity	Cool 4 deg C	PSEG01	H52	OILY-WATER-FROM-BERM	11/06/2017	9040C

Date/Time 11-06-17 6:10 PM
Received by: JP
Relinquished by: CP

Date/Time 11-06-17 7:10 PM
Received by: JP
Relinquished by: JP