



Analytical Summary Report

Analysis Method: SM 4500-PH B

Analyst By : JIGNESH

Parameter: pH

Supervisor Review By : apatel

Run Number: LB91398

Temp.: 19.2 °C

SLOPE.: 98.1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2270
buffer solution pH 7 yellow	W2159
PH 10.01 BUFFER,COLOR CD 475ML	W2154
buffer solution pH 7 yellow	W2219
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.01 Control Limits[+/- 0.1].

True Value of CCV2 = 12.02 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			4.01	11/06/2017	18:15
2	CAL2	1	Water			6.99	11/06/2017	18:16
3	CAL3	1	Water			10.02	11/06/2017	18:17
4	ICV	1	Water			7.00	11/06/2017	18:19
5	CCV1	1	Water			2.01	11/06/2017	18:20
6	I6351-01	1	Water			7.75	11/06/2017	18:21
7	I6351-01DUP	1	Water			7.76	11/06/2017	18:22
8	I6358-01	1	Water			6.62	11/06/2017	18:24
9	CCV2	1	Water			12.02	11/06/2017	18:25

2301398

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH I6351

WorkList ID : 105192

Date : 11/6/2017 6:11:32 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
11/06/2017	Water	I6351-01-17	pH	Cool 4 deg C	PSEG01	H52	DRUMS-1-7	11/06/2017	SM 4500-PH B
11/06/2017	Water	I6358-01-17	pH	Cool 4 deg C	PSEG01	H52	OILY-WATER-FROM-BERM	11/06/2017	SM 4500-PH B

Date/Time 11-06-17 6:11 PM
Received by: JP
Relinquished by: CWP

Date/Time 11-06-17 7:30 AM
Received by:
Relinquished by: JD