


7470A

LB 91405
INSTRUMENT ID: CV2

Linear

 μ Abs.:

34053



A= 0.0000e+000

B= 2.9379e-004

C= 3.5873e-002

Rho= 0.9999132

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.044	0.044	28	0.000	28				
0.2	0.200	0.205	0.005	574	0.0 %	574				
2.5	2.500	2.501	0.001	8392	0.0 %	8392				
5.0	5.000	4.898	-0.102	16549	0.0 %	16549				
7.5	7.500	7.512	0.012	25448	0.0 %	25448				
10.0	10.000	10.040	0.040	34053	0.0 %	34053				

LB91405 INSTRUMENT ID : CV1

Method: 7470A		Operator: Admin		Date of Analysis: 07 Nov 2017 10:04:46				
Sample ID	Extended ID	μ Abs.	Conc.	Std ConcMethod	Units	Date	Type	Type
0.0 - 1	50	28	-	0.00007470A	PPB	07 Nov 2017 10:10:59	S	Std
0.2 - 1	50.2	574	-	0.20007470A	PPB	07 Nov 2017 10:13:07	S	Std
2.5 - 1	52.5	8392	-	2.50007470A	PPB	07 Nov 2017 10:15:16	S	Std
5.0 - 1	55	16549	-	5.00007470A	PPB	07 Nov 2017 10:17:24	S	Std
7.5 - 1	57.5	25448	-	7.50007470A	PPB	07 Nov 2017 10:19:32	S	Std
10.0 - 1	50	34053	-	10.00007470A	PPB	07 Nov 2017 10:21:40	S	Std
ICV48 - 1	ICV48	12810	3.7993	-7470A	PPB	07 Nov 2017 10:24:59	U	SMPL
ICB48 - 1	ICB48	-90	0.0094	-7470A	PPB	07 Nov 2017 10:27:06	U	SMPL
CCV09 - 1	CCV09	16972	5.0220	-7470A	PPB	07 Nov 2017 10:29:13	U	SMPL
CCB09 - 1	CCB09	-114	0.0024	-7470A	PPB	07 Nov 2017 10:31:19	U	SMPL
CRA - 1	CRA	574	0.2045	-7470A	PPB	07 Nov 2017 10:33:28	U	SMPL
HighStd - 1	HighStd	34103	10.0549	-7470A	PPB	07 Nov 2017 10:35:34	U	SMPL
ChkStd - 1	ChkStd	23796	7.0268	-7470A	PPB	07 Nov 2017 10:37:41	U	SMPL
PB104024BL - 1	PBW	-165	-0.0126	-7470A	PPB	07 Nov 2017 10:39:50	U	SMPL
PB104024BS - 1	LCSW	13203	3.9147	-7470A	PPB	07 Nov 2017 10:41:59	U	SMPL
PB104024TB - 1	PB103979TB	152	0.0805	-7470A	PPB	07 Nov 2017 10:44:06	U	SMPL
I6305-02 - 1	PL-01-110317	246	0.1081	-7470A	PPB	07 Nov 2017 10:46:15	U	SMPL
I6334-02 - 1	278	252	0.1099	-7470A	PPB	07 Nov 2017 10:48:23	U	SMPL
I6337-02 - 1	WOODLAKE-COMP-5:1	272	0.1158	-7470A	PPB	07 Nov 2017 10:50:32	U	SMPL
I6338-01 - 1	2015	273	0.1161	-7470A	PPB	07 Nov 2017 10:52:38	U	SMPL
CCV10 - 1	CCV10	17261	5.1069	-7470A	PPB	07 Nov 2017 10:54:45	U	SMPL
CCB10 - 1	CCB10	-109	0.0039	-7470A	PPB	07 Nov 2017 10:56:51	U	SMPL
I6338-03 - 1	TK2021L	234	0.1046	-7470A	PPB	07 Nov 2017 10:58:59	U	SMPL
I6339-01 - 1	DRUM	236	0.1052	-7470A	PPB	07 Nov 2017 11:01:06	U	SMPL
I6341-02 - 1	A6835DW-A6865DW	224	0.1017	-7470A	PPB	07 Nov 2017 11:03:13	U	SMPL
I6341-02DUP - 1	A6835DW-A6865DWDUP	196	0.0935	-7470A	PPB	07 Nov 2017 11:05:20	U	SMPL
I6341-02MS - 1	A6835DW-A6865DWMS	13176	3.9068	-7470A	PPB	07 Nov 2017 11:07:27	U	SMPL
I6341-02MSD - 1	A6835DW-A6865DWMSD	12035	3.5716	-7470A	PPB	07 Nov 2017 11:09:35	U	SMPL
PB104025BL - 1	PBW	165	0.0843	-7470A	PPB	07 Nov 2017 11:11:43	U	SMPL
PB104025BS - 1	LCSW	12794	3.7946	-7470A	PPB	07 Nov 2017 11:13:51	U	SMPL
PB104025TB - 1	PB103977TB	-112	0.0030	-7470A	PPB	07 Nov 2017 11:15:59	U	SMPL
I6333-01 - 1	W-1	39	0.0473	-7470A	PPB	07 Nov 2017 11:18:07	U	SMPL
CCV11 - 1	CCV11	17119	5.0652	-7470A	PPB	07 Nov 2017 11:20:14	U	SMPL
CCB11 - 1	CCB11	-114	0.0024	-7470A	PPB	07 Nov 2017 11:22:21	U	SMPL
I6333-01DUP - 1	W-1DUP	60	0.0535	-7470A	PPB	07 Nov 2017 11:24:29	U	SMPL
I6333-01MS - 1	W-1MS	13869	4.1104	-7470A	PPB	07 Nov 2017 11:26:36	U	SMPL
I6333-01MSD - 1	W-1MSD	13760	4.0784	-7470A	PPB	07 Nov 2017 11:28:43	U	SMPL
PB104026BL - 1	PBW	-78	0.0130	-7470A	PPB	07 Nov 2017 11:30:50	U	SMPL
PB104026BS - 1	LCSW	14412	4.2699	-7470A	PPB	07 Nov 2017 11:32:58	U	SMPL
I6338-02 - 1	TK2021L	418	0.1587	-7470A	PPB	07 Nov 2017 11:35:06	U	SMPL
I6341-01 - 1	A6835DW-A6865DW	246	0.1081	-7470A	PPB	07 Nov 2017 11:37:15	U	SMPL
I6341-01DUP - 1	A6835DW-A6865DWDUP	249	0.1090	-7470A	PPB	07 Nov 2017 11:39:23	U	SMPL
I6341-01MS - 1	A6835DW-A6865DWMS	13482	3.9967	-7470A	PPB	07 Nov 2017 11:41:31	U	SMPL
I6341-01MSD - 1	A6835DW-A6865DWMSD	13255	3.9300	-7470A	PPB	07 Nov 2017 11:46:11	U	SMPL
CCV12 - 1	CCV12	17034	5.0402	-7470A	PPB	07 Nov 2017 11:48:17	U	SMPL
CCB12 - 1	CCB12	-101	0.0062	-7470A	PPB	07 Nov 2017 11:50:25	U	SMPL
I6341-02LX5 - 1		50	0.0506	-7470A	PPB	07 Nov 2017 11:52:33	U	SMPL
I6341-02A - 1		14342	4.2494	-7470A	PPB	07 Nov 2017 11:54:40	U	SMPL
I6333-01LX5 - 1		-67	0.0162	-7470A	PPB	07 Nov 2017 11:56:48	U	SMPL
I6333-01A - 1		14031	4.1580	-7470A	PPB	07 Nov 2017 11:58:55	U	SMPL
I6341-01LX5 - 1		-38	0.0247	-7470A	PPB	07 Nov 2017 12:01:02	U	SMPL
I6341-01A - 1		13679	4.0546	-7470A	PPB	07 Nov 2017 12:03:11	U	SMPL
CCV13 - 1	CCV13	16541	4.8954	-7470A	PPB	07 Nov 2017 12:05:19	U	SMPL
CCB13 - 1	CCB13	-137	-0.0044	-7470A	PPB	07 Nov 2017 12:07:28	U	SMPL