


245.1

LB91414  
INSTRUMENT ID: CYA

Linear

 $\mu$  Abs.:

30758



A= 0.0000e+000

B= 3.2463e-004

C= 2.0691e-003

Rho= 0.9999805

Accept=Accepted

| Std ID | Conc.  | Calc. | Dev.   | Mean  | SD or %RSD | Rep 1 | Rep 2 | Rep 3 | Rep 4 | Rep 5 |
|--------|--------|-------|--------|-------|------------|-------|-------|-------|-------|-------|
| 0.0    | 0.000  | 0.009 | 0.009  | 20    | 0.000      | 20    |       |       |       |       |
| 0.2    | 0.200  | 0.187 | -0.013 | 570   | 0.0 %      | 570   |       |       |       |       |
| 2.5    | 2.500  | 2.477 | -0.023 | 7623  | 0.0 %      | 7623  |       |       |       |       |
| 5.0    | 5.000  | 5.047 | 0.047  | 15540 | 0.0 %      | 15540 |       |       |       |       |
| 7.5    | 7.500  | 7.494 | -0.006 | 23077 | 0.0 %      | 23077 |       |       |       |       |
| 10.0   | 10.000 | 9.987 | -0.013 | 30758 | 0.0 %      | 30758 |       |       |       |       |

## LB91414 INSTRUMENT ID : CV1

Method: 245.1

Operator: Admin

Date of Analysis: 07 Nov 2017 13:58:46

| Sample ID       | Extended ID              | $\mu$ Abs. | Conc.   | Std Conc | Method | Units | Date                 | Type | Type |
|-----------------|--------------------------|------------|---------|----------|--------|-------|----------------------|------|------|
| 0.0 - 1         | SO                       | 20         | -       | 0.0000   | 245.1  | PPB   | 07 Nov 2017 14:01:23 | S    | Std  |
| 0.2 - 1         | SO.2                     | 570        | -       | 0.2000   | 245.1  | PPB   | 07 Nov 2017 14:03:32 | S    | Std  |
| 2.5 - 1         | SO.5                     | 7623       | -       | 2.5000   | 245.1  | PPB   | 07 Nov 2017 14:05:40 | S    | Std  |
| 5.0 - 1         | SO.5                     | 15540      | -       | 5.0000   | 245.1  | PPB   | 07 Nov 2017 14:07:49 | S    | Std  |
| 7.5 - 1         | SO.5                     | 23077      | -       | 7.5000   | 245.1  | PPB   | 07 Nov 2017 14:09:56 | S    | Std  |
| 10.0 - 1        | SO                       | 30758      | -       | 10.0000  | 245.1  | PPB   | 07 Nov 2017 14:12:05 | S    | Std  |
| ICV49 - 1       | ICV49                    | 12148      | 3.9457  | -245.1   |        | PPB   | 07 Nov 2017 14:14:37 | U    | SMPL |
| ICB49 - 1       | ICB49                    | -97        | -0.0294 | -245.1   |        | PPB   | 07 Nov 2017 14:16:43 | U    | SMPL |
| CCV14 - 1       | CCV14                    | 15395      | 4.9998  | -245.1   |        | PPB   | 07 Nov 2017 14:18:51 | U    | SMPL |
| CCB14 - 1       | CCB14                    | -90        | -0.0271 | -245.1   |        | PPB   | 07 Nov 2017 14:20:57 | U    | SMPL |
| CRA - 1         | CRA                      | 543        | 0.1783  | -245.1   |        | PPB   | 07 Nov 2017 14:23:05 | U    | SMPL |
| HighStd - 1     | HighStd                  | 30865      | 10.0219 | -245.1   |        | PPB   | 07 Nov 2017 14:25:12 | U    | SMPL |
| ChkStd - 1      | ChkStd                   | 21984      | 7.1388  | -245.1   |        | PPB   | 07 Nov 2017 14:27:20 | U    | SMPL |
| PB104027BL - 1  | PBW                      | -133       | -0.0411 | -245.1   |        | PPB   | 07 Nov 2017 14:29:29 | U    | SMPL |
| PB104027BS - 1  | LCSW                     | 13368      | 4.3418  | -245.1   |        | PPB   | 07 Nov 2017 14:31:39 | U    | SMPL |
| I6327-01 - 1    | 110217-TIDO-F-D-B        | -65        | -0.0190 | -245.1   |        | PPB   | 07 Nov 2017 14:33:46 | U    | SMPL |
| I6327-02 - 1    | 110217-TIDO-F-D-S        | 36         | 0.0138  | -245.1   |        | PPB   | 07 Nov 2017 14:35:55 | U    | SMPL |
| I6327-03 - 1    | 110217-110217-TIDO-F-U-B | 29         | 0.0115  | -245.1   |        | PPB   | 07 Nov 2017 14:38:03 | U    | SMPL |
| I6327-04 - 1    | 110217-TIDO-F-U-S        | 43         | 0.0160  | -245.1   |        | PPB   | 07 Nov 2017 14:40:12 | U    | SMPL |
| I6327-05 - 1    | 110217-JETT-F-D-B        | 79         | 0.0277  | -245.1   |        | PPB   | 07 Nov 2017 14:42:18 | U    | SMPL |
| CCV15 - 1       | CCV15                    | 15457      | 5.0199  | -245.1   |        | PPB   | 07 Nov 2017 14:44:25 | U    | SMPL |
| CCB15 - 1       | CCB15                    | -92        | -0.0278 | -245.1   |        | PPB   | 07 Nov 2017 14:46:31 | U    | SMPL |
| I6327-06 - 1    | 110217-JETT-F-D-S        | 125        | 0.0426  | -245.1   |        | PPB   | 07 Nov 2017 14:48:39 | U    | SMPL |
| I6327-07 - 1    | 110217-JETT-F-U-B        | 110        | 0.0378  | -245.1   |        | PPB   | 07 Nov 2017 14:50:46 | U    | SMPL |
| I6327-08 - 1    | 110217-JETT-F-U-S        | 86         | 0.0300  | -245.1   |        | PPB   | 07 Nov 2017 14:52:53 | U    | SMPL |
| I6327-09 - 1    | 110217-SIDI-F-D-B        | 37         | 0.0141  | -245.1   |        | PPB   | 07 Nov 2017 14:55:00 | U    | SMPL |
| I6327-10 - 1    | 110217-SIDI-F-D-S        | 16         | 0.0073  | -245.1   |        | PPB   | 07 Nov 2017 14:57:07 | U    | SMPL |
| I6327-11 - 1    | 110217-SIDI-F-U-B        | 47         | 0.0173  | -245.1   |        | PPB   | 07 Nov 2017 14:59:15 | U    | SMPL |
| I6327-12 - 1    | 110217-SIDI-F-U-S        | 44         | 0.0164  | -245.1   |        | PPB   | 07 Nov 2017 15:01:23 | U    | SMPL |
| I6327-12DUP - 1 | 110217-SIDI-F-U-SDUP     | 36         | 0.0138  | -245.1   |        | PPB   | 07 Nov 2017 15:03:31 | U    | SMPL |
| I6327-12MS - 1  | 110217-SIDI-F-U-SMS      | 15087      | 4.8998  | -245.1   |        | PPB   | 07 Nov 2017 15:05:40 | U    | SMPL |
| I6327-12MSD - 1 | 110217-SIDI-F-U-SMSD     | 14992      | 4.8690  | -245.1   |        | PPB   | 07 Nov 2017 15:07:47 | U    | SMPL |
| CCV16 - 1       | CCV16                    | 15735      | 5.1102  | -245.1   |        | PPB   | 07 Nov 2017 15:09:54 | U    | SMPL |
| CCB16 - 1       | CCB16                    | -111       | -0.0340 | -245.1   |        | PPB   | 07 Nov 2017 15:12:00 | U    | SMPL |
| I6327-12LX5 - 1 |                          | 2          | 0.0027  | -245.1   |        | PPB   | 07 Nov 2017 15:14:09 | U    | SMPL |
| I6327-12A - 1   |                          | 15238      | 4.9488  | -245.1   |        | PPB   | 07 Nov 2017 15:16:16 | U    | SMPL |
| CCV17 - 1       | CCV17                    | 15599      | 5.0660  | -245.1   |        | PPB   | 07 Nov 2017 15:18:23 | U    | SMPL |
| CCB17 - 1       | CCB17                    | -117       | -0.0359 | -245.1   |        | PPB   | 07 Nov 2017 15:20:30 | U    | SMPL |