



Analytical Summary Report

Analysis Method: 1010A

Reviewed By: ketankumar

Parameter: Flash Point

Supervisor Review By: apatel

Run Number: LB91422

Ambient Barometric Pressure (mmHg): 760

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2168

Seq	LabID	True Value °F	Dilution	Celsius °C	Result °F	Anal Date	Anal Time
1	ICV	81	1	28.0	82.4	11/07/2017	13:35
2	I6351-02		1	>100.0	>212.0	11/07/2017	14:07
3	I6351-02DUP		1	>100.0	>212.0	11/07/2017	14:40
4	I6358-02		1	>100.0	>212.0	11/07/2017	15:05

Result = (Celsius * 1.8) + 32

LB91422

WORKLIST(Hardcopy Internal Chain)

WorkList Name : FP11717

WorkList ID : 105226

Date : 11/7/2017 12:47:14 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
11/09/2017	Water	I6351-02 C	Flash Point	Cool 4 deg C	PSEG01	H52	DRUMS-1-7	11/03/2017	1010A
11/09/2017	Water	I6358-02 C	Flash Point	Cool 4 deg C	PSEG01	H52	OILY-WATER-FROM-BERM	11/06/2017	1010A

Date/Time 11/7/17 13:00
Received by: KS
Relinquished by: CP

Date/Time 11/7/17 15:10
Received by: CP
Relinquished by: KS