



Analytical Summary Report

Analysis Method: SM5220 D
Parameter: COD
Run Number: LB91425

ANALYST: Indiana
SUPERVISOR REVIEW BY: apatel

Reagent/Standard	Lot/Log #
COD ICV-LCS std, 50ppm	WP59085
COD calibration std. 100 ppm	WP59082
COD calibration std. 10 ppm	WP59081
COD calibration std. 150 ppm	WP59083
COD calibration std. 50 ppm	WP59084
COD calibration std. 0 ppm	WP59087
COD Digestion Vials Low Level 0-150Mg/L	W2217
COD ICV-LCS std, 50ppm	WP59466
COD CCV std, 50ppm	WP59467

Temp In (C): <u>150</u>	Date In: <u>11/07/2017</u>	Time In: <u>10:00</u>
Temp Out (C): <u>150</u>	Date Out: <u>11/07/2017</u>	Time Out: <u>12:00</u>

Intercept: 0.3627 Slope: 1.0071 Regression: 0.9999

Seq	Lab ID	TrueValue (mg/l)	DF	MATRIX	Reading	Anal Date	Anal Time
1	CAL1	0	1	Water	0.000	10/26/2017	11:10
2	CAL2	10	1	Water	11.000	10/26/2017	11:10
3	CAL3	50	1	Water	50.000	10/26/2017	11:11
4	CAL4	100	1	Water	102.000	10/26/2017	11:11
5	CAL5	150	1	Water	151.000	10/26/2017	11:12



Analytical Summary Report

Analysis Method: SM5220 D

ANALYST: Indiana

Parameter: COD

SUPERVISOR REVIEW BY: apatel

Run Number: LB91425

Seq	Lab ID	True Value (mg/l)	Initial Weight (g)	Final Vol (ml)	DF	MATRIX	Reading	Result	AnalDate	AnalTime
1	ICV	50			1	Water	50.000	49.287	10/26/2017	11:12
2	ICB				1	Water	0.000	-0.360	10/26/2017	11:13
3	CCV1	50			1	Water	49.000	48.294	11/07/2017	13:05
4	CCB1				1	Water	0.000	-0.360	11/07/2017	13:05
5	LB91425BL				1	Water	0.000	-0.360	11/07/2017	13:06
6	LB91425BS	50			1	Water	51.000	50.280	11/07/2017	13:06
7	I6338-02				1	Water	50.000	49.287	11/07/2017	13:07
8	I6341-01				1	Water	0.000	-0.360	11/07/2017	13:07
9	I6351-01				1	Water	6.000	5.598	11/07/2017	13:08
10	I6358-01				1	Water	40.000	39.358	11/07/2017	13:08
11	I6358-01DUP				1	Water	39.000	38.365	11/07/2017	13:09
12	I6358-01MS	50			1	Water	89.000	88.012	11/07/2017	13:09
13	I6358-01MSD	50			1	Water	90.000	89.005	11/07/2017	13:10
14	CCV2	50			1	Water	52.000	51.273	11/07/2017	13:10
15	CCB2				1	Water	0.000	-0.360	11/07/2017	13:10

2B91425

WORKLIST(Hardcopy Internal Chain)

WorkList Name : COD110617

WorkList ID : 105141

Date : 11/6/2017 8:22:33 AM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
11/08/2017	Water	I6338-02	COD	Conc H2SO4 4c	PSEG01	H52	TK2021L	11/03/2017	SM5220 D
11/08/2017	Water	I6341-01	COD	Conc H2SO4 4c	PSEG01	H52	A6835DW-A6865DW	11/03/2017	SM5220 D
11/06/2017	Water	I6351-01	COD	Conc H2SO4 4c	PSEG01	H52	DRUMS-1-7	11/03/2017	SM5220 D
11/09/2017	Water	I6358-01	COD	Conc H2SO4 4c	PSEG01	H52	OILY-WATER-FROM-BERM	11/06/2017	SM5220 D

Date/Time 11/07/17 9:30
 Received by: IB
 Relinquished by: IB

Date/Time 11/07/17 11:55
 Received by: IB
 Relinquished by: IB