



Analytical Summary Report

Analysis Method: 9045C

Analyst By : Jignesh

Parameter: pH

Supervisor Review By : apatel

Run Number: LB91427

Temp.: 21.3 °C

BalanceID: WC SC-4

SLOPE.: 98.4

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2270
buffer solution pH 7 yellow	W2159
PH 10.01 BUFFER,COLOR CD 475ML	W2154
buffer solution pH 7 yellow	W2219
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.01 Control Limits[+/- 0.1].

True Value of CCV2 = 12.02 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			4.01	11/07/2017	17:35
2	CAL2	1	Water			7.00	11/07/2017	17:36
3	CAL3	1	Water			10.02	11/07/2017	17:37
4	ICV	1	Water			7.01	11/07/2017	17:38
5	CCV1	1	Water			2.02	11/07/2017	17:40
6	I6374-01	1	Solid	20.02	20	6.35	11/07/2017	17:41
7	I6374-01DUP	1	Solid	20.03	20	6.36	11/07/2017	17:42
8	I6377-01	1	Solid	20.04	20	6.28	11/07/2017	17:43
9	CCV2	1	Water			12.01	11/07/2017	17:45

2891427

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH I6374

WorkList ID : 105230

Date : 11/7/2017 5:30:05 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
11/07/2017	Solid	I6374-01-A	pH	Cool 4 deg C	TULL02	I11	FK-PS-01-020	11/07/2017	9045C
11/07/2017	Solid	I6377-01-A	pH	Cool 4 deg C	TULL02	I12	FK-PS-01-021	11/07/2017	9045C

Date/Time 11-07-17 5:30 PM
Received by: JD
Relinquished by: JD

Date/Time 11-07-17 6:35 PM
Received by: CD
Relinquished by: JP