



Analytical Summary Report

Analysis Method: 9045C

Analyst By : Jignesh

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB91428

Temp.: 22.3 °C

BalanceID: WC SC-4

SLOPE.: 98.6

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2270
buffer solution pH 7 yellow	W2159
PH 10.01 BUFFER,COLOR CD 475ML	W2154
buffer solution pH 7 yellow	W2219
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.01 Control Limits[+/- 0.1].

True Value of CCV2 = 12.01 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			4.01	11/07/2017	18:15
2	CAL2	1	Water			7.00	11/07/2017	18:16
3	CAL3	1	Water			10.03	11/07/2017	18:19
4	ICV	1	Water			6.99	11/07/2017	18:20
5	CCV1	1	Water			2.01	11/07/2017	18:21
6	I6304-10	1	Solid	20.02	20	8.94	11/07/2017	18:22
7	I6304-10DUP	1	Solid	20.03	20	8.95	11/07/2017	18:23
8	I6310-04	1	Solid	20.01	20	8.38	11/07/2017	18:25
9	I6379-03	1	Solid	20.04	20	5.43	11/07/2017	18:26
10	I6379-06	1	Solid	20.02	20	5.69	11/07/2017	18:28
11	I6379-09	1	Solid	20.04	20	5.93	11/07/2017	18:30
12	I6381-02	1	Solid	20.01	20	8.03	11/07/2017	18:31
13	I6381-04	1	Solid	20.03	20	7.96	11/07/2017	18:32
14	CCV2	1	Water			12.02	11/07/2017	18:33

WORKLIST(Hardcopy Internal Chain)

WorkList Name : CORROSIONITY I6381

WorkList ID : 105233

Date : 11/7/2017 5:55:42 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
11/10/2017	Solid	I6304-10	Corrosivity	Cool 4 deg C	PSEG05	H52	CL-02-110717	11/07/2017	9045C
11/10/2017	Solid	I6310-04	Corrosivity	Cool 4 deg C	PSEG05	H52	1-5-RO	11/07/2017	9045C
11/10/2017	Solid	I6379-03	Corrosivity	Cool 4 deg C	PSEG01	H52	TP-1-(2-8)	11/07/2017	9045C
11/10/2017	Solid	I6379-06	Corrosivity	Cool 4 deg C	PSEG01	H52	TP-2-(1-4)	11/07/2017	9045C
11/10/2017	Solid	I6379-09	Corrosivity	Cool 4 deg C	PSEG01	H52	TP-3-(1-4)	11/07/2017	9045C
11/10/2017	Solid	I6381-02	Corrosivity	Cool 4 deg C	PSEG01	H52	RT-4651	11/07/2017	9045C
11/10/2017	Solid	I6381-04	Corrosivity	Cool 4 deg C	PSEG01	H52	RT-1500	11/07/2017	9045C

Date/Time 11-07-17 5:55 PM
 Received by: JP
 Relinquished by: JP

Date/Time 11-07-17 7:20 PM
 Received by: JP
 Relinquished by: JP