

TOTAL SUSPENDED SOLIDS - SM2540D

SUPERVISOR: apatel

ANALYST: Aparna

Date: 11/08/2017

DATE & TIME IN: 11/08/2017 09:45

OVEN TEMP: 103 °C

Run Number: LB91431

DATE & TIME OUT: 11/08/2017 13:45

OVEN TEMP: 105 °C

BalanceID: WC Sc-3

OvenID: WC Oven-2

Dish #	Lab ID	Client ID	Empty Dish Weight (g)	Final Empty Dish Weight (g)	Sample Volume (ml)	1st Empty Dish+Sample weight after 1.5hr drying @103-@105°C (g)	2nd Empty Dish+Sample weight after 1.5hr drying @103-@105°C (g)	Final Empty Dish+Sample weight after 1.5hr drying @103-@105°C (g)	Weight (g)	Result mg/L
1	LB91431BL	LB91431BL	1.4	1.4	100	1.4	1.4	1.4	0	0
2	LB91431BS	LB91431BS	1.4216	1.4216	100	1.475	1.475	1.475	0.0534	534
3	I6369-01	110617-TIDO-F-D-B	1.402	1.402	100	1.4175	1.4175	1.4175	0.0155	155
4	I6369-02	110617-TIDO-F-D-S	1.4187	1.4187	100	1.435	1.435	1.435	0.0163	163
5	I6369-03	110617-TIDO-F-U-B	1.42	1.42	100	1.4347	1.4347	1.4347	0.0147	147
6	I6369-04	110617-TIDO-F-U-S	1.3973	1.3973	100	1.4107	1.4107	1.4107	0.0134	134
7	I6369-05	110617-JETT-F-D-B	1.3996	1.3996	100	1.4189	1.4189	1.4189	0.0193	193
8	I6369-06	110617-JETT-F-D-S	1.3965	1.3965	100	1.4139	1.4139	1.4139	0.0174	174
9	I6369-07	110617-JETT-F-U-B	1.3905	1.3905	100	1.4064	1.4064	1.4064	0.0159	159
10	I6369-08	110617-JETT-F-U-S	1.4029	1.4029	100	1.4154	1.4154	1.4154	0.0125	125
11	I6369-08DUP	110617-JETT-F-U-SDUP	1.3857	1.3857	100	1.3979	1.3979	1.3979	0.0122	122
12	I6369-09	110617-SIDI-F-D-B	1.4168	1.4168	100	1.4214	1.4214	1.4214	0.0046	46
13	I6369-10	110617-SIDI-F-D-S	1.3984	1.3984	100	1.4031	1.4031	1.4031	0.0047	47
14	I6369-11	110617-SIDI-F-U-B	1.3945	1.3945	100	1.414	1.414	1.414	0.0195	195
15	I6369-12	110617-SIDI-F-U-S	1.3954	1.3954	100	1.4087	1.4087	1.4087	0.0133	133

WORKLIST(Hardcopy Internal Chain)

WorkList Name : TSS110817

WorkList ID : 105234

Date : 11/8/2017 7:58:54 AM

LR91431

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
11/13/2017	Water	I6369-01-C	TSS	Cool 4 deg C	HDR103	H52	110617-TIDO-F-D-B	11/06/2017	SM2540 D
11/13/2017	Water	I6369-02-C	TSS	Cool 4 deg C	HDR103	H52	110617-TIDO-F-D-S	11/06/2017	SM2540 D
11/13/2017	Water	I6369-03-C	TSS	Cool 4 deg C	HDR103	H52	110617-TIDO-F-U-B	11/06/2017	SM2540 D
11/13/2017	Water	I6369-04-C	TSS	Cool 4 deg C	HDR103	H52	110617-TIDO-F-U-S	11/06/2017	SM2540 D
11/13/2017	Water	I6369-05-C	TSS	Cool 4 deg C	HDR103	H52	110617-JETT-F-D-B	11/06/2017	SM2540 D
11/13/2017	Water	I6369-06-C	TSS	Cool 4 deg C	HDR103	H52	110617-JETT-F-D-S	11/06/2017	SM2540 D
11/13/2017	Water	I6369-07-C	TSS	Cool 4 deg C	HDR103	H52	110617-JETT-F-U-B	11/06/2017	SM2540 D
11/13/2017	Water	I6369-08-C	TSS	Cool 4 deg C	HDR103	H52	110617-JETT-F-U-S	11/06/2017	SM2540 D
11/13/2017	Water	I6369-09-C	TSS	Cool 4 deg C	HDR103	H52	110617-SIDI-F-D-B	11/06/2017	SM2540 D
11/13/2017	Water	I6369-10-C	TSS	Cool 4 deg C	HDR103	H52	110617-SIDI-F-D-S	11/06/2017	SM2540 D
11/13/2017	Water	I6369-11-C	TSS	Cool 4 deg C	HDR103	H52	110617-SIDI-F-U-B	11/06/2017	SM2540 D
11/13/2017	Water	I6369-12-C	TSS	Cool 4 deg C	HDR103	H52	110617-SIDI-F-U-S	11/06/2017	SM2540 D

Date/Time 11/08/17 9:10
 Received by: AS
 Relinquished by: AS

Date/Time 11/08/17 10:00
 Received by: AS
 Relinquished by: AS