

245.1


LB 91435

INSTRUMENT ID: CV1

Linear

 μ Abs.:

32426



A= 0.0000e+000

B= 3.0967e-004

C= 1.3685e-002

Rho= 0.9999465

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.015	0.015	4	0.000	4				
0.2	0.200	0.230	0.030	698	0.0 %	698				
2.5	2.500	2.468	-0.032	7924	0.0 %	7924				
5.0	5.000	4.992	-0.008	16075	0.0 %	16075				
7.5	7.500	7.441	-0.059	23985	0.0 %	23985				
10.0	10.000	10.055	0.055	32426	0.0 %	32426				

LB91435 INSTRUMENT ID : CV1

Method: 245.1

Operator: Admin

Date of Analysis: 08 Nov 2017 10:44:30

Sample ID	Extended ID	µ Abs.	Conc.	Std ConcMethod	Units	Date	Type	Type
0.0 - 1	50	4	-	0.0000245.1	PPB	08 Nov 2017 10:50:18	S	Std
0.2 - 1	50.2	698	-	0.2000245.1	PPB	08 Nov 2017 10:52:27	S	Std
2.5 - 1	52.5	7924	-	2.5000245.1	PPB	08 Nov 2017 10:54:36	S	Std
5.0 - 1	55	16075	-	5.0000245.1	PPB	08 Nov 2017 10:56:44	S	Std
7.5 - 1	57.5	23985	-	7.5000245.1	PPB	08 Nov 2017 10:58:52	S	Std
10.0 - 1	57.0	32426	-	10.0000245.1	PPB	08 Nov 2017 11:05:52	S	Std
ICV51 - 1	ICV51	12862	3.9967	-245.1	PPB	08 Nov 2017 11:09:14	U	SMPL
ICB51 - 1	ICB51	-119	-0.0232	-245.1	PPB	08 Nov 2017 11:11:20	U	SMPL
CCV22 - 1	CCV22	16108	5.0018	-245.1	PPB	08 Nov 2017 11:13:28	U	SMPL
CCB22 - 1	CCB22	-121	-0.0238	-245.1	PPB	08 Nov 2017 11:15:34	U	SMPL
CRA - 1	CRA	694	0.2286	-245.1	PPB	08 Nov 2017 11:17:42	U	SMPL
HighStd - 1	HighStd	32587	10.1049	-245.1	PPB	08 Nov 2017 11:19:49	U	SMPL
ChkStd - 1	ChkStd	21884	6.7905	-245.1	PPB	08 Nov 2017 11:21:56	U	SMPL
PB104066BL - 1	PBW	-165	-0.0374	-245.1	PPB	08 Nov 2017 11:24:06	U	SMPL
PB104066BS - 1	LCSW	12985	4.0347	-245.1	PPB	08 Nov 2017 11:26:14	U	SMPL
I6355-01 - 1	110317-TIDO-F-D-B	-97	-0.0164	-245.1	PPB	08 Nov 2017 11:28:22	U	SMPL
I6355-02 - 1	110317-TIDO-F-D-S	-14	0.0093	-245.1	PPB	08 Nov 2017 11:30:30	U	SMPL
I6355-03 - 1	110317-TIDO-F-U-B	8	0.0162	-245.1	PPB	08 Nov 2017 11:32:39	U	SMPL
I6355-04 - 1	110317-TIDO-F-U-S	7	0.0159	-245.1	PPB	08 Nov 2017 11:34:47	U	SMPL
I6355-05 - 1	110317-TIFR-F-D-B	26	0.0217	-245.1	PPB	08 Nov 2017 11:36:54	U	SMPL
CCV23 - 1	CCV23	16159	5.0176	-245.1	PPB	08 Nov 2017 11:39:01	U	SMPL
CCB23 - 1	CCB23	-108	-0.0198	-245.1	PPB	08 Nov 2017 11:41:07	U	SMPL
I6355-06 - 1	110317-TIFR-F-D-M	31	0.0233	-245.1	PPB	08 Nov 2017 11:43:15	U	SMPL
I6355-07 - 1	110317-TIFR-F-D-S	59	0.0320	-245.1	PPB	08 Nov 2017 11:45:21	U	SMPL
I6355-08 - 1	110317-TIFR-F-U-B	62	0.0329	-245.1	PPB	08 Nov 2017 11:47:28	U	SMPL
I6355-09 - 1	110317-TIFR-F-U-M	17	0.0189	-245.1	PPB	08 Nov 2017 11:49:35	U	SMPL
I6355-10 - 1	110317-TIFR-F-U-S	21	0.0202	-245.1	PPB	08 Nov 2017 11:51:43	U	SMPL
I6355-11 - 1	110317-SIDI-E-D-B	38	0.0255	-245.1	PPB	08 Nov 2017 11:53:51	U	SMPL
I6355-12 - 1	110317-SIDI-E-D-S	28	0.0224	-245.1	PPB	08 Nov 2017 11:55:59	U	SMPL
I6355-13 - 1	110317-SIDI-E-U-B	27	0.0220	-245.1	PPB	08 Nov 2017 11:58:07	U	SMPL
I6355-14 - 1	110317-SIDI-E-U-S	34	0.0242	-245.1	PPB	08 Nov 2017 12:00:16	U	SMPL
I6355-15 - 1	110317-SIDI-E-D-B	15	0.0183	-245.1	PPB	08 Nov 2017 12:02:23	U	SMPL
CCV24 - 1	CCV24	16425	5.1000	-245.1	PPB	08 Nov 2017 12:04:29	U	SMPL
CCB24 - 1	CCB24	-131	-0.0269	-245.1	PPB	08 Nov 2017 12:06:36	U	SMPL
I6355-16 - 1	110317-SIDI-E-D-S	31	0.0233	-245.1	PPB	08 Nov 2017 12:08:44	U	SMPL
I6355-16DUP - 1	110317-SIDI-E-D-SDUP	7	0.0159	-245.1	PPB	08 Nov 2017 12:10:51	U	SMPL
I6355-16MS - 1	110317-SIDI-E-D-SMS	15450	4.7981	-245.1	PPB	08 Nov 2017 12:12:58	U	SMPL
I6355-16MSD - 1	110317-SIDI-E-D-SMSD	15610	4.8476	-245.1	PPB	08 Nov 2017 12:15:05	U	SMPL
PB104077BL - 1	PBW	-110	-0.0204	-245.1	PPB	08 Nov 2017 12:17:12	U	SMPL
PB104077BS - 1	LCSW	12850	3.9929	-245.1	PPB	08 Nov 2017 12:19:20	U	SMPL
I6369-01 - 1	110617-TIDO-F-D-B	-74	-0.0092	-245.1	PPB	08 Nov 2017 12:21:28	U	SMPL
I6369-02 - 1	110617-TIDO-F-D-S	29	0.0227	-245.1	PPB	08 Nov 2017 12:23:36	U	SMPL
I6369-03 - 1	110617-TIDO-F-U-B	14	0.0180	-245.1	PPB	08 Nov 2017 12:25:45	U	SMPL
I6369-04 - 1	110617-TIDO-F-U-S	7	0.0159	-245.1	PPB	08 Nov 2017 12:27:52	U	SMPL
CCV25 - 1	CCV25	16268	5.0514	-245.1	PPB	08 Nov 2017 12:29:59	U	SMPL
CCB25 - 1	CCB25	-117	-0.0225	-245.1	PPB	08 Nov 2017 12:32:06	U	SMPL
I6369-05 - 1	110617-JETT-F-D-B	48	0.0285	-245.1	PPB	08 Nov 2017 12:34:14	U	SMPL
I6369-06 - 1	110617-JETT-F-D-S	54	0.0304	-245.1	PPB	08 Nov 2017 12:36:21	U	SMPL
I6369-07 - 1	110617-JETT-F-U-B	19	0.0196	-245.1	PPB	08 Nov 2017 12:38:28	U	SMPL
I6369-08 - 1	110617-JETT-F-U-S	27	0.0220	-245.1	PPB	08 Nov 2017 12:40:36	U	SMPL
I6369-09 - 1	110617-SIDI-F-D-B	29	0.0227	-245.1	PPB	08 Nov 2017 12:42:43	U	SMPL
I6369-10 - 1	110617-SIDI-F-D-S	15	0.0183	-245.1	PPB	08 Nov 2017 12:44:51	U	SMPL
I6369-11 - 1	110617-SIDI-F-U-B	24	0.0211	-245.1	PPB	08 Nov 2017 12:46:59	U	SMPL
I6369-12 - 1	110617-SIDI-F-U-S	22	0.0205	-245.1	PPB	08 Nov 2017 12:49:07	U	SMPL
I6369-12DUP - 1	110617-SIDI-F-U-SDUP	45	0.0276	-245.1	PPB	08 Nov 2017 12:51:16	U	SMPL
I6369-12MS - 1	110617-SIDI-F-U-SMS	15192	4.7182	-245.1	PPB	08 Nov 2017 12:53:23	U	SMPL
CCV26 - 1	CCV26	16161	5.0182	-245.1	PPB	08 Nov 2017 12:55:30	U	SMPL
CCB26 - 1	CCB26	-116	-0.0222	-245.1	PPB	08 Nov 2017 12:57:37	U	SMPL
I6369-12MSD - 1	110617-SIDI-F-U-SMSD	15292	4.7491	-245.1	PPB	08 Nov 2017 12:59:45	U	SMPL
I6355-16LX5 - 1		-64	-0.0061	-245.1	PPB	08 Nov 2017 13:01:52	U	SMPL
I6355-16A - 1		15822	4.9133	-245.1	PPB	08 Nov 2017 13:03:59	U	SMPL
I6369-12LX5 - 1		-71	-0.0083	-245.1	PPB	08 Nov 2017 13:06:07	U	SMPL
I6369-12A - 1		15704	4.8767	-245.1	PPB	08 Nov 2017 13:08:14	U	SMPL
CCV27 - 1	CCV27	16276	5.0539	-245.1	PPB	08 Nov 2017 13:10:22	U	SMPL
CCB27 - 1	CCB27	-126	-0.0253	-245.1	PPB	08 Nov 2017 13:12:30	U	SMPL