

Analytical Summary ReportAnalysis Method: 365.3ANALYST: AMANDEEPParameter: Phosphorus-TotalSUPERVISOR REVIEW BY: apatelRun Number: LB91437

Reagent/Standard	Lot/Log #
5N sulfuric acid	WP56414
calibration std. phosphate 1 ppm	WP59489
calibration std. phosphate 0.5 ppm	WP59488
calibration std. phosphate 0.3 ppm	WP59487
calibration std. 0 ppm	WP59484
phosphate ICV std.	WP59491
phosphate CCV std.	WP59490
calibration std. phosphate 0.1 ppm	WP59486
calibration std. phosphate 0.05 ppm	WP59485
Combined reagent	WP59493
11 N sulfuric acid	WP57045
Phenolphthalein indicator	WP57038
Sodium hydroxide, 1N	WP57036

Intercept: -0.0009Slope: 0.6504Regression: 0.999829

Seq	Lab ID	True Value (mg/L)	DF	Initial Volume (mL)	Final Volume (mL)	Absorbance Reading at 880nm	AnalDate	AnalTime
1	CAL1	0.00	1	50	50	0.000	11/08/2017	11:26
2	CAL2	0.05	1	50	50	0.033	11/08/2017	11:26
3	CAL3	0.10	1	50	50	0.065	11/08/2017	11:27
4	CAL4	0.30	1	50	50	0.186	11/08/2017	11:27
5	CAL5	0.50	1	50	50	0.330	11/08/2017	11:28
6	CAL6	1.00	1	50	50	0.649	11/08/2017	11:28



Analytical Summary Report

Analysis Method: 365.3

ANALYST: AMANDEEP

Parameter: Phosphorus-Total

SUPERVISOR REVIEW BY: apatel

Run Number: LB91437

Seq	Lab ID	True Value (mg/l)	DF	Initial Volume (mL)	Final Volume (mL)	Absorbance Reading at 880nm	Result (mg/L)	AnalDate	AnalTime
1	ICV	0.50	1	50	50	0.326	0.503	11/08/2017	11:29
2	ICB		1	50	50	0.000	0.001	11/08/2017	11:29
3	CCV1	0.50	1	50	50	0.331	0.510	11/09/2017	11:30
4	CCB1		1	50	50	0.000	0.001	11/09/2017	11:30
5	PB104094BL		1	50	50	0.000	0.001	11/08/2017	11:31
6	PB104094BS	0.50	1	50	50	0.338	0.521	11/08/2017	11:31
7	I6291-01		1	50	50	0.786	1.210	11/08/2017	11:32
8	I6291-01DUP		1	50	50	0.833	1.282	11/08/2017	11:32
9	I6291-01MS	0.50	1	50	50	0.873	1.344	11/08/2017	11:33
10	I6291-01MSD	0.50	1	50	50	0.890	1.370	11/08/2017	11:33
11	I6291-01		10	50	50	0.211	0.326	11/08/2017	11:50
12	I6291-01DUP		10	50	50	0.229	0.353	11/08/2017	11:50
13	I6291-01MS	0.50	10	50	50	0.239	0.369	11/08/2017	11:51
14	I6291-01MSD	0.50	10	50	50	0.240	0.370	11/08/2017	11:51
15	CCV2	0.50	1	50	50	0.329	0.507	11/08/2017	11:52
16	CCB2		1	50	50	0.000	0.001	11/08/2017	11:52