



Analytical Summary Report

Analysis Method: 1010A

Reviewed By: HIRAL

Parameter: Flash Point

Supervisor Review By: apatel

Run Number: LB91442

Ambient Barometric Pressure (mmHg): 760

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2168

Seq	LabID	True Value °F	Dilution	Celsius °C	Result °F	Anal Date	Anal Time
1	ICV	81	1	27.0	80.6	11/08/2017	09:10
2	I6351-05		1	>100.0	>212.0	11/08/2017	10:15
3	I6351-05DUP		1	>100.0	>212.0	11/08/2017	10:45

Result = (Celsius * 1.8) + 32

LB91442

WORKLIST(Hardcopy Internal Chain)

WorkList Name : FP11817

WorkList ID : 105245

Date : 11/8/2017 8:26:47 AM

Due Date	Matrix	Sample	Test	Flash Point	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
11/06/2017	Solid	I6351-05	✓	Flash Point	Cool 4 deg C	PSEG01	H52	DRUM-7	11/03/2017	1010A

Date/Time 11/8/17 8:30 AM
Received by: KS
Relinquished by: KS

Date/Time 11/8/17 10:50 AM
Received by: CP
Relinquished by: KS