


7470A

LB91444
INSTRUMENT ID: CV1

Linear

 μ Abs.:

35426



A= 0.0000e+000

B= 2.8313e-004

C= -3.6026e-002

Rho= 0.9999634

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	-0.038	-0.038	-6	0.000	-6				
0.2	0.200	0.191	-0.009	801	0.0 %	801				
2.5	2.500	2.561	0.061	9171	0.0 %	9171				
5.0	5.000	5.015	0.015	17842	0.0 %	17842				
7.5	7.500	7.477	-0.023	26536	0.0 %	26536				
10.0	10.000	9.994	-0.006	35426	0.0 %	35426				

LB91444 INSTRUMENT ID : CV1

Method: 7470A		Operator: Admin		Date of Analysis: 08 Nov 2017 13:21:52					
Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	-6	-	0.0000	7470A	PPB	08 Nov 2017 13:22:04	S	Std
0.2 - 1	50.2	801	-	0.2000	7470A	PPB	08 Nov 2017 13:24:12	S	Std
2.5 - 1	52.5	9171	-	2.5000	7470A	PPB	08 Nov 2017 13:26:21	S	Std
5.0 - 1	55	17842	-	5.0000	7470A	PPB	08 Nov 2017 13:28:30	S	Std
7.5 - 1	57.5	26536	-	7.5000	7470A	PPB	08 Nov 2017 13:30:37	S	Std
10.0 - 1	510	35426	-	10.0000	7470A	PPB	08 Nov 2017 13:35:11	S	Std
ICV52 - 1	ICV52	13751	3.8572	-	7470A	PPB	08 Nov 2017 13:37:40	U	SMPL
ICB52 - 1	ICB52	-113	-0.0680	-	7470A	PPB	08 Nov 2017 13:39:46	U	SMPL
CCV28 - 1	CCV28	17644	4.9594	-	7470A	PPB	08 Nov 2017 13:41:54	U	SMPL
CCB28 - 1	CCB28	-117	-0.0692	-	7470A	PPB	08 Nov 2017 13:44:00	U	SMPL
CRA - 1	CRA	769	0.1817	-	7470A	PPB	08 Nov 2017 13:46:08	U	SMPL
HighStd - 1	HighStd	35284	9.9538	-	7470A	PPB	08 Nov 2017 13:48:14	U	SMPL
ChkStd - 1	ChkStd	24241	6.8272	-	7470A	PPB	08 Nov 2017 13:50:21	U	SMPL
PB104064BL - 1	PBW	-180	-0.0870	-	7470A	PPB	08 Nov 2017 13:52:30	U	SMPL
PB104064BS - 1	LCSW	12968	3.6355	-	7470A	PPB	08 Nov 2017 13:54:39	U	SMPL
PB104064TB - 1	PB104033TB	-75	-0.0573	-	7470A	PPB	08 Nov 2017 13:56:47	U	SMPL
I6300-02 - 1	OR-02-110617	9	-0.0335	-	7470A	PPB	08 Nov 2017 13:58:55	U	SMPL
I6311-02 - 1	TR-04-110617	25	-0.0289	-	7470A	PPB	08 Nov 2017 14:01:03	U	SMPL
I6311-02DUP - 1	TR-04-110617DUP	28	-0.0281	-	7470A	PPB	08 Nov 2017 14:03:12	U	SMPL
I6311-02MS - 1	TR-04-110617MS	12839	3.5990	-	7470A	PPB	08 Nov 2017 14:05:18	U	SMPL
CCV29 - 1	CCV29	17779	4.9977	-	7470A	PPB	08 Nov 2017 14:07:25	U	SMPL
CCB29 - 1	CCB29	-121	-0.0703	-	7470A	PPB	08 Nov 2017 14:09:32	U	SMPL
I6311-02MSD - 1	TR-04-110617MSD	12813	3.5917	-	7470A	PPB	08 Nov 2017 14:11:39	U	SMPL
I6351-02 - 1	DRUMS-1-7	-81	-0.0590	-	7470A	PPB	08 Nov 2017 14:13:46	U	SMPL
I6353-02 - 1	ROLLOFF-292	6	-0.0343	-	7470A	PPB	08 Nov 2017 14:15:53	U	SMPL
I6358-02 - 1	OILY-WATER-FROM-BERM	33	-0.0267	-	7470A	PPB	08 Nov 2017 14:18:00	U	SMPL
I6359-04 - 1	TP-1	22	-0.0298	-	7470A	PPB	08 Nov 2017 14:20:07	U	SMPL
PB104065BL - 1	PBW	12	-0.0326	-	7470A	PPB	08 Nov 2017 14:22:15	U	SMPL
PB104065BS - 1	LCSW	12800	3.5880	-	7470A	PPB	08 Nov 2017 14:24:23	U	SMPL
I6351-01 - 1	DRUMS-1-7	458	0.0936	-	7470A	PPB	08 Nov 2017 14:26:31	U	SMPL
I6358-01 - 1	OILY-WATER-FROM-BERM	-4	-0.0372	-	7470A	PPB	08 Nov 2017 14:28:40	U	SMPL
I6358-01DUP - 1	OILY-WATER-FROM-BERMDUP	-18	-0.0411	-	7470A	PPB	08 Nov 2017 14:30:47	U	SMPL
CCV30 - 1	CCV30	17698	4.9747	-	7470A	PPB	08 Nov 2017 14:32:54	U	SMPL
CCB30 - 1	CCB30	-115	-0.0686	-	7470A	PPB	08 Nov 2017 14:35:01	U	SMPL
I6358-01MS - 1	OILY-WATER-FROM-BERMMS	11784	3.3003	-	7470A	PPB	08 Nov 2017 14:37:08	U	SMPL
I6358-01MSD - 1	OILY-WATER-FROM-BERMMSD	12237	3.4286	-	7470A	PPB	08 Nov 2017 14:39:15	U	SMPL
I6311-02LX5 - 1		16	-0.0315	-	7470A	PPB	08 Nov 2017 14:46:33	U	SMPL
I6311-02A - 1		13544	3.7986	-	7470A	PPB	08 Nov 2017 14:48:40	U	SMPL
I6358-01LX5 - 1		-108	-0.0666	-	7470A	PPB	08 Nov 2017 14:50:47	U	SMPL
I6358-01A - 1		13789	3.8680	-	7470A	PPB	08 Nov 2017 14:52:55	U	SMPL
CCV31 - 1	CCV31	17761	4.9926	-	7470A	PPB	08 Nov 2017 14:55:03	U	SMPL
CCB31 - 1	CCB31	-91	-0.0618	-	7470A	PPB	08 Nov 2017 14:57:11	U	SMPL