


7471A

LB91455
INSTRUMENT ID: CV1

Linear

 μ Abs.:

42922



A= 0.0000e+000

B= 2.3516e-004

C= -1.4394e-002

Rho= 0.9998551

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	-0.009	-0.009	21	0.000	21				
0.2	0.200	0.210	0.010	956	0.0 %	956				
2.5	2.500	2.542	0.042	10871	0.0 %	10871				
5.0	5.000	5.002	0.002	21334	0.0 %	21334				
7.5	7.500	7.376	-0.124	31426	0.0 %	31426	0			
10.0	10.000	10.079	0.079	42922	0.0 %	42922				

LB91455 INSTRUMENT ID : CV1

Method: 7471A		Operator: Admin		Date of Analysis: 08 Nov 2017 16:10:11					
Sample ID	Extended ID	µ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	21	-	0.0000	7471A	PPB	08 Nov 2017 16:12:05	S	Std
0.2 - 1	50.2	956	-	0.2000	7471A	PPB	08 Nov 2017 16:14:23	S	Std
2.5 - 1	52.5	10871	-	2.5000	7471A	PPB	08 Nov 2017 16:16:40	S	Std
5.0 - 1	55	21334	-	5.0000	7471A	PPB	08 Nov 2017 16:18:58	S	Std
7.5 - 1	57.5	31426	-	7.5000	7471A	PPB	08 Nov 2017 16:23:46	S	Std
10.0 - 1	50	42922	-	10.0000	7471A	PPB	08 Nov 2017 16:28:45	S	Std
ICV53 - 1	ICV53	15928	3.7312	-	7471A	PPB	08 Nov 2017 16:31:54	U	SMPL
ICB53 - 1	ICB53	-129	-0.0447	-	7471A	PPB	08 Nov 2017 16:34:10	U	SMPL
CCV32 - 1	CCV32	21314	4.9977	-	7471A	PPB	08 Nov 2017 16:36:26	U	SMPL
CCB32 - 1	CCB32	-157	-0.0513	-	7471A	PPB	08 Nov 2017 16:38:42	U	SMPL
CRA - 1	CRA	944	0.2076	-	7471A	PPB	08 Nov 2017 16:40:59	U	SMPL
HighStd - 1	HighStd	43474	10.2088	-	7471A	PPB	08 Nov 2017 16:43:14	U	SMPL
ChkStd - 1	ChkStd	28663	6.7259	-	7471A	PPB	08 Nov 2017 16:45:30	U	SMPL
PB104069BL - 1	PBS	-51	-0.0264	-	7471A	PPB	08 Nov 2017 16:47:48	U	SMPL
PB104069BS - 1	LCSS	17711	4.1505	-	7471A	PPB	08 Nov 2017 16:50:07	U	SMPL
I6304-09 - 1	CL-02-110717	6760	1.5753	-	7471A	PPB	08 Nov 2017 16:52:23	U	SMPL
I6310-01 - 1	1-5-RO	3140	0.7240	-	7471A	PPB	08 Nov 2017 16:54:41	U	SMPL
I6364-01 - 1	N-P001-001-0001-01	3367	0.7774	-	7471A	PPB	08 Nov 2017 16:56:58	U	SMPL
I6364-01DUP - 1	N-P001-001-0001-01DUP	3366	0.7771	-	7471A	PPB	08 Nov 2017 16:59:16	U	SMPL
I6364-02 - 1	I6364-01MS	18436	4.3209	-	7471A	PPB	08 Nov 2017 17:01:32	U	SMPL
CCV33 - 1	CCV33	21052	4.9361	-	7471A	PPB	08 Nov 2017 17:03:48	U	SMPL
CCB33 - 1	CCB33	-168	-0.0539	-	7471A	PPB	08 Nov 2017 17:06:04	U	SMPL
I6364-03 - 1	I6364-01MSD	18894	4.4287	-	7471A	PPB	08 Nov 2017 17:08:21	U	SMPL
I6364-04 - 1	N-P001-001-0001-02	2725	0.6264	-	7471A	PPB	08 Nov 2017 17:10:37	U	SMPL
I6367-01 - 1	P002-5001-0001-01	3068	0.7071	-	7471A	PPB	08 Nov 2017 17:12:53	U	SMPL
I6367-02 - 1	I6367-01MS	19117	4.4811	-	7471A	PPB	08 Nov 2017 17:15:10	U	SMPL
I6367-03 - 1	I6367-01MSD	19054	4.4663	-	7471A	PPB	08 Nov 2017 17:17:26	U	SMPL
I6367-04 - 1	P002-5001-0001-02	2658	0.6107	-	7471A	PPB	08 Nov 2017 17:19:44	U	SMPL
I6373-01 - 1	208	151	0.0211	-	7471A	PPB	08 Nov 2017 17:22:02	U	SMPL
I6374-01 - 1	FK-PS-01-020	392	0.0778	-	7471A	PPB	08 Nov 2017 17:24:20	U	SMPL
I6377-01 - 1	FK-PS-01-021	472	0.0966	-	7471A	PPB	08 Nov 2017 17:26:38	U	SMPL
I6379-01 - 1	TP-1-(2-8)	100192	23.5464	-	7471A	PPB	08 Nov 2017 17:28:54	U	SMPL
CCV34 - 1	CCV34	20904	4.9013	-	7471A	PPB	08 Nov 2017 17:31:10	U	SMPL
CCB34 - 1	CCB34	-203	-0.0621	-	7471A	PPB	08 Nov 2017 17:33:28	U	SMPL
I6379-04 - 1	TP-2-(1-4)	11836	2.7689	-	7471A	PPB	08 Nov 2017 17:35:46	U	SMPL
I6379-07 - 1	TP-3-(1-4)	59565	13.9927	-	7471A	PPB	08 Nov 2017 17:38:03	U	SMPL
I6381-01 - 1	RT-4651	5654	1.3152	-	7471A	PPB	08 Nov 2017 17:40:19	U	SMPL
I6381-03 - 1	RT-1500	2332	0.5340	-	7471A	PPB	08 Nov 2017 17:42:38	U	SMPL
I6383-01 - 1	SB1-GARAGE-FRONT	2241	0.5126	-	7471A	PPB	08 Nov 2017 17:44:54	U	SMPL
I6383-02 - 1	SB2-GARAGE-BACK	2146	0.4903	-	7471A	PPB	08 Nov 2017 17:47:11	U	SMPL
I6383-03 - 1	SB1-SOUTH-PARKING-LOT	1893	0.4308	-	7471A	PPB	08 Nov 2017 17:49:28	U	SMPL
I6383-04 - 1	SB2-NORTH-PARKING-LOT	2184	0.4992	-	7471A	PPB	08 Nov 2017 17:51:46	U	SMPL
I6364-01LX5 - 1		621	0.1316	-	7471A	PPB	08 Nov 2017 17:54:04	U	SMPL
I6364-01A - 1		20034	4.6967	-	7471A	PPB	08 Nov 2017 17:56:20	U	SMPL
CCV35 - 1	CCV35	21426	5.0241	-	7471A	PPB	08 Nov 2017 17:58:36	U	SMPL
CCB35 - 1	CCB35	-155	-0.0508	-	7471A	PPB	08 Nov 2017 18:00:54	U	SMPL
I6379-01DLX5 - 1		19580	4.5900	-	7471A	PPB	08 Nov 2017 18:03:10	U	SMPL
I6379-07DLX5 - 1		11325	2.6488	-	7471A	PPB	08 Nov 2017 18:05:26	U	SMPL
CCV36 - 1	CCV36	21362	5.0090	-	7471A	PPB	08 Nov 2017 18:07:44	U	SMPL
CCB36 - 1	CCB36	-141	-0.0476	-	7471A	PPB	08 Nov 2017 18:10:00	U	SMPL