



Analytical Summary Report

Analysis Method: 9045C

Analyst By : Jignesh

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB91467

Temp.: 22.1 °C

BalanceID: WC SC-4

SLOPE.: 98.3

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2270
buffer solution pH 7 yellow	W2159
PH 10.01 BUFFER,COLOR CD 475ML	W2154
buffer solution pH 7 yellow	W2219
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.01 Control Limits[+/- 0.1].

True Value of CCV2 = 12.01 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			4.01	11/08/2017	16:35
2	CAL2	1	Water			7.01	11/08/2017	16:36
3	CAL3	1	Water			10.04	11/08/2017	16:37
4	ICV	1	Water			6.99	11/08/2017	16:39
5	CCV1	1	Water			2.01	11/08/2017	16:40
6	I6289-04	1	Solid	20.03	20	9.13	11/08/2017	16:41
7	I6289-04DUP	1	Solid	20.05	20	9.14	11/08/2017	16:42
8	I6382-10	1	Solid	20.02	20	7.35	11/08/2017	16:43
9	I6382-20	1	Solid	20.03	20	7.94	11/08/2017	16:45
10	I6384-02	1	Solid	20.01	20	9.21	11/08/2017	16:46
11	CCV2	1	Water			12.02	11/08/2017	16:47

WORKLIST(Hardcopy Internal Chain)

LB91467

WorkList Name : CORROSIONITY I6382

WorkList ID : 105279

Date : 11/8/2017 2:24:57 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
11/12/2017	Solid	I6289-04	Corrosivity	Cool 4 deg C	PSEG05	H52	SU-02-110817	11/08/2017	9045C
11/10/2017	Solid	I6382-10	Corrosivity	Cool 4 deg C	PSEG01	H52	TP-3	11/07/2017	9045C
11/10/2017	Solid	I6382-20	Corrosivity	Cool 4 deg C	PSEG01	H52	TP-2	11/07/2017	9045C
11/10/2017	Solid	I6384-02	Corrosivity	Cool 4 deg C	PSEG01	H52	EAST-RUTHERFORD-ROLLC	11/07/2017	9045C

Date/Time 11-08-17 2:30 PM
 Received by: Jm
 Relinquished by: Jm

Date/Time 11-08-17 6:35 PM
 Received by: Jm
 Relinquished by: Jm