



Analytical Summary Report

Analysis Method: 9040C

Analyst By : JIGNESH

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB91468

Temp.: 18.9 °C

SLOPE.: 97.6

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2270
buffer solution pH 7 yellow	W2159
PH 10.01 BUFFER,COLOR CD 475ML	W2154
buffer solution pH 7 yellow	W2219
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.01 Control Limits[+/- 0.1].

True Value of CCV2 = 12.01 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			4.01	11/09/2017	11:30
2	CAL2	1	Water			6.99	11/09/2017	11:31
3	CAL3	1	Water			10.02	11/09/2017	11:32
4	ICV	1	Water			7.00	11/09/2017	11:33
5	CCV1	1	Water			2.01	11/09/2017	11:35
6	I6387-02	1	Water			7.45	11/09/2017	11:36
7	I6387-02DUP	1	Water			7.46	11/09/2017	11:37
8	CCV2	1	Water			12.01	11/09/2017	11:39

LB91468

WORKLIST(Hardcopy Internal Chain)

WorkList Name : CORROSIVITY I6387 WorkList ID : 105316 Date : 11/9/2017 11:13:18 AM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
11/08/2017	Water	I6387-02-β	Corrosivity	Cool 4 deg C	PSEG01	H52	SEWAGE-INJECTION-TANK-1	11/08/2017	9040C

Date/Time 11-09-17 11:13 AM
Received by: JP
Relinquished by: CP

Date/Time 11-09-17 12:00 PM
Received by: JP
Relinquished by: CP