


7471A

LB 91474
INSTRUMENT ID: CV1

Linear

 μ Abs.:

40994



A= 0.0000e+000

B= 2.4277e-004

C= -2.2301e-003

Rho= 0.9996258

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.007	0.007	39	0.000	39				
0.2	0.200	0.210	0.010	875	0.0 %	875				
2.5	2.500	2.549	0.049	10510	0.0 %	10510				
5.0	5.000	4.824	-0.176	19881	0.0 %	19881				
7.5	7.500	7.659	0.159	31556	0.0 %	31556	0			
10.0	10.000	9.950	-0.050	40994	0.0 %	40994				

LB91474 INSTRUMENT ID : CV1

Method: 7471A		Operator: Admin		Date of Analysis: 09 Nov 2017 14:06:03				
Sample ID	Extended ID	µ Abs.	Conc.	Std ConcMethod	Units	Date	Type	Type
0.0 - 1	50	39	-	0.00007471A	PPB	09 Nov 2017 14:09:01	S	Std
0.2 - 1	50.2	875	-	0.20007471A	PPB	09 Nov 2017 14:11:19	S	Std
2.5 - 1	52.5	10510	-	2.50007471A	PPB	09 Nov 2017 14:13:36	S	Std
5.0 - 1	55	19881	-	5.00007471A	PPB	09 Nov 2017 14:15:54	S	Std
7.5 - 1	57.5	31556	-	7.50007471A	PPB	09 Nov 2017 14:18:11	S	Std
10.0 - 1	510	40994	-	10.00007471A	PPB	09 Nov 2017 14:20:29	S	Std
ICV55 - 1	ICV55	16206	3.9322	-7471A	PPB	09 Nov 2017 14:23:14	U	SMPL
ICB55 - 1	ICB55	-112	-0.0294	-7471A	PPB	09 Nov 2017 14:25:30	U	SMPL
CCV42 - 1	CCV42	20965	5.0875	-7471A	PPB	09 Nov 2017 14:27:46	U	SMPL
CCB42 - 1	CCB42	-143	-0.0369	-7471A	PPB	09 Nov 2017 14:30:02	U	SMPL
CRA - 1	CRA	840	0.2017	-7471A	PPB	09 Nov 2017 14:32:19	U	SMPL
HighStd - 1	HighStd	41254	10.0132	-7471A	PPB	09 Nov 2017 14:34:35	U	SMPL
ChkStd - 1	ChkStd	29526	7.1659	-7471A	PPB	09 Nov 2017 14:36:51	U	SMPL
PB104110BL - 1	PBS	-218	-0.0552	-7471A	PPB	09 Nov 2017 14:39:10	U	SMPL
PB104110BS - 1	LCSS	15145	3.6746	-7471A	PPB	09 Nov 2017 14:41:28	U	SMPL
I6289-03 - 1	SU-02-110817	473	0.1126	-7471A	PPB	09 Nov 2017 14:43:45	U	SMPL
I6368-04 - 1	TEC-RCA-SOIL	2391	0.5782	-7471A	PPB	09 Nov 2017 14:46:02	U	SMPL
I6382-01 - 1	TP-3	868	0.2085	-7471A	PPB	09 Nov 2017 14:48:20	U	SMPL
I6382-11 - 1	TP-2	952	0.2289	-7471A	PPB	09 Nov 2017 14:50:38	U	SMPL
I6384-01 - 1	EAST-RUTHERFORD-ROLLOFFS	8791	2.1320	-7471A	PPB	09 Nov 2017 14:52:54	U	SMPL
CCV43 - 1	CCV43	20841	5.0574	-7471A	PPB	09 Nov 2017 14:55:10	U	SMPL
CCB43 - 1	CCB43	-118	-0.0309	-7471A	PPB	09 Nov 2017 14:57:25	U	SMPL
I6388-01 - 1	3N-1	813	0.1951	-7471A	PPB	09 Nov 2017 14:59:43	U	SMPL
I6388-03 - 1	3N-2	478	0.1138	-7471A	PPB	09 Nov 2017 15:01:58	U	SMPL
I6388-05 - 1	3N-3	311	0.0733	-7471A	PPB	09 Nov 2017 15:04:14	U	SMPL
I6388-07 - 1	3N-4	406	0.0963	-7471A	PPB	09 Nov 2017 15:06:30	U	SMPL
I6388-09 - 1	3N-5	801	0.1922	-7471A	PPB	09 Nov 2017 15:08:47	U	SMPL
I6388-11 - 1	3N-6	333	0.0786	-7471A	PPB	09 Nov 2017 15:11:04	U	SMPL
I6388-13 - 1	3N-7	586	0.1400	-7471A	PPB	09 Nov 2017 15:13:21	U	SMPL
I6388-15 - 1	3N-8	883	0.2121	-7471A	PPB	09 Nov 2017 15:15:38	U	SMPL
I6388-15DUP - 1	3N-8DUP	873	0.2097	-7471A	PPB	09 Nov 2017 15:17:57	U	SMPL
I6388-15MS - 1	3N-8MS	16553	4.0164	-7471A	PPB	09 Nov 2017 15:20:13	U	SMPL
CCV44 - 1	CCV44	20887	5.0686	-7471A	PPB	09 Nov 2017 15:22:30	U	SMPL
CCB44 - 1	CCB44	-128	-0.0333	-7471A	PPB	09 Nov 2017 15:24:46	U	SMPL
I6388-15MSD - 1	3N-8MSD	16482	3.9992	-7471A	PPB	09 Nov 2017 15:27:03	U	SMPL
I6388-17 - 1	3N-9	805	0.1932	-7471A	PPB	09 Nov 2017 15:29:20	U	SMPL
I6388-19 - 1	3N-10	1118	0.2692	-7471A	PPB	09 Nov 2017 15:31:36	U	SMPL
PB104111BL - 1	PBS	13	0.0009	-7471A	PPB	09 Nov 2017 15:33:52	U	SMPL
PB104111BS - 1	LCSS	15585	3.7814	-7471A	PPB	09 Nov 2017 15:36:09	U	SMPL
I6389-01 - 1	3N-11	797	0.1913	-7471A	PPB	09 Nov 2017 15:38:26	U	SMPL
I6389-03 - 1	3N-12	1034	0.2488	-7471A	PPB	09 Nov 2017 15:40:43	U	SMPL
I6389-05 - 1	3N-13	890	0.2138	-7471A	PPB	09 Nov 2017 15:43:01	U	SMPL
I6389-05DUP - 1	3N-13DUP	826	0.1983	-7471A	PPB	09 Nov 2017 15:45:19	U	SMPL
I6389-05MS - 1	3N-13MS	16330	3.9623	-7471A	PPB	09 Nov 2017 15:47:35	U	SMPL
CCV45 - 1	CCV45	20604	4.9999	-7471A	PPB	09 Nov 2017 15:49:51	U	SMPL
CCB45 - 1	CCB45	-107	-0.0282	-7471A	PPB	09 Nov 2017 15:52:08	U	SMPL
I6389-05MSD - 1	3N-13MSD	16590	4.0254	-7471A	PPB	09 Nov 2017 15:54:25	U	SMPL
I6389-07 - 1	3N-14	577	0.1379	-7471A	PPB	09 Nov 2017 15:56:41	U	SMPL
I6389-09 - 1	3N-15	332	0.0784	-7471A	PPB	09 Nov 2017 15:58:58	U	SMPL
I6389-11 - 1	3N-16	493	0.1175	-7471A	PPB	09 Nov 2017 16:01:14	U	SMPL
I6389-13 - 1	3N-17	551	0.1315	-7471A	PPB	09 Nov 2017 16:03:31	U	SMPL
I6389-15 - 1	3N-18	371	0.0878	-7471A	PPB	09 Nov 2017 16:05:48	U	SMPL
I6389-17 - 1	3N-19	298	0.0701	-7471A	PPB	09 Nov 2017 16:08:05	U	SMPL
I6389-19 - 1	3N-20	2227	0.5384	-7471A	PPB	09 Nov 2017 16:10:22	U	SMPL
I6390-01 - 1	3N-21	341	0.0806	-7471A	PPB	09 Nov 2017 16:12:40	U	SMPL
I6390-03 - 1	3N-22	300	0.0706	-7471A	PPB	09 Nov 2017 16:14:56	U	SMPL
CCV46 - 1	CCV46	19873	4.8224	-7471A	PPB	09 Nov 2017 16:17:13	U	SMPL
CCB46 - 1	CCB46	-107	-0.0282	-7471A	PPB	09 Nov 2017 16:19:29	U	SMPL
I6388-15LX5 - 1		182	0.0420	-7471A	PPB	09 Nov 2017 16:21:46	U	SMPL
I6388-15A - 1		17368	4.2143	-7471A	PPB	09 Nov 2017 16:24:03	U	SMPL
I6389-05LX5 - 1		52	0.0104	-7471A	PPB	09 Nov 2017 16:26:19	U	SMPL
I6389-05A - 1		17519	4.2509	-7471A	PPB	09 Nov 2017 16:28:35	U	SMPL
CCV47 - 1	CCV47	20907	5.0735	-7471A	PPB	09 Nov 2017 16:30:52	U	SMPL
CCB47 - 1	CCB47	-113	-0.0297	-7471A	PPB	09 Nov 2017 16:33:09	U	SMPL