


ISM02.4

LB 91487
INSTRUMENT ID: CV1

Linear

 μ Abs.:

43520



A= 0.0000e+000

B= 2.2946e-004

C= -4.3707e-002

Rho= 0.9999138

Accept=Accepted

slope
1/10153epl

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%OD
0.00	0.000	-0.035	-0.035	36	0.000	36					-
0.2	0.200	0.188	-0.012	1010	0.0 %	1010					-6
2.5	2.500	2.497	-0.003	11071	0.0 %	11071					0
5.0	5.000	5.096	0.096	22399	0.0 %	22399					2
7.5	7.500	7.512	0.012	32929	0.0 %	32929					0
10.0	10.000	9.942	-0.058	43520	0.0 %	43520					-1

NB
11/09/12

LB91487
INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0	36	-		0 ISM02.4	PPB	11/9/2017 16:54	Std
	0.2 S0.2	1010	-		0.2 ISM02.4	PPB	11/9/2017 16:56	Std
	2.5 S2.5	11071	-		2.5 ISM02.4	PPB	11/9/2017 16:58	Std
	5 S5	22399	-		5 ISM02.4	PPB	11/9/2017 17:01	Std
	7.5 S7.5	32929	-		7.5 ISM02.4	PPB	11/9/2017 17:03	Std
	10 S10	43520	-		10 ISM02.4	PPB	11/9/2017 17:05	Std
ICV	ICV	16699	3.7881 -		ISM02.4	PPB	11/9/2017 17:08	SMPL
ICB	ICB	-118	-0.0708 -		ISM02.4	PPB	11/9/2017 17:10	SMPL
CCV048	CCV048	22350	5.0848 -		ISM02.4	PPB	11/9/2017 17:13	SMPL
CCB048	CCB048	-136	-0.0749 -		ISM02.4	PPB	11/9/2017 17:15	SMPL
PB104109BL	PBS004	16	-0.04 -		ISM02.4	PPB	11/9/2017 17:17	SMPL
I6246-06	MBDXT9	1160	0.2225 -		ISM02.4	PPB	11/9/2017 17:19	SMPL
I6246-07	MBDXX1	764	0.1316 -		ISM02.4	PPB	11/9/2017 17:22	SMPL
I6246-11	MBDXM1	544	0.0811 -		ISM02.4	PPB	11/9/2017 17:29	SMPL
I6246-15	MBDXM1D	532	0.0784 -		ISM02.4	PPB	11/9/2017 17:31	SMPL
I6246-16	MBDXM1S	12131	2.7399 -		ISM02.4	PPB	11/9/2017 17:33	SMPL
CCV049	CCV049	22232	5.0577 -		ISM02.4	PPB	11/9/2017 17:36	SMPL
CCB049	CCB049	-128	-0.0731 -		ISM02.4	PPB	11/9/2017 17:38	SMPL