



Analytical Summary Report

Analysis Method: SM 4500-PH B

Analyst By : JIGNESH

Parameter: pH

Supervisor Review By : apatel

Run Number: LB91488

Temp.: 21.1 °C

SLOPE.: 98.2

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2270
buffer solution pH 7 yellow	W2159
PH 10.01 BUFFER,COLOR CD 475ML	W2154
buffer solution pH 7 yellow	W2219
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.01 Control Limits[+/- 0.1].

True Value of CCV2 = 12.02 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			4.01	11/09/2017	17:30
2	CAL2	1	Water			7.00	11/09/2017	17:31
3	CAL3	1	Water			10.03	11/09/2017	17:33
4	ICV	1	Water			6.99	11/09/2017	17:35
5	CCV1	1	Water			2.01	11/09/2017	17:36
6	I6399-24	1	Water			6.84	11/09/2017	17:37
7	I6399-27	1	Water			7.14	11/09/2017	17:39
8	I6407-08	1	Water			12.01	11/09/2017	17:41
9	I6409-05	1	Water			6.56	11/09/2017	17:42
10	I6409-08	1	Water			6.30	11/09/2017	17:43
11	I6413-01	1	Water			9.24	11/09/2017	17:45
12	I6413-01DUP	1	Water			9.25	11/09/2017	17:46
13	CCV2	1	Water			12.02	11/09/2017	17:47

2391488

WORKLIST(Hardcopy Internal Chain)

WorkList Name : ph i6413

WorkList ID : 105334

Date : 11/9/2017 6:02:12 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
11/08/2017	Water	16399-24	pH	Cool 4 deg C	PSEG01	H52	TP-12	11/08/2017	SM 4500-PH B
11/08/2017	Water	16399-27	pH	Cool 4 deg C	PSEG01	H52	TP-14	11/08/2017	SM 4500-PH B
11/09/2017	Water	16407-08	pH	Cool 4 deg C	PSEG01	H52	FM1034	11/09/2017	SM 4500-PH B
11/09/2017	Water	16409-05	pH	Cool 4 deg C	PSEG01	H52	MH-110917-01	11/09/2017	SM 4500-PH B
11/09/2017	Water	16409-08	pH	Cool 4 deg C	PSEG01	H52	MH-110917-02-FREEZE-PIT	11/09/2017	SM 4500-PH B
11/09/2017	Water	16413-01	pH	Cool 4 deg C	PSEG01	H52	FM728	11/09/2017	SM 4500-PH B

Date/Time 11-09-17 6:02 PM
Received by: TL
Relinquished by: _____

Date/Time 11-09-17 7:20 PM
Received by: CP
Relinquished by: JP