

LB 91498


7470A

INSTRUMENT ID: CV4

Linear

μ Abs.:

34258



A= 0.0000e+000

B= 2.8799e-004

C= 7.1230e-003

Rho= 0.9996475

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	-0.007	-0.007	-49	0.000	-49				
0.2	0.200	0.202	0.002	676	0.0 %	676				
2.5	2.500	2.478	-0.022	8578	0.0 %	8578				
5.0	5.000	4.957	-0.043	17187	0.0 %	17187				
7.5	7.500	7.698	0.198	26704	0.0 %	26704				
10.0	10.000	9.873	-0.127	34258	0.0 %	34258				

LB91498 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 10 Nov 2017 08:12:06

Sample ID	Extended ID	μ Abs	Conc	Std Conc/Method	Units	Date	Type	Type
0.0 - 1	50	-49	-	0.00007470A	PPB	10 Nov 2017 08:12:14	S	Std
0.2 - 1	50.2	676	-	0.20007470A	PPB	10 Nov 2017 08:14:23	S	Std
2.5 - 1	52.5	8578	-	2.50007470A	PPB	10 Nov 2017 08:16:31	S	Std
5.0 - 1	55	17187	-	5.00007470A	PPB	10 Nov 2017 08:18:40	S	Std
7.5 - 1	57.5	26704	-	7.50007470A	PPB	10 Nov 2017 08:20:47	S	Std
10.0 - 1	510	34258	-	10.00007470A	PPB	10 Nov 2017 08:30:21	S	Std
ICV56 - 1	ICV56	13362	3.8553	-7470A	PPB	10 Nov 2017 08:32:53	U	SMPL
ICB56 - 1	ICB56	-181	-0.0450	-7470A	PPB	10 Nov 2017 08:34:59	U	SMPL
CCV50 - 1	CCV50	17103	4.9327	-7470A	PPB	10 Nov 2017 08:37:07	U	SMPL
CCB50 - 1	CCB50	-198	-0.0499	-7470A	PPB	10 Nov 2017 08:39:13	U	SMPL
CRA - 1	CRA	676	0.2018	-7470A	PPB	10 Nov 2017 08:41:21	U	SMPL
HighStd - 1	HighStd	34347	9.8988	-7470A	PPB	10 Nov 2017 08:43:28	U	SMPL
ChkStd - 1	ChkStd	24440	7.0456	-7470A	PPB	10 Nov 2017 08:45:34	U	SMPL
PB104135BL - 1	PBW	-239	-0.0617	-7470A	PPB	10 Nov 2017 08:47:43	U	SMPL
PB104135BS - 1	LCSW	12895	3.7208	-7470A	PPB	10 Nov 2017 08:49:53	U	SMPL
PB104135TB - 1	PB104099TB	-154	-0.0372	-7470A	PPB	10 Nov 2017 08:52:00	U	SMPL
I6289-04 - 1	SU-02-110817	-59	-0.0099	-7470A	PPB	10 Nov 2017 08:54:09	U	SMPL
I6289-04DUP - 1	SU-02-110817DUP	-57	-0.0093	-7470A	PPB	10 Nov 2017 08:56:17	U	SMPL
I6289-04MS - 1	SU-02-110817MS	12498	3.6064	-7470A	PPB	10 Nov 2017 08:58:26	U	SMPL
I6289-04MSD - 1	SU-02-110817MSD	13113	3.7836	-7470A	PPB	10 Nov 2017 09:00:33	U	SMPL
CCV51 - 1	CCV51	17132	4.9410	-7470A	PPB	10 Nov 2017 09:02:40	U	SMPL
CCB51 - 1	CCB51	-184	-0.0459	-7470A	PPB	10 Nov 2017 09:04:47	U	SMPL
I6382-10 - 1	TP-3	-79	-0.0156	-7470A	PPB	10 Nov 2017 09:06:55	U	SMPL
I6382-20 - 1	TP-2	-71	-0.0133	-7470A	PPB	10 Nov 2017 09:09:01	U	SMPL
I6382-22 - 1	DRUM-5-6	-53	-0.0081	-7470A	PPB	10 Nov 2017 09:11:08	U	SMPL
I6384-02 - 1	EAST-RUTHERFORD-ROLLOFFS	-57	-0.0093	-7470A	PPB	10 Nov 2017 09:13:15	U	SMPL
I6387-02 - 1	SEWAGE-INJECTION-TANK-WATER	-29	-0.0012	-7470A	PPB	10 Nov 2017 09:15:22	U	SMPL
I6396-04 - 1	LS-STONE	-64	-0.0113	-7470A	PPB	10 Nov 2017 09:17:30	U	SMPL
I6396-08 - 1	LS-SOIL	-40	-0.0044	-7470A	PPB	10 Nov 2017 09:19:38	U	SMPL
I6396-12 - 1	LS-CONCRETE	-48	-0.0067	-7470A	PPB	10 Nov 2017 09:21:47	U	SMPL
PB104137BL - 1	PBW	-61	-0.0104	-7470A	PPB	10 Nov 2017 09:23:55	U	SMPL
PB104137BS - 1	LCSW	13029	3.7594	-7470A	PPB	10 Nov 2017 09:26:02	U	SMPL
CCV52 - 1	CCV52	16955	4.8900	-7470A	PPB	10 Nov 2017 09:28:09	U	SMPL
CCB52 - 1	CCB52	-182	-0.0453	-7470A	PPB	10 Nov 2017 09:30:16	U	SMPL
I6387-01 - 1	SEWAGE-INJECTION-TANK-WATER	54295	15.6437	-7470A	PPB	10 Nov 2017 09:32:24	U	SMPL
I6399-24 - 1	TP-12	2268	0.6603	-7470A	PPB	10 Nov 2017 09:34:31	U	SMPL
I6399-27 - 1	TP-14	343	0.1059	-7470A	PPB	10 Nov 2017 09:36:40	U	SMPL
I6399-27DUP - 1	TP-14DUP	346	0.1068	-7470A	PPB	10 Nov 2017 09:38:47	U	SMPL
I6399-27MS - 1	TP-14MS	15524	4.4779	-7470A	PPB	10 Nov 2017 09:40:55	U	SMPL
I6399-27MSD - 1	TP-14MSD	15708	4.5309	-7470A	PPB	10 Nov 2017 09:43:02	U	SMPL
I6289-04LX5 - 1		-155	-0.0375	-7470A	PPB	10 Nov 2017 09:45:11	U	SMPL
I6289-04A - 1		13704	3.9538	-7470A	PPB	10 Nov 2017 09:47:19	U	SMPL
I6399-27LX5 - 1		-67	-0.0122	-7470A	PPB	10 Nov 2017 09:49:27	U	SMPL
I6399-27A - 1		16368	4.7210	-7470A	PPB	10 Nov 2017 09:51:35	U	SMPL
CCV53 - 1	CCV53	16956	4.8903	-7470A	PPB	10 Nov 2017 09:53:42	U	SMPL
CCB53 - 1	CCB53	-167	-0.0410	-7470A	PPB	10 Nov 2017 09:55:49	U	SMPL
I6387-01DLX2 - 1		27524	7.9338	-7470A	PPB	10 Nov 2017 09:57:57	U	SMPL
CCV54 - 1	CCV54	17017	4.9079	-7470A	PPB	10 Nov 2017 10:00:04	U	SMPL
CCB54 - 1	CCB54	-169	-0.0415	-7470A	PPB	10 Nov 2017 10:02:14	U	SMPL