



Analytical Summary Report

Analysis Method: 9045C

Analyst By : Jignesh

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB91503

Temp.: 22.3 °C

BalanceID: WC SC-4

SLOPE.: 98.4

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2270
buffer solution pH 7 yellow	W2159
PH 10.01 BUFFER,COLOR CD 475ML	W2154
buffer solution pH 7 yellow	W2219
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.01 Control Limits[+/- 0.1].

True Value of CCV2 = 12.01 Control Limits[+/- 0.1].

True Value of CCV3 = 12.01 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			4.01	10/27/2017	16:00
2	CAL2	1	Water			7.00	10/27/2017	16:01
3	CAL3	1	Water			10.04	10/27/2017	16:03
4	ICV	1	Water			7.00	10/27/2017	16:05
5	CCV1	1	Water			2.01	10/27/2017	16:06
6	I6167-03	1	Solid	20.02	20	7.44	10/27/2017	16:07
7	I6174-03	1	Solid	20.03	20	7.50	10/27/2017	16:09
8	I6182-02	1	Solid	20.01	20	9.72	10/27/2017	16:11
9	I6194-02	1	Solid	20.04	20	8.47	10/27/2017	16:13
10	I6196-02	1	Solid	20.03	20	8.54	10/27/2017	16:16
11	I6200-02	1	Solid	20.02	20	7.98	10/27/2017	16:18
12	I6202-02	1	Solid	20.05	20	8.17	10/27/2017	16:20
13	I6206-02	1	Solid	20.03	20	8.17	10/27/2017	16:21
14	I6206-04	1	Solid	20.04	20	8.23	10/27/2017	16:23
15	I6206-06	1	Solid	20.05	20	8.65	10/27/2017	16:25
16	CCV2	1	Water			12.02	10/27/2017	16:27
17	I6206-06DUP	1	Solid	20.02	20	8.66	10/27/2017	16:30
18	CCV3	1	Water			12.01	10/27/2017	16:31

2391503

WORKLIST(Hardcopy Internal Chain)

WorkList Name : CORROSIONITY I6200 WorkList ID : 104802

Date : 10/27/2017 3:27:15 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
11/01/2017	Solid	I6202-02	Corrosivity	Cool 4 deg C	BEMS01	F31	ROLL-OFF-2	10/27/2017	9045C
11/01/2017	Solid	I6182-02	Corrosivity	Cool 4 deg C	PSEG01	D52	RT-3026	10/27/2017	9045C
11/01/2017	Solid	I6194-02	Corrosivity	Cool 4 deg C	PSEG05	D52	BU-01-102717	10/27/2017	9045C
11/01/2017	Solid	I6196-02	Corrosivity	Cool 4 deg C	PSEG05	D52	OK-03-102717	10/27/2017	9045C
11/01/2017	Solid	I6200-02	Corrosivity	Cool 4 deg C	PSEG05	D52	OR-01-102717	10/27/2017	9045C
11/01/2017	Solid	I6206-02	Corrosivity	Cool 4 deg C	PSEG05	D52	EO-01-102717-A	10/27/2017	9045C
11/01/2017	Solid	I6206-04	Corrosivity	Cool 4 deg C	PSEG05	D52	EO-01-102717-B	10/27/2017	9045C
11/01/2017	Solid	I6206-06	Corrosivity	Cool 4 deg C	PSEG05	D52	EO-01-102717-C	10/27/2017	9045C
10/31/2017	Solid	I6167-03	Corrosivity	Cool 4 deg C	SKAN06	F21	S13-SP2C	10/26/2017	9045C
11/09/2017	Solid	I6174-03	Corrosivity	Cool 4 deg C	WSP001	F41	B7(0-10)COMP	10/26/2017	9045C

Date/Time 10-27-17 5:27 AM
Received by: JP
Relinquished by: CA

Date/Time 10-27-17 7:10 AM
Received by: LP
Relinquished by: JP