



Analytical Summary Report

Analysis Method: SM 4500-PH B

Analyst By : JIGNESH

Parameter: pH

Supervisor Review By : apatel

Run Number: LB91518

Temp.: 19.6 °C

SLOPE.: 97.7

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2270
buffer solution pH 7 yellow	W2159
PH 10.01 BUFFER,COLOR CD 475ML	W2154
buffer solution pH 7 yellow	W2219
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.01 Control Limits[+/- 0.1].

True Value of CCV2 = 12.01 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			4.01	11/10/2017	13:45
2	CAL2	1	Water			7.00	11/10/2017	13:46
3	CAL3	1	Water			10.03	11/10/2017	13:47
4	ICV	1	Water			7.00	11/10/2017	13:48
5	CCV1	1	Water			2.01	11/10/2017	13:50
6	I6434-01	1	Water			4.50	11/10/2017	13:51
7	I6434-01DUP	1	Water			4.51	11/10/2017	13:52
8	I6434-02	1	Water			7.81	11/10/2017	14:00
9	I6435-01	1	Water			5.55	11/10/2017	14:01
10	CCV2	1	Water			12.01	11/10/2017	14:02

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH I6434

WorkList ID : 105385

Date : 11/10/2017 1:34:45 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
11/08/2017	Water	I6434-01	pH	Cool 4 deg C	STER10	J62	CELL-A-POOL-WATER	11/08/2017	SM 4500-PH B
11/08/2017	Water	I6434-02	pH	Cool 4 deg C	STER10	J62	CITY-WATER	11/08/2017	SM 4500-PH B
11/09/2017	Water	I6435-01	pH	Cool 4 deg C	STER12	J63	CELL-POOL-WATER	11/09/2017	SM 4500-PH B

Date/Time 11-10-17 1:34 PM
 Received by: JH
 Relinquished by: CJ

Date/Time 11-10-17 2:45 PM
 Received by: CD
 Relinquished by: JH