


ISM02.4

LB91526
INSTRUMENT ID: CV1

Linear

 μ Abs.:

39943



A= 0.0000e+000

B= 2.5141e-004

C= 1.4799e-002

Rho= 0.9999077

Accept=Accepted

Slope
x-intercept

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	σ/σD
0.00	0.000	0.007	0.007	-33	0.000	-33					-
0.2	0.200	0.223	0.023	829	0.0 %	829					12
2.5	2.500	2.536	0.036	10028	0.0 %	10028					1
5.0	5.000	4.904	-0.096	19446	0.0 %	19446					-2
7.5	7.500	7.473	-0.027	29667	0.0 %	29667					0
10.0	10.000	10.057	0.057	39943	0.0 %	39943					1

n03
11/10/12

LB91526
INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0	-33	-		0 ISM02.4	PPB	11/10/2017 18:06	Std
	0.2 S0.2	829	-		0.2 ISM02.4	PPB	11/10/2017 18:08	Std
	2.5 S2.5	10028	-		2.5 ISM02.4	PPB	11/10/2017 18:10	Std
	5 S5	19446	-		5 ISM02.4	PPB	11/10/2017 18:13	Std
	7.5 S7.5	29667	-		7.5 ISM02.4	PPB	11/10/2017 18:15	Std
	10 S10	39943	-		10 ISM02.4	PPB	11/10/2017 18:17	Std
ICV	ICV	15280	3.8564 -		ISM02.4	PPB	11/10/2017 18:20	SMPL
ICB	ICB	-177	-0.0297 -		ISM02.4	PPB	11/10/2017 18:22	SMPL
CCV066	CCV066	19718	4.9722 -		ISM02.4	PPB	11/10/2017 18:27	SMPL
CCB066	CCB066	-168	-0.0274 -		ISM02.4	PPB	11/10/2017 18:29	SMPL
PB104138BL	PBW003	162	0.0555 -		ISM02.4	PPB	11/10/2017 18:31	SMPL
I6362-01	MBE2Q4	585	0.1619 -		ISM02.4	PPB	11/10/2017 18:34	SMPL
I6362-02	MBE2Q5	216	0.0691 -		ISM02.4	PPB	11/10/2017 18:36	SMPL
I6362-03	MBE2Q7	302	0.0907 -		ISM02.4	PPB	11/10/2017 18:38	SMPL
I6362-04	MBE2Q8	218	0.0696 -		ISM02.4	PPB	11/10/2017 18:40	SMPL
I6362-05	MBE2Q9	294	0.0887 -		ISM02.4	PPB	11/10/2017 18:43	SMPL
I6362-06	MBE2R1	203	0.0658 -		ISM02.4	PPB	11/10/2017 18:45	SMPL
I6362-07	MBE2R2	305	0.0915 -		ISM02.4	PPB	11/10/2017 18:47	SMPL
I6362-08	MBE2R3	214	0.0686 -		ISM02.4	PPB	11/10/2017 18:50	SMPL
I6362-09	MBE2R4	226	0.0716 -		ISM02.4	PPB	11/10/2017 18:52	SMPL
I6362-10	MBE2R5	273	0.0834 -		ISM02.4	PPB	11/10/2017 18:54	SMPL
I6362-11	MBE2R6	343	0.101 -		ISM02.4	PPB	11/10/2017 18:56	SMPL
I6362-12	MBE2R6D	347	0.102 -		ISM02.4	PPB	11/10/2017 18:59	SMPL
I6362-13	MBE2R6S	4652	1.1844 -		ISM02.4	PPB	11/10/2017 19:01	SMPL
I6362-14	MBE2R7	280	0.0852 -		ISM02.4	PPB	11/10/2017 19:03	SMPL
I6362-15	MBE2R8	219	0.0699 -		ISM02.4	PPB	11/10/2017 19:05	SMPL
I6362-16	MBE2R9	253	0.0784 -		ISM02.4	PPB	11/10/2017 19:08	SMPL
I6362-17	MBE2S0	226	0.0716 -		ISM02.4	PPB	11/10/2017 19:10	SMPL
I6362-18	MBE2S1	242	0.0756 -		ISM02.4	PPB	11/10/2017 19:12	SMPL
I6362-19	MBE2S2	233	0.0734 -		ISM02.4	PPB	11/10/2017 19:15	SMPL
CCV067	CCV067	19485	4.9136 -		ISM02.4	PPB	11/10/2017 19:17	SMPL
CCB067	CCB067	-145	-0.0217 -		ISM02.4	PPB	11/10/2017 19:19	SMPL
I6362-20	MBE2S4	253	0.0784 -		ISM02.4	PPB	11/10/2017 19:21	SMPL
I6362-21	MBE2S5	510	0.143 -		ISM02.4	PPB	11/10/2017 19:24	SMPL
I6362-22	MBE2S6	258	0.0797 -		ISM02.4	PPB	11/10/2017 19:26	SMPL
CCV068	CCV068	19737	4.977 -		ISM02.4	PPB	11/10/2017 19:28	SMPL
CCB068	CCB068	-142	-0.0209 -		ISM02.4	PPB	11/10/2017 19:30	SMPL