



Analytical Summary Report

Analysis Method: 9045C

Analyst By : Jignesh

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB91528

Temp.: 22.6 °C

BalanceID: WC SC-4

SLOPE.: 98.7

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2270
buffer solution pH 7 yellow	W2159
PH 10.01 BUFFER,COLOR CD 475ML	W2154
buffer solution pH 7 yellow	W2219
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.01 Control Limits[+/- 0.1].

True Value of CCV2 = 2.01 Control Limits[+/- 0.1].

True Value of CCV3 = 12.02 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			4.01	11/10/2017	17:00
2	CAL2	1	Water			7.00	11/10/2017	17:01
3	CAL3	1	Water			10.03	11/10/2017	17:03
4	ICV	1	Water			6.99	11/10/2017	17:05
5	CCV1	1	Water			2.01	11/10/2017	17:06
6	I6300-04	1	Solid	20.02	20	7.88	11/10/2017	17:09
7	I6300-04DUP	1	Solid	20.03	20	7.89	11/10/2017	17:10
8	I6307-02	1	Solid	20.04	20	8.68	11/10/2017	17:11
9	I6307-04	1	Solid	20.02	20	8.80	11/10/2017	17:12
10	I6307-06	1	Solid	20.05	20	8.63	11/10/2017	17:13
11	I6311-04	1	Solid	20.01	20	8.33	11/10/2017	17:16
12	I6408-04	1	Solid	20.02	20	7.76	11/10/2017	17:19
13	I6423-03	1	Solid	20.04	20	8.64	11/10/2017	17:20
14	I6427-03	1	Solid	20.01	20	7.80	11/10/2017	17:21
15	I6429-02	1	Solid	20.04	20	6.65	11/10/2017	17:22
16	CCV2	1	Water			2.01	11/10/2017	17:25
17	I6436-02	1	Solid	20.02	20	8.38	11/10/2017	17:26
18	I6441-04	1	Solid	20.03	20	7.70	11/10/2017	17:29
19	I6441-08	1	Solid	20.00	20	7.24	11/10/2017	17:30
20	I6441-13	1	Solid	20.02	20	8.71	11/10/2017	17:31

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Result (pH)	Anal Date	Anal Time
21	I6441-19	1	Solid	20.03	20	8.65	11/10/2017	17:32
22	I6441-24	1	Solid	20.04	20	8.79	11/10/2017	17:33
23	I6441-30	1	Solid	20.01	20	8.00	11/10/2017	17:35
24	I6441-30DUP	1	Solid	20.03	20	8.01	11/10/2017	17:36
25	CCV3	1	Water			12.01	11/10/2017	17:39

LB91528

WORKLIST(Hardcopy Internal Chain)

WorkList Name : CORROSIONITY I6349

WorkList ID : 105399

Date : 11/10/2017 6:20:30 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
11/15/2017	Solid	I6300-04	Corrosivity	Cool 4 deg C	PSEG05		OR-03-111017	11/10/2017	9045C
11/15/2017	Solid	I6307-02	Corrosivity	Cool 4 deg C	PSEG05	J52	EO-02-111017-A	11/10/2017	9045C
11/15/2017	Solid	I6307-04	Corrosivity	Cool 4 deg C	PSEG05	J52	EO-02-111017-B	11/10/2017	9045C
11/15/2017	Solid	I6307-06	Corrosivity	Cool 4 deg C	PSEG05	J52	EO-02-111017-C	11/10/2017	9045C
11/15/2017	Solid	I6311-04	Corrosivity	Cool 4 deg C	PSEG05		TR-05-111017	11/10/2017	9045C
11/16/2017	Solid	I6408-04	Corrosivity	Cool 4 deg C	DVIR01	I51	WASTE-CHARACTERIZATION	11/08/2017	9045C
11/23/2017	Solid	I6423-03	Corrosivity	Cool 4 deg C	PARS01	H52	B8(0-10)COMP	11/09/2017	9045C
11/15/2017	Solid	I6427-03	Corrosivity	Cool 4 deg C	PSEG01	H52	277-216-156-343-350-342	11/10/2017	9045C
11/17/2017	Solid	I6429-02	Corrosivity	Cool 4 deg C	AREC01	I31	SP-1	11/09/2017	9045C
11/15/2017	Solid	I6436-02	Corrosivity	Cool 4 deg C	PSEG01	H52	153RD-63RD-ST-DRUMS	11/10/2017	9045C
11/15/2017	Solid	I6441-04	Corrosivity	Cool 4 deg C	PSEG01	H52	TP-11A-11B	11/10/2017	9045C
11/15/2017	Solid	I6441-08	Corrosivity	Cool 4 deg C	PSEG01	H52	TP-10	11/10/2017	9045C
11/15/2017	Solid	I6441-13	Corrosivity	Cool 4 deg C	PSEG01	H52	TP-9A-9B	11/10/2017	9045C
11/15/2017	Solid	I6441-19	Corrosivity	Cool 4 deg C	PSEG01	H52	TP-8	11/10/2017	9045C
11/15/2017	Solid	I6441-24	Corrosivity	Cool 4 deg C	PSEG01	H52	TP-7A-7B	11/10/2017	9045C
11/15/2017	Solid	I6441-30	Corrosivity	Cool 4 deg C	PSEG01	H52	TP-16	11/10/2017	9045C

Date/Time 11-10-17 6:20 PM

Received by: J

Relinquished by: S

Date/Time 11-10-17 8:00 PM

Received by: CD

Relinquished by: JD