



Analytical Summary Report

Analysis Method: SM 4500-PH B

Analyst By : JIGNESH

Parameter: pH

Supervisor Review By : apatel

Run Number: LB91530

Temp.: 20.1 °C

SLOPE.: 97.6

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2270
buffer solution pH 7 yellow	W2159
PH 10.01 BUFFER,COLOR CD 475ML	W2154
buffer solution pH 7 yellow	W2219
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.01 Control Limits[+/- 0.1].

True Value of CCV2 = 12.01 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			4.01	11/10/2017	18:20
2	CAL2	1	Water			6.99	11/10/2017	18:21
3	CAL3	1	Water			10.02	11/10/2017	18:22
4	ICV	1	Water			7.00	11/10/2017	18:24
5	CCV1	1	Water			2.01	11/10/2017	18:25
6	I6436-03	1	Water			7.65	11/10/2017	18:26
7	I6436-03DUP	1	Water			7.66	11/10/2017	18:27
8	I6441-31	1	Water			7.51	11/10/2017	18:29
9	I6443-01	1	Water			7.83	11/10/2017	18:30
10	CCV2	1	Water			12.01	11/10/2017	18:31

WORKLIST(Hardcopy Internal Chain)

23 91530

WorkList Name : PH I6441

WorkList ID : 105401

Date : 11/10/2017 6:22:22 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
11/10/2017	Water	I6436-03	pH	Cool 4 deg C	PSEG01	H52	153RD-63RD-ST-DRUMS	11/10/2017	SM 4500-PH B
11/10/2017	Water	I6441-31	pH	Cool 4 deg C	PSEG01	H52	TP-8	11/10/2017	SM 4500-PH B
11/10/2017	Water	I6443-01	pH	Cool 4 deg C	PSEG01	J52	FADW4870L	11/10/2017	SM 4500-PH B

Date/Time 11-10-17 6:20 PM
 Received by: JP
 Relinquished by: JP

Date/Time 11-10-17 8:00 PM
 Received by: JP
 Relinquished by: JP