

TOTAL SUSPENDED SOLIDS - SM2540D

SUPERVISOR: apatel

ANALYST: Aparna

Date: 11/13/2017

DATE & TIME IN: 11/13/2017 10:00

OVEN TEMP: 103 °C

Run Number: LB91534

DATE & TIME OUT: 11/13/2017 14:00

OVEN TEMP: 105 °C

BalanceID: WC SC-3

OvenID: WC OVEN-2

Dish #	Lab ID	Client ID	Empty Dish Weight (g)	Final Empty Dish Weight (g)	Sample Volume (ml)	1st Empty Dish+Sample weight after 1.5hr drying @103-@105°C (g)	2nd Empty Dish+Sample weight after 1.5hr drying @103-@105°C (g)	Final Empty Dish+Sample weight after 1.5hr drying @103-@105°C (g)	Weight (g)	Result mg/L
1	LB91534BL	LB91534BL	1.3966	1.3966	100	1.3966	1.3966	1.3966	0	0
2	LB91534BS	LB91534BS	1.3997	1.3997	100	1.4498	1.4498	1.4498	0.0501	501
3	I6426-01	110917-TIDO-E-D-B	1.4031	1.4031	300	1.4194	1.4194	1.4194	0.0163	54.3
4	I6426-02	110917-TIDO-E-D-S	1.3987	1.3987	200	1.41	1.41	1.41	0.0113	56.5
5	I6426-03	110917-TIDO-E-U-B	1.3974	1.3974	300	1.4166	1.4166	1.4166	0.0192	64
6	I6426-04	110917-TIDO-E-U-S	1.4	1.4	300	1.4163	1.4163	1.4163	0.0163	54.3
7	I6426-05	110917-SIDI-E-D-B	1.4032	1.4032	200	1.4194	1.4194	1.4194	0.0162	81
8	I6426-06	110917-SIDI-E-D-S	1.4048	1.4048	300	1.4241	1.4241	1.4241	0.0193	64.3
9	I6426-07	110917-SIDI-E-U-B	1.3963	1.3963	200	1.4101	1.4101	1.4101	0.0138	69
10	I6426-07DUP	110917-SIDI-E-U-BDUP	1.3998	1.3998	200	1.4139	1.4139	1.4139	0.0141	70.5
11	I6426-08	110917-SIDI-E-U-S	1.4045	1.4045	300	1.4172	1.4172	1.4172	0.0127	42.3
12	I6426-09	110917-JETT-F-D-B	1.4076	1.4076	300	1.4151	1.4151	1.4151	0.0075	25
13	I6426-10	110917-JETT-F-D-S	1.4084	1.4084	300	1.413	1.413	1.413	0.0046	15.3
14	I6426-11	110917-JETT-F-U-B	1.4134	1.4134	300	1.423	1.423	1.423	0.0096	32
15	I6426-12	110917-JETT-F-U-S	1.3939	1.3939	300	1.4004	1.4004	1.4004	0.0065	21.7

TOTAL SUSPENDED SOLIDS - SM2540D

SUPERVISOR: apatel

ANALYST: Aparna

Date: 11/13/2017

DATE & TIME IN: 11/13/2017 10:00

OVEN TEMP: 103 °C

Run Number: LB91534

DATE & TIME OUT: 11/13/2017 14:00

OVEN TEMP: 105 °C

BalanceID: WC SC-3

OvenID: WC OVEN-2

Dish #	Lab ID	Client ID	Empty Dish Weight (g)	Final Empty Dish Weight (g)	Sample Volume (ml)	1st Empty Dish+Sample weight after 1.5hr drying @103-@105°C (g)	2nd Empty Dish+Sample weight after 1.5hr drying @103-@105°C (g)	Final Empty Dish+Sample weight after 1.5hr drying @103-@105°C (g)	Weight (g)	Result mg/L
16	I6436-03	153RD-63RD-ST-DRUM	1.3982	1.3982	400	1.4035	1.4035	1.4035	0.0053	13.3
17	I6436-03DUP	153RD-63RD-ST-DRUMDUP	1.3925	1.3925	400	1.398	1.398	1.398	0.0055	13.7
18	I6441-31	TP-8	1.4017	1.4017	100	1.5633	1.5633	1.5633	0.1616	1616
19	I6443-01	FADW4870L	1.4051	1.4051	300	1.4167	1.4167	1.4167	0.0116	38.7

A = Sample Volume (ml)

B = Empty Dish Weight (g)

C = Final Empty Dish + Sample weight after 1.5 hr drying @105°C(g)

D = Weight (g)

Weight (g) = C - B

Result mg/L = $\frac{D}{A} \times 1000 \times 1000$

WORKLIST(Hardcopy Internal Chain)

WorkList Name : TSS111317

WorkList ID : 105407

Date : 11/13/2017 8:14:14 AM

1291534

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
11/16/2017	Water	I6426-01 -C	TSS	Cool 4 deg C	HDR103	H52	110917-TIDO-E-D-B	11/09/2017	SM2540 D
11/16/2017	Water	I6426-02 -C	TSS	Cool 4 deg C	HDR103	H52	110917-TIDO-E-D-S	11/09/2017	SM2540 D
11/16/2017	Water	I6426-03 -C	TSS	Cool 4 deg C	HDR103	H52	110917-TIDO-E-U-B	11/09/2017	SM2540 D
11/16/2017	Water	I6426-04 -C	TSS	Cool 4 deg C	HDR103	H52	110917-TIDO-E-U-S	11/09/2017	SM2540 D
11/16/2017	Water	I6426-05 -C	TSS	Cool 4 deg C	HDR103	H52	110917-SIDI-E-D-B	11/09/2017	SM2540 D
11/16/2017	Water	I6426-06 -C	TSS	Cool 4 deg C	HDR103	H52	110917-SIDI-E-D-S	11/09/2017	SM2540 D
11/16/2017	Water	I6426-07 -C	TSS	Cool 4 deg C	HDR103	H52	110917-SIDI-E-U-B	11/09/2017	SM2540 D
11/16/2017	Water	I6426-08 -C	TSS	Cool 4 deg C	HDR103	H52	110917-SIDI-E-U-S	11/09/2017	SM2540 D
11/16/2017	Water	I6426-09 -C	TSS	Cool 4 deg C	HDR103	H52	110917-JETT-F-D-B	11/09/2017	SM2540 D
11/16/2017	Water	I6426-10 -C	TSS	Cool 4 deg C	HDR103	H52	110917-JETT-F-D-S	11/09/2017	SM2540 D
11/16/2017	Water	I6426-11 -C	TSS	Cool 4 deg C	HDR103	H52	110917-JETT-F-U-B	11/09/2017	SM2540 D
11/16/2017	Water	I6426-12 -C	TSS	Cool 4 deg C	HDR103	H52	110917-JETT-F-U-S	11/09/2017	SM2540 D
11/15/2017	Water	I6436-03 -C	TSS	Cool 4 deg C	PSEG01	H52	153RD-63RD-ST-DRUMS	11/10/2017	SM2540 D
11/15/2017	Water	I6441-31 -C	TSS	Cool 4 deg C	PSEG01	H52	TP-8	11/10/2017	SM2540 D
11/15/2017	Water	I6443-01 -C	TSS	Cool 4 deg C	PSEG01	J52	FADW4870L	11/10/2017	SM2540 D

Date/Time 11/13/17, 9:10
 Received by: AS
 Relinquished by: AS

Date/Time 11/13/17, 10:15
 Received by: AS
 Relinquished by: AS