


7470A

LB 91535
INSTRUMENT ID: CY2

Linear

 μ Abs.:

34324



A= 0.0000e+000

B= 2.9055e-004

C= 7.5303e-002

Rho= 0.9998659

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.037	0.037	-131	0.000	-131				
0.2	0.200	0.226	0.026	519	0.0 %	519				
2.5	2.500	2.500	-0.000	8345	0.0 %	8345				
5.0	5.000	4.869	-0.131	16499	0.0 %	16499				
7.5	7.500	7.519	0.019	25621	0.0 %	25621				
10.0	10.000	10.048	0.048	34324	0.0 %	34324				

LB91535 INSTRUMENT ID : CV1

Method: 7470A		Operator: Admin		Date of Analysis: 13 Nov 2017 09:40:29					
Sample ID	Extended ID	µ Abs	Conc.	Std CondMethod	Units	Date	Type	Type	
0.0 - 1	50	-131	-	0.00007470A	PPB	13 Nov 2017 10:23:07	S	Std	
0.2 - 1	50.2	519	-	0.20007470A	PPB	13 Nov 2017 10:25:16	S	Std	
2.5 - 1	52.5	8345	-	2.50007470A	PPB	13 Nov 2017 10:27:24	S	Std	
5.0 - 1	55	16499	-	5.00007470A	PPB	13 Nov 2017 10:29:33	S	Std	
7.5 - 1	57.5	25621	-	7.50007470A	PPB	13 Nov 2017 10:31:41	S	Std	
10.0 - 1	510	34324	-	10.00007470A	PPB	13 Nov 2017 10:33:49	S	Std	
ICV59 - 1	ICV59	13577	4.0201	-7470A	PPB	13 Nov 2017 10:36:21	U	SMPL	
ICB59 - 1	ICB59	-188	0.0207	-7470A	PPB	13 Nov 2017 10:38:27	U	SMPL	
CCV69 - 1	CCV69	17187	5.0690	-7470A	PPB	13 Nov 2017 10:40:34	U	SMPL	
CCB69 - 1	CCB69	-181	0.0227	-7470A	PPB	13 Nov 2017 10:42:41	U	SMPL	
CRA - 1	CRA	524	0.2276	-7470A	PPB	13 Nov 2017 10:44:48	U	SMPL	
HighStd - 1	HighStd	34436	10.0807	-7470A	PPB	13 Nov 2017 10:46:55	U	SMPL	
ChkStd - 1	ChkStd	25037	7.3498	-7470A	PPB	13 Nov 2017 10:49:02	U	SMPL	
PB104175BL - 1	PBW	-259	0.0001	-7470A	PPB	13 Nov 2017 10:51:11	U	SMPL	
PB104175BS - 1	LCSW	13793	4.0829	-7470A	PPB	13 Nov 2017 10:53:19	U	SMPL	
PB104175TB - 1	PB104139TB	-157	0.0297	-7470A	PPB	13 Nov 2017 10:55:27	U	SMPL	
I6299-04 - 1	OK-01-110917	-30	0.0666	-7470A	PPB	13 Nov 2017 10:57:34	U	SMPL	
I6301-02 - 1	JC-02-110917-A	-49	0.0611	-7470A	PPB	13 Nov 2017 10:59:43	U	SMPL	
I6301-04 - 1	JC-02-110917-B	-47	0.0616	-7470A	PPB	13 Nov 2017 11:01:51	U	SMPL	
I6301-06 - 1	JC-02-110917-C	-52	0.0602	-7470A	PPB	13 Nov 2017 11:03:58	U	SMPL	
CCV70 - 1	CCV70	17416	5.1355	-7470A	PPB	13 Nov 2017 11:06:04	U	SMPL	
CCB70 - 1	CCB70	-169	0.0262	-7470A	PPB	13 Nov 2017 11:08:11	U	SMPL	
I6301-08 - 1	JC-02-110917-D	-60	0.0579	-7470A	PPB	13 Nov 2017 11:10:18	U	SMPL	
I6301-10 - 1	JC-02-110917-E	-47	0.0616	-7470A	PPB	13 Nov 2017 11:12:25	U	SMPL	
I6301-12 - 1	JC-02-110917-F	-48	0.0614	-7470A	PPB	13 Nov 2017 11:14:32	U	SMPL	
I6301-14 - 1	JC-02-110917-G	-48	0.0614	-7470A	PPB	13 Nov 2017 11:16:39	U	SMPL	
I6301-16 - 1	JC-02-110917-H	-39	0.0640	-7470A	PPB	13 Nov 2017 11:18:46	U	SMPL	
I6399-07 - 1	TP-12	160	0.1218	-7470A	PPB	13 Nov 2017 11:20:54	U	SMPL	
I6399-12 - 1	TP-13-A-13-B	-59	0.0582	-7470A	PPB	13 Nov 2017 11:23:02	U	SMPL	
I6399-18 - 1	TP-14	-52	0.0602	-7470A	PPB	13 Nov 2017 11:25:10	U	SMPL	
I6399-23 - 1	TP-15	-54	0.0596	-7470A	PPB	13 Nov 2017 11:27:18	U	SMPL	
I6407-02 - 1	3091-4235	-55	0.0593	-7470A	PPB	13 Nov 2017 11:29:25	U	SMPL	
CCV71 - 1	CCV71	16981	5.0091	-7470A	PPB	13 Nov 2017 11:31:32	U	SMPL	
CCB71 - 1	CCB71	-174	0.0247	-7470A	PPB	13 Nov 2017 11:33:39	U	SMPL	
I6407-05 - 1	4632	-65	0.0564	-7470A	PPB	13 Nov 2017 11:35:47	U	SMPL	
I6407-07 - 1	2334	-51	0.0605	-7470A	PPB	13 Nov 2017 11:37:54	U	SMPL	
I6409-02 - 1	MH-110917-01	-38	0.0643	-7470A	PPB	13 Nov 2017 11:40:01	U	SMPL	
I6409-04 - 1	MH-110917-02-FREEZE-PIT	-48	0.0614	-7470A	PPB	13 Nov 2017 11:42:08	U	SMPL	
I6409-04DUP - 1	MH-110917-02-FREEZE-PITDUP	-43	0.0628	-7470A	PPB	13 Nov 2017 11:44:16	U	SMPL	
I6409-04MS - 1	MH-110917-02-FREEZE-PITMS	12916	3.8280	-7470A	PPB	13 Nov 2017 11:46:23	U	SMPL	
I6409-04MSD - 1	MH-110917-02-FREEZE-PITMSD	12570	3.7275	-7470A	PPB	13 Nov 2017 11:48:31	U	SMPL	
PB104176BL - 1	PBW	-136	0.0358	-7470A	PPB	13 Nov 2017 11:50:40	U	SMPL	
PB104176BS - 1	LCSW	13290	3.9367	-7470A	PPB	13 Nov 2017 11:52:48	U	SMPL	
PB104176TB - 1	PB104146TB	-122	0.0399	-7470A	PPB	13 Nov 2017 11:54:55	U	SMPL	
CCV72 - 1	CCV72	17316	5.1065	-7470A	PPB	13 Nov 2017 11:57:03	U	SMPL	
CCB72 - 1	CCB72	-176	0.0242	-7470A	PPB	13 Nov 2017 11:59:10	U	SMPL	
I6399-25 - 1	TP-12	-39	0.0640	-7470A	PPB	13 Nov 2017 12:01:19	U	SMPL	
I6399-25DUP - 1	TP-12DUP	-34	0.0654	-7470A	PPB	13 Nov 2017 12:03:26	U	SMPL	
I6399-25MS - 1	TP-12MS	12984	3.8478	-7470A	PPB	13 Nov 2017 12:05:33	U	SMPL	
I6399-25MSD - 1	TP-12MSD	12893	3.8214	-7470A	PPB	13 Nov 2017 12:07:40	U	SMPL	
I6399-28 - 1	TP-14	-140	0.0346	-7470A	PPB	13 Nov 2017 12:09:47	U	SMPL	
I6407-09 - 1	FM1034	-48	0.0614	-7470A	PPB	13 Nov 2017 12:11:55	U	SMPL	
I6409-06 - 1	MH-110917-01	-19	0.0698	-7470A	PPB	13 Nov 2017 12:14:03	U	SMPL	
I6409-09 - 1	MH-110917-02-FREEZE-PIT	-16	0.0707	-7470A	PPB	13 Nov 2017 12:16:11	U	SMPL	
I6413-02 - 1	FM728	-5	0.0739	-7470A	PPB	13 Nov 2017 12:18:20	U	SMPL	
PB104177BL - 1	PBW	-28	0.0672	-7470A	PPB	13 Nov 2017 12:20:27	U	SMPL	
CCV73 - 1	CCV73	17011	5.0178	-7470A	PPB	13 Nov 2017 12:22:34	U	SMPL	
CCB73 - 1	CCB73	-159	0.0291	-7470A	PPB	13 Nov 2017 12:24:41	U	SMPL	
PB104177BS - 1	LCSW	13492	3.9954	-7470A	PPB	13 Nov 2017 12:26:49	U	SMPL	
I6407-08 - 1	FM1034	-103	0.0454	-7470A	PPB	13 Nov 2017 12:28:56	U	SMPL	
I6409-05 - 1	MH-110917-01	39644	11.5939	-7470A	PPB	13 Nov 2017 12:31:04	U	SMPL	
I6409-08 - 1	MH-110917-02-FREEZE-PIT	7780	2.3358	-7470A	PPB	13 Nov 2017 12:33:11	U	SMPL	
I6413-01 - 1	FM728	221	0.1395	-7470A	PPB	13 Nov 2017 12:35:20	U	SMPL	
I6413-01DUP - 1	FM728DUP	235	0.1436	-7470A	PPB	13 Nov 2017 12:37:28	U	SMPL	
I6413-01MS - 1	FM728MS	15632	4.6172	-7470A	PPB	13 Nov 2017 12:39:36	U	SMPL	
I6413-01MSD - 1	FM728MSD	15504	4.5800	-7470A	PPB	13 Nov 2017 12:41:44	U	SMPL	
CCV74 - 1	CCV74	17957	5.2927	-7470A	PPB	13 Nov 2017 12:43:53	U	SMPL	
CCB74 - 1	CCB74	-174	0.0247	-7470A	PPB	13 Nov 2017 12:45:59	U	SMPL	
I6409-04LX5 - 1		-65	0.0564	-7470A	PPB	13 Nov 2017 12:48:08	U	SMPL	
I6409-04A - 1		13719	4.0614	-7470A	PPB	13 Nov 2017 12:50:15	U	SMPL	
I6399-25LX5 - 1		-133	0.0367	-7470A	PPB	13 Nov 2017 12:52:22	U	SMPL	
I6399-25A - 1		13387	3.9649	-7470A	PPB	13 Nov 2017 12:54:29	U	SMPL	
I6413-01LX5 - 1		-53	0.0599	-7470A	PPB	13 Nov 2017 12:56:37	U	SMPL	
I6413-01A - 1		16097	4.7523	-7470A	PPB	13 Nov 2017 12:58:44	U	SMPL	
I6409-05DLX2 - 1		16955	5.0016	-7470A	PPB	13 Nov 2017 13:00:51	U	SMPL	
CCV75 - 1	CCV75	17124	5.0507	-7470A	PPB	13 Nov 2017 13:02:59	U	SMPL	
CCB75 - 1	CCB75	-166	0.0271	-7470A	PPB	13 Nov 2017 13:05:07	U	SMPL	