

7471A


LB91547

INSTRUMENT ID: CV1

Linear

 μ Abs.:

43377



A= 0.0000e+000

B= 2.3147e-004

C= 1.7232e-002

Rho= 0.9997152

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.033	0.033	70	0.000	70				
0.2	0.200	0.197	-0.003	777	0.0 %	777				
2.5	2.500	2.428	-0.072	10417	0.0 %	10417				
5.0	5.000	5.128	0.128	22078	0.0 %	22078				
7.5	7.500	7.356	-0.144	31706	0.0 %	31706	0			
10.0	10.000	10.057	0.057	43377	0.0 %	43377				

LB91547 INSTRUMENT ID : CV1

Method: 7471A

Operator: Admin

Date of Analysis: 13 Nov 2017 14:50:11

Sample ID	Extended ID	μ Abs.	Conc.	Std Cond	Method	Units	Date	Type	Type
0.0 - 1	30	70	-	0.00007471A	PPB		13 Nov 2017 15:04:56	S	Std
0.2 - 1	30.2	777	-	0.20007471A	PPB		13 Nov 2017 15:07:14	S	Std
2.5 - 1	30.5	10417	-	2.50007471A	PPB		13 Nov 2017 15:09:31	S	Std
5.0 - 1	35	22078	-	5.00007471A	PPB		13 Nov 2017 15:14:49	S	Std
7.5 - 1	37.5	31706	-	7.50007471A	PPB		13 Nov 2017 15:17:06	S	Std
10.0 - 1	3.10	43377	-	10.00007471A	PPB		13 Nov 2017 15:19:24	S	Std
ICV60 - 1	ICV60	16259	3.7806	-7471A	PPB		13 Nov 2017 15:22:07	U	SMPL
ICB60 - 1	ICB60	-182	-0.0249	-7471A	PPB		13 Nov 2017 15:24:23	U	SMPL
CCV76 - 1	CCV76	22160	5.1465	-7471A	PPB		13 Nov 2017 15:29:10	U	SMPL
CCB76 - 1	CCB76	-198	-0.0286	-7471A	PPB		13 Nov 2017 15:31:25	U	SMPL
CRA - 1	CRA	740	0.1885	-7471A	PPB		13 Nov 2017 15:33:42	U	SMPL
HighStd - 1	HighStd	43745	10.1427	-7471A	PPB		13 Nov 2017 15:35:57	U	SMPL
ChkStd - 1	ChkStd	28520	6.6186	-7471A	PPB		13 Nov 2017 15:38:14	U	SMPL
PB104193BL - 1	PBS	-263	-0.0436	-7471A	PPB		13 Nov 2017 15:40:32	U	SMPL
PB104193BS - 1	LCSS	15547	3.6158	-7471A	PPB		13 Nov 2017 15:42:50	U	SMPL
I6300-03 - 1	OR-03-111017	9015	2.1039	-7471A	PPB		13 Nov 2017 15:45:07	U	SMPL
I6301-01 - 1	JC-02-110917-A	30014	6.9644	-7471A	PPB		13 Nov 2017 15:47:24	U	SMPL
I6301-03 - 1	JC-02-110917-B	1215	0.2985	-7471A	PPB		13 Nov 2017 15:49:42	U	SMPL
I6301-05 - 1	JC-02-110917-C	13554	3.1545	-7471A	PPB		13 Nov 2017 15:52:01	U	SMPL
I6301-07 - 1	JC-02-110917-D	38223	8.8645	-7471A	PPB		13 Nov 2017 15:54:17	U	SMPL
CCV77 - 1	CCV77	20776	4.8262	-7471A	PPB		13 Nov 2017 15:56:32	U	SMPL
CCB77 - 1	CCB77	-230	-0.0360	-7471A	PPB		13 Nov 2017 15:58:50	U	SMPL
I6301-09 - 1	JC-02-110917-E	9443	2.2030	-7471A	PPB		13 Nov 2017 16:01:07	U	SMPL
I6301-11 - 1	JC-02-110917-F	5832	1.3671	-7471A	PPB		13 Nov 2017 16:03:23	U	SMPL
I6301-13 - 1	JC-02-110917-G	25241	5.8596	-7471A	PPB		13 Nov 2017 16:05:39	U	SMPL
I6301-15 - 1	JC-02-110917-H	22155	5.1453	-7471A	PPB		13 Nov 2017 16:07:55	U	SMPL
I6407-01 - 1	3091-4235	1283	0.3142	-7471A	PPB		13 Nov 2017 16:10:13	U	SMPL
I6407-04 - 1	4632	2675	0.6364	-7471A	PPB		13 Nov 2017 16:12:31	U	SMPL
I6407-06 - 1	2334	2769	0.6582	-7471A	PPB		13 Nov 2017 16:14:48	U	SMPL
I6407-06DUP - 1	2334DUP	2834	0.6732	-7471A	PPB		13 Nov 2017 16:17:06	U	SMPL
I6407-06MS - 1	2334MS	19181	4.4570	-7471A	PPB		13 Nov 2017 16:19:23	U	SMPL
I6407-06MSD - 1	2334MSD	19416	4.5114	-7471A	PPB		13 Nov 2017 16:21:39	U	SMPL
CCV78 - 1	CCV78	20214	4.6961	-7471A	PPB		13 Nov 2017 16:23:56	U	SMPL
CCB78 - 1	CCB78	-189	-0.0265	-7471A	PPB		13 Nov 2017 16:26:12	U	SMPL
I6408-03 - 1	WASTE-CHARACTERIZATION	422	0.1149	-7471A	PPB		13 Nov 2017 16:28:30	U	SMPL
I6409-01 - 1	MH-110917-01	15306	3.5600	-7471A	PPB		13 Nov 2017 16:30:46	U	SMPL
I6409-03 - 1	MH-110917-02-FREEZE-PIT	7206	1.6852	-7471A	PPB		13 Nov 2017 16:33:02	U	SMPL
I6419-01 - 1	LEW-1-E(7-8)	94	0.0390	-7471A	PPB		13 Nov 2017 16:35:19	U	SMPL
I6419-02 - 1	LEW-2-E(1-2)	11995	2.7937	-7471A	PPB		13 Nov 2017 16:37:37	U	SMPL
I6419-03 - 1	LEW-2-E(4-5)	13008	3.0281	-7471A	PPB		13 Nov 2017 16:39:54	U	SMPL
I6419-04 - 1	LEW-2-E(6-7)	12303	2.8649	-7471A	PPB		13 Nov 2017 16:42:11	U	SMPL
PB104194BL - 1	PBS	-128	-0.0124	-7471A	PPB		13 Nov 2017 16:44:28	U	SMPL
PB104194BS - 1	PB104194BS	15887	3.6945	-7471A	PPB		13 Nov 2017 16:46:46	U	SMPL
I6311-03 - 1	TR-05-111017	1479	0.3596	-7471A	PPB		13 Nov 2017 16:49:02	U	SMPL
CCV79 - 1	CCV79	21236	4.9326	-7471A	PPB		13 Nov 2017 16:51:19	U	SMPL
CCB79 - 1	CCB79	-198	-0.0286	-7471A	PPB		13 Nov 2017 16:53:35	U	SMPL
I6422-01 - 1	EPE-110-3(0-5)	15366	3.5739	-7471A	PPB		13 Nov 2017 16:55:53	U	SMPL
I6422-02 - 1	EPE-110-3(5-10)	53638	12.4326	-7471A	PPB		13 Nov 2017 16:58:09	U	SMPL
I6422-03 - 1	EPE-P-10(0-5)	17298	4.0211	-7471A	PPB		13 Nov 2017 17:00:26	U	SMPL
I6422-04 - 1	EPE-P-10(5-10)	2523	0.6012	-7471A	PPB		13 Nov 2017 17:02:45	U	SMPL
I6422-05 - 1	EPE-120-2(0-5)	7885	1.8423	-7471A	PPB		13 Nov 2017 17:05:02	U	SMPL
I6422-06 - 1	EPE-120-2(5-10)	2789	0.6628	-7471A	PPB		13 Nov 2017 17:07:19	U	SMPL
I6422-06DUP - 1	EPE-120-2(5-10)DUP	2634	0.6269	-7471A	PPB		13 Nov 2017 17:09:36	U	SMPL
I6422-06MS - 1	EPE-120-2(5-10)MS	19272	4.4780	-7471A	PPB		13 Nov 2017 17:11:54	U	SMPL
I6422-06MSD - 1	EPE-120-2(5-10)MSD	19805	4.6014	-7471A	PPB		13 Nov 2017 17:14:11	U	SMPL
I6427-01 - 1	277-216-156-343-350-342	3701	0.8739	-7471A	PPB		13 Nov 2017 17:16:28	U	SMPL
CCV80 - 1	CCV80	21211	4.9268	-7471A	PPB		13 Nov 2017 17:18:45	U	SMPL
CCB80 - 1	CCB80	-207	-0.0307	-7471A	PPB		13 Nov 2017 17:21:02	U	SMPL
6429-01 - 1	SP-1	521	0.1378	-7471A	PPB		13 Nov 2017 17:23:19	U	SMPL
6436-01 - 1	153RD-63RD-ST-DRUM	1088	0.2691	-7471A	PPB		13 Nov 2017 17:25:36	U	SMPL
6441-03 - 1	TP-11A-11B	19054	4.4276	-7471A	PPB		13 Nov 2017 17:27:52	U	SMPL
6441-07 - 1	TP-10	87010	20.1570	-7471A	PPB		13 Nov 2017 17:30:09	U	SMPL
6441-12 - 1	TP-9A-9B	1649	0.3989	-7471A	PPB		13 Nov 2017 17:32:25	U	SMPL
6441-18 - 1	TP-8	198290	45.9145	-7471A	PPB		13 Nov 2017 17:34:45	U	SMPL
6441-23 - 1	TP-7A-7B	17293	4.0200	-7471A	PPB		13 Nov 2017 17:37:02	U	SMPL
6441-29 - 1	TP-16	6981	1.6331	-7471A	PPB		13 Nov 2017 17:39:30	U	SMPL
CCV81 - 1	CCV81	20979	4.8731	-7471A	PPB		13 Nov 2017 17:41:49	U	SMPL
CCB81 - 1	CCB81	-225	-0.0348	-7471A	PPB		13 Nov 2017 17:44:05	U	SMPL
6407-06LX5 - 1		590	0.1538	-7471A	PPB		13 Nov 2017 17:46:24	U	SMPL
6407-06A - 1		19974	4.6405	-7471A	PPB		13 Nov 2017 17:48:41	U	SMPL
6422-06LX5 - 1		492	0.1311	-7471A	PPB		13 Nov 2017 17:50:58	U	SMPL
6422-06A - 1		19556	4.5438	-7471A	PPB		13 Nov 2017 17:53:15	U	SMPL
6422-02DLX2 - 1		26519	6.1555	-7471A	PPB		13 Nov 2017 17:55:32	U	SMPL
6441-07DLX5 - 1		17260	4.0123	-7471A	PPB		13 Nov 2017 17:57:48	U	SMPL
6441-18DLX10 - 1		20606	4.7868	-7471A	PPB		13 Nov 2017 18:00:06	U	SMPL
CCV82 - 1	CCV82	20912	4.8576	-7471A	PPB		13 Nov 2017 18:02:23	U	SMPL
CCB82 - 1	CCB82	-192	-0.0272	-7471A	PPB		13 Nov 2017 18:04:41	U	SMPL