

Analytical Summary ReportAnalysis Method: 365.3ANALYST: AMANDEEPParameter: Phosphorus, TotalSUPERVISOR REVIEW BY: apatelRun Number: LB91549BALANCE ID: WC SC-4

Reagent/Standard	Lot/Log #
5N sulfuric acid	WP56414
calibration std. phosphate 1 ppm	WP59583
calibration std. phosphate 0.5 ppm	WP59582
calibration std. phosphate 0.3 ppm	WP59581
calibration std. 0 ppm	WP59578
phosphate ICV std.	WP59585
phosphate CCV std.	WP59584
calibration std. phosphate 0.1 ppm	WP59580
calibration std. phosphate 0.05 ppm	WP59579
Combined reagent	WP59587
11 N sulfuric acid	WP57045
Phenolphthalein indicator	WP57038
Sodium hydroxide, 1N	WP57036

Intercept: -0.0009Slope: 0.6504Regression: 0.999829

Seq	Lab ID	True Value (mg/L)	DF	Initial Volume (mL)	Final Volume (mL)	Absorbance Reading at 880nm	AnalDate	AnalTime
1	CAL1	0.00	1	50	50	0.000	11/13/2017	13:45
2	CAL2	0.05	1	50	50	0.033	11/13/2017	13:45
3	CAL3	0.10	1	50	50	0.065	11/13/2017	13:46
4	CAL4	0.30	1	50	50	0.186	11/13/2017	13:46
5	CAL5	0.50	1	50	50	0.330	11/13/2017	13:47
6	CAL6	1.00	1	50	50	0.649	11/13/2017	13:47



Analytical Summary Report

Analysis Method: 365.3

ANALYST: AMANDEEP

Parameter: Phosphorus, Total

SUPERVISOR REVIEW BY: apatel

Run Number: LB91549

BALANCE ID: WC SC-4

Seq	Lab ID	True Value (mg/l)	DF	Initial Volume (mL)	Final Volume (mL)	Absorbance Reading at 880nm	Result (mg/L)	AnalDate	AnalTime
1	ICV	0.50	1	50	50	0.320	0.493	11/13/2017	13:48
2	ICB		1	50	50	0.000	0.001	11/13/2017	13:48
3	CCV1	0.50	1	50	50	0.329	0.507	11/13/2017	13:49
4	CCB1		1	50	50	0.000	0.001	11/13/2017	13:49
5	PB104215BL		1	1.00	50	0.000	0.001	11/13/2017	13:50
6	PB104215BS	0.50	1	1.01	50	0.319	0.492	11/13/2017	13:50
7	I6298-11		1	1.00	50	3.098	4.765	11/13/2017	13:51
8	I6298-11		50	1.00	50	0.450	0.693	11/13/2017	13:51
9	CCV2	0.50	1	50	50	0.328	0.506	11/13/2017	13:52
10	CCB2		1	50	50	0.000	0.001	11/13/2017	13:52