



Analytical Summary Report

Analysis Method: 9040C

Analyst By : JIGNESH

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB91560

Temp.: 21.4 °C

SLOPE.: 98.7

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2270
buffer solution pH 7 yellow	W2159
PH 10.01 BUFFER,COLOR CD 475ML	W2154
buffer solution pH 7 yellow	W2219
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.01 Control Limits[+/- 0.1].

True Value of CCV2 = 12.01 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			4.01	11/14/2017	08:10
2	CAL2	1	Water			7.00	11/14/2017	08:11
3	CAL3	1	Water			10.03	11/14/2017	08:12
4	ICV	1	Water			6.99	11/14/2017	08:14
5	CCV1	1	Water			2.01	11/14/2017	08:16
6	I6469-02	1	Water			7.53	11/14/2017	08:18
7	I6469-02DUP	1	Water			7.54	11/14/2017	08:19
8	CCV2	1	Water			12.01	11/14/2017	08:20

2B91560

WORKLIST(Hardcopy Internal Chain)

WorkList Name : CORROSIVITY I6469 WorkList ID : 105445 Date : 11/14/2017 7:56:07 AM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
11/13/2017	Water	I6469-02 - 1A	Corrosivity	Cool 4 deg C	PSEG01	J52	SV30666L	11/13/2017	9040C

Date/Time 11-14-17 8:00 AM
Received by: Jn
Relinquished by: CB

Date/Time 11-14-17 9:00 AM
Received by: CB
Relinquished by: JD