



Analytical Summary Report

Analysis Method: SM 4500-PH B

Analyst By : JIGNESH

Parameter: pH

Supervisor Review By : _____

Run Number: LB91565

Temp.: 20.1 °C

SLOPE.: 98.2

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2270
buffer solution pH 7 yellow	W2159
PH 10.01 BUFFER,COLOR CD 475ML	W2154
buffer solution pH 7 yellow	W2219
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.01 Control Limits[+/- 0.1].

True Value of CCV2 = 12.01 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			4.01	11/14/2017	08:25
2	CAL2	1	Water			6.99	11/14/2017	08:26
3	CAL3	1	Water			10.02	11/14/2017	08:27
4	ICV	1	Water			7.00	11/14/2017	08:29
5	CCV1	1	Water			2.01	11/14/2017	08:30
6	I5274-20	1	Water			8.35	11/14/2017	08:31
7	CCV2	1	Water			12.01	11/14/2017	08:33