

LB91566


7470A

INSTRUMENT ID: CV1

Linear

 μ Abs.:

33985



A= 0.0000e+000

B= 2.9198e-004

C= -2.0450e-003

Rho= 0.9999085

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	-0.023	-0.023	-73	0.000	-73				
0.2	0.200	0.180	-0.020	624	0.0 %	624				
2.5	2.500	2.516	0.016	8625	0.0 %	8625				
5.0	5.000	5.022	0.022	17208	0.0 %	17208				
7.5	7.500	7.584	0.084	25980	0.0 %	25980				
10.0	10.000	9.921	-0.079	33985	0.0 %	33985				

LB91566 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 14 Nov 2017 10:16:33

Sample ID	Extended ID	μ Abs	Conc	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	-73	-	0.0000	7470A	PPB	14 Nov 2017 10:17:50	S	Std
0.2 - 1	50.2	624	-	0.2000	7470A	PPB	14 Nov 2017 10:19:59	S	Std
2.5 - 1	52.5	8625	-	2.5000	7470A	PPB	14 Nov 2017 10:22:07	S	Std
5.0 - 1	55	17208	-	5.0000	7470A	PPB	14 Nov 2017 10:24:15	S	Std
7.5 - 1	57.5	25980	-	7.5000	7470A	PPB	14 Nov 2017 10:26:23	S	Std
10.0 - 1	58	33985	-	10.0000	7470A	PPB	14 Nov 2017 10:28:31	S	Std
ICV61 - 1	ICV61	13005	3.7952	-7470A	PPB	14 Nov 2017 10:31:06	U	SMPL	
ICB61 - 1	ICB61	-194	-0.0587	-7470A	PPB	14 Nov 2017 10:33:12	U	SMPL	
CCV83 - 1	CCV83	17380	5.0726	-7470A	PPB	14 Nov 2017 10:35:19	U	SMPL	
CCB83 - 1	CCB83	-205	-0.0619	-7470A	PPB	14 Nov 2017 10:37:25	U	SMPL	
CRA - 1	CRA	635	0.1834	-7470A	PPB	14 Nov 2017 10:39:33	U	SMPL	
HighStd - 1	HighStd	34481	10.0658	-7470A	PPB	14 Nov 2017 10:41:39	U	SMPL	
ChkStd - 1	ChkStd	24015	7.0099	-7470A	PPB	14 Nov 2017 10:43:46	U	SMPL	
PB104221BL - 1	PBW	-262	-0.0785	-7470A	PPB	14 Nov 2017 10:45:55	U	SMPL	
PB104221BS - 1	LCSW	12444	3.6314	-7470A	PPB	14 Nov 2017 10:48:04	U	SMPL	
PB104221TB - 1	PB104178TB	-167	-0.0508	-7470A	PPB	14 Nov 2017 10:50:12	U	SMPL	
I6300-04 - 1	OR-03-111017	-78	-0.0248	-7470A	PPB	14 Nov 2017 10:52:20	U	SMPL	
I6307-02 - 1	EO-02-111017-A	-79	-0.0251	-7470A	PPB	14 Nov 2017 10:54:28	U	SMPL	
I6307-04 - 1	EO-02-111017-B	-78	-0.0248	-7470A	PPB	14 Nov 2017 10:56:36	U	SMPL	
I6307-06 - 1	EO-02-111017-C	-63	-0.0204	-7470A	PPB	14 Nov 2017 10:58:43	U	SMPL	
CCV84 - 1	CCV84	17216	5.0247	-7470A	PPB	14 Nov 2017 11:00:50	U	SMPL	
CCB84 - 1	CCB84	-191	-0.0578	-7470A	PPB	14 Nov 2017 11:02:56	U	SMPL	
I6311-04 - 1	TR-05-111017	-76	-0.0242	-7470A	PPB	14 Nov 2017 11:05:04	U	SMPL	
I6408-04 - 1	WASTE-CHARACTERIZATION	-64	-0.0207	-7470A	PPB	14 Nov 2017 11:07:10	U	SMPL	
I6423-03 - 1	B8(0-10)COMP	-92	-0.0289	-7470A	PPB	14 Nov 2017 11:09:17	U	SMPL	
I6423-03DUP - 1	B8(0-10)COMP DUP	-74	-0.0237	-7470A	PPB	14 Nov 2017 11:11:24	U	SMPL	
I6423-03MS - 1	B8(0-10)COMPMS	12379	3.6124	-7470A	PPB	14 Nov 2017 11:13:32	U	SMPL	
I6423-03MSD - 1	B8(0-10)COMPMSD	12854	3.7511	-7470A	PPB	14 Nov 2017 11:15:39	U	SMPL	
I6427-03 - 1	277-216-156-343-350-342	-188	-0.0569	-7470A	PPB	14 Nov 2017 11:17:48	U	SMPL	
I6429-02 - 1	SP-1	-86	-0.0272	-7470A	PPB	14 Nov 2017 11:19:56	U	SMPL	
I6436-02 - 1	153RD-63RD-ST-DRUM	-71	-0.0228	-7470A	PPB	14 Nov 2017 11:22:05	U	SMPL	
I6436-04 - 1	153RD-63RD-ST-DRUM	-75	-0.0239	-7470A	PPB	14 Nov 2017 11:24:12	U	SMPL	
CCV85 - 1	CCV85	17033	4.9713	-7470A	PPB	14 Nov 2017 11:26:18	U	SMPL	
CCB85 - 1	CCB85	-191	-0.0578	-7470A	PPB	14 Nov 2017 11:28:25	U	SMPL	
I6441-04 - 1	TP-11A-11B	-81	-0.0257	-7470A	PPB	14 Nov 2017 11:30:33	U	SMPL	
I6441-08 - 1	TP-10	-10	-0.0050	-7470A	PPB	14 Nov 2017 11:32:40	U	SMPL	
I6441-13 - 1	TP-9A-9B	-89	-0.0280	-7470A	PPB	14 Nov 2017 11:34:46	U	SMPL	
I6441-19 - 1	TP-8	-60	-0.0196	-7470A	PPB	14 Nov 2017 11:36:54	U	SMPL	
I6441-24 - 1	TP-7A-7B	-73	-0.0234	-7470A	PPB	14 Nov 2017 11:39:01	U	SMPL	
I6441-30 - 1	TP-16	18	0.0032	-7470A	PPB	14 Nov 2017 11:41:09	U	SMPL	
I6441-32 - 1	TP-8	-49	-0.0164	-7470A	PPB	14 Nov 2017 11:43:17	U	SMPL	
I6443-02 - 1	FADW4870L	-60	-0.0196	-7470A	PPB	14 Nov 2017 11:45:25	U	SMPL	
I6449-01 - 1	ADT-WC-01	-40	-0.0137	-7470A	PPB	14 Nov 2017 11:47:33	U	SMPL	
PB104222BL - 1	PBW	-63	-0.0204	-7470A	PPB	14 Nov 2017 11:49:40	U	SMPL	
CCV86 - 1	CCV86	17385	5.0740	-7470A	PPB	14 Nov 2017 11:51:48	U	SMPL	
CCB86 - 1	CCB86	-198	-0.0599	-7470A	PPB	14 Nov 2017 11:53:55	U	SMPL	
PB104222BS - 1	LCSW	12599	3.6766	-7470A	PPB	14 Nov 2017 11:56:03	U	SMPL	
I6436-03 - 1	153RD-63RD-ST-DRUM	-170	-0.0517	-7470A	PPB	14 Nov 2017 11:58:10	U	SMPL	
I6441-31 - 1	TP-8	218896	63.9114	-7470A	PPB	14 Nov 2017 12:00:18	U	SMPL	
I6443-01 - 1	FADW4870L	-99	-0.0310	-7470A	PPB	14 Nov 2017 12:05:53	U	SMPL	
I6443-01DUP - 1	FADW4870LDUP	-23	-0.0088	-7470A	PPB	14 Nov 2017 12:08:00	U	SMPL	
I6443-01MS - 1	FADW4870LMS	13295	3.8798	-7470A	PPB	14 Nov 2017 12:10:08	U	SMPL	
I6443-01MSD - 1	FADW4870LMSD	13655	3.9850	-7470A	PPB	14 Nov 2017 12:12:16	U	SMPL	
CCV87 - 1	CCV87	16875	4.9251	-7470A	PPB	14 Nov 2017 12:14:24	U	SMPL	
CCB87 - 1	CCB87	-218	-0.0657	-7470A	PPB	14 Nov 2017 12:16:33	U	SMPL	
6423-03LX5 - 1		-82	-0.0260	-7470A	PPB	14 Nov 2017 12:18:41	U	SMPL	
6423-03A - 1		13444	3.9233	-7470A	PPB	14 Nov 2017 12:20:48	U	SMPL	
6443-01LX5 - 1		-149	-0.0456	-7470A	PPB	14 Nov 2017 12:22:55	U	SMPL	
6443-01A - 1		14033	4.0953	-7470A	PPB	14 Nov 2017 12:25:03	U	SMPL	
6441-31DLX10 - 1		20184	5.8913	-7470A	PPB	14 Nov 2017 12:27:10	U	SMPL	
CCV88 - 1	CCV88	17336	5.0597	-7470A	PPB	14 Nov 2017 12:29:17	U	SMPL	
CCB88 - 1	CCB88	-225	-0.0677	-7470A	PPB	14 Nov 2017 12:31:26	U	SMPL	