



Analytical Summary Report

Analysis Method: SM5220 D

ANALYST: Indiana

Parameter: COD

SUPERVISOR REVIEW BY: apatel

Run Number: LB91573

Reagent/Standard	Lot/Log #
COD ICV-LCS std, 50ppm	WP59085
COD calibration std. 100 ppm	WP59082
COD calibration std. 10 ppm	WP59081
COD calibration std. 150 ppm	WP59083
COD calibration std. 50 ppm	WP59084
COD calibration std. 0 ppm	WP59087
COD ICV-LCS std, 50ppm	WP59559
COD CCV std, 50ppm	WP59560
COD Digestion Vials Low Level 0-150Mg/L	W2291

Temp In (C): <u>150</u>	Date In: <u>11/14/2017</u>	Time In: <u>10:00</u>
Temp Out (C): <u>150</u>	Date Out: <u>11/14/2017</u>	Time Out: <u>12:05</u>

Intercept: 0.3627

Slope: 1.0071

Regression: 0.9999

Seq	Lab ID	TrueValue (mg/l)	DF	MATRIX	Reading	Anal Date	Anal Time
1	CAL1	0	1	Water	0.000	10/26/2017	11:10
2	CAL2	10	1	Water	11.000	10/26/2017	11:10
3	CAL3	50	1	Water	50.000	10/26/2017	11:11
4	CAL4	100	1	Water	102.000	10/26/2017	11:11
5	CAL5	150	1	Water	151.000	10/26/2017	11:12



Analytical Summary Report

Analysis Method: SM5220 D

ANALYST: Indiana

Parameter: COD

SUPERVISOR REVIEW BY: apatel

Run Number: LB91573

Seq	Lab ID	True Value (mg/l)	Initial Weight (g)	Final Vol (ml)	DF	MATRIX	Reading	Result	AnalDate	AnalTime
1	ICV	50			1	Water	50.000	49.287	10/26/2017	11:12
2	ICB				1	Water	0.000	-0.360	10/26/2017	11:13
3	CCV1	50			1	Water	51.000	50.280	11/14/2017	13:00
4	CCB1				1	Water	0.000	-0.360	11/14/2017	13:00
5	LB91573BL				1	Water	0.000	-0.360	11/14/2017	13:01
6	LB91573BS	50			1	Water	51.000	50.280	11/14/2017	13:01
7	I6436-03				1	Water	17.000	16.520	11/14/2017	13:02
8	I6436-03DUP				1	Water	17.000	16.520	11/14/2017	13:02
9	I6436-03MS	50			1	Water	68.000	67.160	11/14/2017	13:03
10	I6436-03MSD	50			1	Water	67.000	66.168	11/14/2017	13:03
11	I6441-31				10	Water	59.000	58.224	11/14/2017	13:04
12	I6443-01				1	Water	46.000	45.316	11/14/2017	13:04
13	CCV2	50			1	Water	50.000	49.287	11/14/2017	13:05
14	CCB2				1	Water	0.000	-0.360	11/14/2017	13:05