

245.1


LB91574

INSTRUMENT ID: CV1

Linear

 μ Abs.:

34082



A= 0.0000e+000

B= 2.9143e-004

C= 2.2736e-002

Rho= 0.9998058

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.004	0.004	-64	0.000	-64				
0.2	0.200	0.217	0.017	668	0.0 %	668				
2.5	2.500	2.510	0.010	8533	0.0 %	8533				
5.0	5.000	4.886	-0.114	16687	0.0 %	16687				
7.5	7.500	7.628	0.128	26095	0.0 %	26095				
10.0	10.000	9.955	-0.045	34082	0.0 %	34082				

LB91574 INSTRUMENT ID : CV1

Method: 245.1

Operator: Admin

Date of Analysis: 14 Nov 2017 13:18:23

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	-64	-	0.0000	245.1	PPB	14 Nov 2017 14:07:39	S	Std
0.2 - 1	50.2	668	-	0.2000	245.1	PPB	14 Nov 2017 14:09:48	S	Std
2.5 - 1	52.5	8533	-	2.5000	245.1	PPB	14 Nov 2017 14:11:56	S	Std
5.0 - 1	55	16687	-	5.0000	245.1	PPB	14 Nov 2017 14:14:04	S	Std
7.5 - 1	57.5	26095	-	7.5000	245.1	PPB	14 Nov 2017 14:16:13	S	Std
10.0 - 1	610	34082	-	10.0000	245.1	PPB	14 Nov 2017 14:18:21	S	Std
ICV62 - 1	ICV62	13519	3.9626	-245.1	PPB	PPB	14 Nov 2017 14:20:59	U	SMPL
ICB62 - 1	ICB62	-179	-0.0294	-245.1	PPB	PPB	14 Nov 2017 14:23:05	U	SMPL
CCV89 - 1	CCV89	17284	5.0599	-245.1	PPB	PPB	14 Nov 2017 14:25:12	U	SMPL
CCB89 - 1	CCB89	-188	-0.0321	-245.1	PPB	PPB	14 Nov 2017 14:27:18	U	SMPL
CRA - 1	CRA	637	0.2084	-245.1	PPB	PPB	14 Nov 2017 14:29:26	U	SMPL
HighStd - 1	HighStd	34358	10.0358	-245.1	PPB	PPB	14 Nov 2017 14:31:32	U	SMPL
ChkStd - 1	ChkStd	23378	6.8359	-245.1	PPB	PPB	14 Nov 2017 14:33:39	U	SMPL
PB104223BL - 1	PBW	-256	-0.0519	-245.1	PPB	PPB	14 Nov 2017 14:35:48	U	SMPL
PB104223BS - 1	LCSW	13528	3.9653	-245.1	PPB	PPB	14 Nov 2017 14:37:57	U	SMPL
I6426-01 - 1	110917-TIDO-E-D-B	-175	-0.0283	-245.1	PPB	PPB	14 Nov 2017 14:40:04	U	SMPL
I6426-02 - 1	110917-TIDO-E-D-S	-76	0.0006	-245.1	PPB	PPB	14 Nov 2017 14:42:12	U	SMPL
I6426-03 - 1	110917-TIDO-E-U-B	-75	0.0009	-245.1	PPB	PPB	14 Nov 2017 14:44:21	U	SMPL
I6426-04 - 1	110917-TIDO-E-U-S	-60	0.0053	-245.1	PPB	PPB	14 Nov 2017 14:46:30	U	SMPL
I6426-05 - 1	110917-SIDI-E-D-B	-67	0.0032	-245.1	PPB	PPB	14 Nov 2017 14:48:36	U	SMPL
CCV90 - 1	CCV90	16875	4.9407	-245.1	PPB	PPB	14 Nov 2017 14:50:43	U	SMPL
CCB90 - 1	CCB90	-181	-0.0300	-245.1	PPB	PPB	14 Nov 2017 14:52:49	U	SMPL
I6426-06 - 1	110917-SIDI-E-D-S	-79	-0.0003	-245.1	PPB	PPB	14 Nov 2017 14:54:57	U	SMPL
I6426-07 - 1	110917-SIDI-E-U-B	-65	0.0038	-245.1	PPB	PPB	14 Nov 2017 14:57:03	U	SMPL
I6426-08 - 1	110917-SIDI-E-U-S	-70	0.0023	-245.1	PPB	PPB	14 Nov 2017 14:59:10	U	SMPL
I6426-09 - 1	110917-JETT-F-D-B	-64	0.0041	-245.1	PPB	PPB	14 Nov 2017 15:01:17	U	SMPL
I6426-10 - 1	110917-JETT-F-D-S	-48	0.0087	-245.1	PPB	PPB	14 Nov 2017 15:03:25	U	SMPL
I6426-11 - 1	110917-JETT-F-U-B	-52	0.0076	-245.1	PPB	PPB	14 Nov 2017 15:05:32	U	SMPL
I6426-12 - 1	110917-JETT-F-U-S	-63	0.0044	-245.1	PPB	PPB	14 Nov 2017 15:07:40	U	SMPL
I6426-12DUP - 1	110917-JETT-F-U-SDUP	-48	0.0087	-245.1	PPB	PPB	14 Nov 2017 15:09:49	U	SMPL
I6426-12MS - 1	110917-JETT-F-U-SMS	15496	4.5388	-245.1	PPB	PPB	14 Nov 2017 15:11:57	U	SMPL
I6426-12MSD - 1	110917-JETT-F-U-SMSD	14948	4.3791	-245.1	PPB	PPB	14 Nov 2017 15:14:04	U	SMPL
CCV91 - 1	CCV91	17151	5.0211	-245.1	PPB	PPB	14 Nov 2017 15:16:11	U	SMPL
CCB91 - 1	CCB91	-201	-0.0358	-245.1	PPB	PPB	14 Nov 2017 15:18:18	U	SMPL
PB104257BL - 1	PBW	-72	0.0018	-245.1	PPB	PPB	14 Nov 2017 15:20:26	U	SMPL
PB104257BS - 1	LCSW	13406	3.9297	-245.1	PPB	PPB	14 Nov 2017 15:24:59	U	SMPL
I6402-01 - 1	110717-TIDO-F-D-B	-146	-0.0198	-245.1	PPB	PPB	14 Nov 2017 15:27:05	U	SMPL
I6402-02 - 1	110717-TIDO-F-D-S	-65	0.0038	-245.1	PPB	PPB	14 Nov 2017 15:29:13	U	SMPL
I6402-03 - 1	110717-TIDO-F-U-B	-64	0.0041	-245.1	PPB	PPB	14 Nov 2017 15:31:20	U	SMPL
I6402-04 - 1	110717-TIDO-F-U-S	-63	0.0044	-245.1	PPB	PPB	14 Nov 2017 15:33:28	U	SMPL
I6402-05 - 1	110717-SIDI-F-D-B	-60	0.0053	-245.1	PPB	PPB	14 Nov 2017 15:35:36	U	SMPL
I6402-06 - 1	110717-SIDI-F-D-S	-39	0.0114	-245.1	PPB	PPB	14 Nov 2017 15:37:45	U	SMPL
I6402-07 - 1	110717-SIDI-F-U-B	-60	0.0053	-245.1	PPB	PPB	14 Nov 2017 15:39:53	U	SMPL
I6402-08 - 1	110717-SIDI-F-U-S	-62	0.0047	-245.1	PPB	PPB	14 Nov 2017 15:42:00	U	SMPL
CCV92 - 1	CCV92	16959	4.9652	-245.1	PPB	PPB	14 Nov 2017 15:44:08	U	SMPL
CCB92 - 1	CCB92	-173	-0.0277	-245.1	PPB	PPB	14 Nov 2017 15:46:15	U	SMPL
6402-09 - 1	110717-JETT-E-D-B	-62	0.0047	-245.1	PPB	PPB	14 Nov 2017 15:48:23	U	SMPL
6402-10 - 1	110717-JETT-E-D-S	-67	0.0032	-245.1	PPB	PPB	14 Nov 2017 15:50:30	U	SMPL
6402-11 - 1	110717-JETT-E-U-B	-48	0.0087	-245.1	PPB	PPB	14 Nov 2017 15:52:37	U	SMPL
6402-12 - 1	110717-JETT-E-U-S	-32	0.0134	-245.1	PPB	PPB	14 Nov 2017 15:54:45	U	SMPL
6403-06 - 1	110817-TIDO-F-D-S	-55	0.0067	-245.1	PPB	PPB	14 Nov 2017 15:56:52	U	SMPL
6403-07 - 1	110817-TIDO-F-U-B	-64	0.0041	-245.1	PPB	PPB	14 Nov 2017 15:59:00	U	SMPL
6403-08 - 1	110817-TIDO-F-U-S	-50	0.0082	-245.1	PPB	PPB	14 Nov 2017 16:01:08	U	SMPL
6403-12 - 1	110817-SIDI-F-U-B	-9	0.0201	-245.1	PPB	PPB	14 Nov 2017 16:03:17	U	SMPL
6403-13 - 1	110817-SIDI-F-U-M	-48	0.0087	-245.1	PPB	PPB	14 Nov 2017 16:05:25	U	SMPL
CCV93 - 1	CCV93	16798	4.9182	-245.1	PPB	PPB	14 Nov 2017 16:07:32	U	SMPL
CCB93 - 1	CCB93	-170	-0.0268	-245.1	PPB	PPB	14 Nov 2017 16:09:39	U	SMPL
6426-12LX5 - 1		-75	0.0009	-245.1	PPB	PPB	14 Nov 2017 16:11:47	U	SMPL
6426-12A - 1		15271	4.4732	-245.1	PPB	PPB	14 Nov 2017 16:13:54	U	SMPL
CCV94 - 1	CCV94	17130	5.0150	-245.1	PPB	PPB	14 Nov 2017 16:16:01	U	SMPL
CCB94 - 1	CCB94	-187	-0.0318	-245.1	PPB	PPB	14 Nov 2017 16:18:09	U	SMPL