

===== LB91577 =====
Test results

Aquakem 7.2AQ1

Page: 1

CHEMTECH
284 Sheffield Street,
Mountainside, NJ 07092
Reviewed by : AS11/14/2017 14:50
-----Test: ^{Reactive}~~Total~~ CN Mg/L

Sample Id	Result	Dil. 1 +	Response	Errors
ICV1	98.439	0.0	0.099	
ICB1	-1.120	0.0	0.004	
CCV1	246.947	0.0	0.242	
CCB1	-0.942	0.0	0.004	
PB104265BL	-1.080	0.0	0.004	
I6302-02	-0.921	0.0	0.004	
I6302-02DUP	-1.165	0.0	0.003	
I6302-04	-1.188	0.0	0.003	
I6302-06	-1.142	0.0	0.004	
I6305-04	0.138	0.0	0.005	
I6305-06	-1.296	0.0	0.003	
I6305-08	-1.526	0.0	0.003	
I6305-10	-1.130	0.0	0.004	
I6311-04	-1.108	0.0	0.004	
CCV2	248.529	0.0	0.243	
CCB2	-1.094	0.0	0.004	
PB104266BL	-1.307	0.0	0.003	
I6469-02	-1.214	0.0	0.003	
I6469-02DUP	-0.927	0.0	0.004	
CCV3	259.689	0.0	0.254	
CCB3	-1.029	0.0	0.004	

N	21
Mean	39.788
SD	91.2836
CV%	229.42

AS

Aquakem v. 7.2AQ1

Results from time period:

Tue Nov 14 09:06:24 2017

Tue Nov 14 14:48:18 2017

Sample Id	Sam/Ctr/c/	Test short name	Test type	Result	Result unit	Result date and time	Stat
0.OPPBCN	A	Reactive CN	P	-1.2805	µg/l	11/14/2017 9:06:24	
5.OPPBCN	A	Reactive CN	P	3.9377	µg/l	11/14/2017 9:06:25	
10PPBCN	A	Reactive CN	P	8.4046	µg/l	11/14/2017 9:06:26	
50PPBCN	A	Reactive CN	P	50.3361	µg/l	11/14/2017 9:06:27	
100PPBCN	A	Reactive CN	P	102.075	µg/l	11/14/2017 9:06:28	
250PPBCN	A	Reactive CN	P	253.8664	µg/l	11/14/2017 9:06:29	
500PPBCN	A	Reactive CN	P	497.6607	µg/l	11/14/2017 9:06:30	
ICV1	S	Reactive CN	P	98.4389	µg/l	11/14/2017 14:36:24	
ICB1	S	Reactive CN	P	-1.1205	µg/l	11/14/2017 14:36:25	
CCV1	S	Reactive CN	P	246.9469	µg/l	11/14/2017 14:36:28	
CCB1	S	Reactive CN	P	-0.942	µg/l	11/14/2017 14:36:29	
PB104265BL	S	Reactive CN	P	-1.08	µg/l	11/14/2017 14:36:31	
I6302-02	S	Reactive CN	P	-0.9214	µg/l	11/14/2017 14:36:32	
I6302-02DUP	S	Reactive CN	P	-1.1652	µg/l	11/14/2017 14:36:33	
I6302-04	S	Reactive CN	P	-1.1881	µg/l	11/14/2017 14:43:56	
I6302-06	S	Reactive CN	P	-1.1418	µg/l	11/14/2017 14:43:57	
I6305-04	S	Reactive CN	P	0.138	µg/l	11/14/2017 14:43:58	
I6305-06	S	Reactive CN	P	-1.2964	µg/l	11/14/2017 14:43:59	
I6305-08	S	Reactive CN	P	-1.5258	µg/l	11/14/2017 14:44:00	
I6305-10	S	Reactive CN	P	-1.1303	µg/l	11/14/2017 14:44:01	
I6311-04	S	Reactive CN	P	-1.1078	µg/l	11/14/2017 14:44:02	
CCV2	S	Reactive CN	P	248.5285	µg/l	11/14/2017 14:44:03	
CCB2	S	Reactive CN	P	-1.0941	µg/l	11/14/2017 14:44:04	
PB104266BL	S	Reactive CN	P	-1.3071	µg/l	11/14/2017 14:44:05	
I6469-02	S	Reactive CN	P	-1.2137	µg/l	11/14/2017 14:44:06	
I6469-02DUP	S	Reactive CN	P	-0.927	µg/l	11/14/2017 14:48:15	
CCV3	S	Reactive CN	P	259.6891	µg/l	11/14/2017 14:48:16	
CCB3	S	Reactive CN	P	-1.0293	µg/l	11/14/2017 14:48:17	

CHEMTECH
284 Sheffield Street,
Mountainside, NJ 07092
Reviewed by : AS

11/14/2017 9:10


Test Total CN

Accepted 11/14/2017 9:10

Factor 1042
Bias 0.005

Coeff. of det. 0.999853

Errors



	Calibrator	Response	Calc. con.	Conc.	Errors
1	0.0PPBCN	0.003	-1.2805	0.0000	
2	5.0PPBCN	0.008	3.9377	5.0000	
3	10PPBCN	0.013	8.4046	10.0000	
4	50PPBCN	0.053	50.3361	50.0000	
5	100PPBCN	0.103	102.0750	100.0000	
6	250PPBCN	0.248	253.8664	250.0000	
7	500PPBCN	0.482	497.6607	500.0000	