



Analytical Summary Report

Analysis Method: 1010A

Reviewed By: HIRAL

Parameter: Flash Point

Supervisor Review By: apatel

Run Number: LB91584

Ambient Barometric Pressure (mmHg): 760

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2168

Seq	LabID	True Value °F	Dilution	Celsius °C	Result °F	Anal Date	Anal Time
1	ICV	81	1	27.0	80.6	11/14/2017	09:20
2	I6446-01		1	>100.0	>212.0	11/14/2017	09:50
3	I6446-01DUP		1	>100.0	>212.0	11/14/2017	10:15
4	I6446-02		1	>100.0	>212.0	11/14/2017	10:32
5	I6446-03		1	>100.0	>212.0	11/14/2017	10:50
6	I6446-04		1	>100.0	>212.0	11/14/2017	11:15
7	I6446-05		1	>100.0	>212.0	11/14/2017	11:35
8	I6446-06		1	>100.0	>212.0	11/14/2017	11:50
9	I6446-07		1	>100.0	>212.0	11/14/2017	12:10
10	I6446-08		1	>100.0	>212.0	11/14/2017	12:35
11	I6446-09		1	>100.0	>212.0	11/14/2017	12:50
12	I6446-10		1	>100.0	>212.0	11/14/2017	13:15
13	I6446-11		1	>100.0	>212.0	11/14/2017	13:35
14	I6446-12		1	>100.0	>212.0	11/14/2017	13:52
15	I6446-13		1	>100.0	>212.0	11/14/2017	14:05
16	I6446-14		1	>100.0	>212.0	11/14/2017	14:25
17	I6446-15		1	>100.0	>212.0	11/14/2017	14:40
18	I6446-16		1	>100.0	>212.0	11/14/2017	14:52
19	I6446-17		1	>100.0	>212.0	11/14/2017	15:10
20	I6446-18		1	>100.0	>212.0	11/14/2017	15:32
21	I6446-19		1	>100.0	>212.0	11/14/2017	15:45
22	I6446-20		1	>100.0	>212.0	11/14/2017	16:05

Result = (Celsius * 1.8) + 32

WORKLIST(Hardcopy Internal Chain)

WorkList Name : FP111417

WorkList ID : 105461

Date : 11/14/2017 8:51:21 AM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
11/15/2017	Water	✓ 16446-01	Flash Point	Cool 4 deg C	THEP04	J62	D-1	11/10/2017	1010A
11/15/2017	Water	✓ 16446-02	Flash Point	Cool 4 deg C	THEP04	J62	D-2	11/10/2017	1010A
11/15/2017	Water	✓ 16446-03	Flash Point	Cool 4 deg C	THEP04	J62	D-3	11/10/2017	1010A
11/15/2017	Water	✓ 16446-04	Flash Point	Cool 4 deg C	THEP04	J62	D-4	11/10/2017	1010A
11/15/2017	Water	✓ 16446-05	Flash Point	Cool 4 deg C	THEP04	J62	D-5	11/10/2017	1010A
11/15/2017	Water	✓ 16446-06	Flash Point	Cool 4 deg C	THEP04	J62	D-6	11/10/2017	1010A
11/15/2017	Water	✓ 16446-07	Flash Point	Cool 4 deg C	THEP04	J62	D-8	11/10/2017	1010A
11/15/2017	Water	✓ 16446-08	Flash Point	Cool 4 deg C	THEP04	J62	D-9	11/10/2017	1010A
11/15/2017	Water	✓ 16446-09	Flash Point	Cool 4 deg C	THEP04	J62	D-10	11/10/2017	1010A
11/15/2017	Water	✓ 16446-10	Flash Point	Cool 4 deg C	THEP04	J62	D-13	11/10/2017	1010A
11/15/2017	Water	✓ 16446-11	Flash Point	Cool 4 deg C	THEP04	J62	D-14	11/10/2017	1010A
11/15/2017	Water	✓ 16446-12	Flash Point	Cool 4 deg C	THEP04	J62	D-15	11/10/2017	1010A
11/15/2017	Water	✓ 16446-13	Flash Point	Cool 4 deg C	THEP04	J62	D-16	11/10/2017	1010A
11/15/2017	Water	✓ 16446-14	Flash Point	Cool 4 deg C	THEP04	J62	D-18	11/10/2017	1010A
11/15/2017	Water	✓ 16446-15	Flash Point	Cool 4 deg C	THEP04	J62	D-19	11/10/2017	1010A
11/15/2017	Water	✓ 16446-16	Flash Point	Cool 4 deg C	THEP04	J62	D-20	11/10/2017	1010A
11/15/2017	Water	✓ 16446-17	Flash Point	Cool 4 deg C	THEP04	J62	D-21	11/10/2017	1010A
11/15/2017	Water	✓ 16446-18	Flash Point	Cool 4 deg C	THEP04	J62	D-23	11/10/2017	1010A
11/15/2017	Water	✓ 16446-19	Flash Point	Cool 4 deg C	THEP04	J62	D-25	11/10/2017	1010A
11/15/2017	Water	✓ 16446-20	Flash Point	Cool 4 deg C	THEP04	J62	D-26	11/10/2017	1010A

Date/Time

Received by:

Relinquished by:

Date/Time

Received by:

Relinquished by: