



Analytical Summary Report

Analysis Method: 1010A

Reviewed By: ketankumar

Parameter: Flash Point

Supervisor Review By: apatel

Run Number: LB91585

Ambient Barometric Pressure (mmHg): 760

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2168

Seq	LabID	True Value °F	Dilution	Celsius °C	Result °F	Anal Date	Anal Time
1	ICV	81	1	27.0	80.6	11/14/2017	16:15
2	I6436-04		1	>100.0	>212.0	11/14/2017	16:32
3	I6436-04DUP		1	>100.0	>212.0	11/14/2017	16:45
4	I6441-32		1	>100.0	>212.0	11/14/2017	16:58
5	I6443-02		1	>100.0	>212.0	11/14/2017	17:15
6	I6447-01		1	>100.0	>212.0	11/14/2017	17:32
7	I6447-02		1	>100.0	>212.0	11/14/2017	17:46
8	I6447-03		1	>100.0	>212.0	11/14/2017	17:58
9	I6447-04		1	>100.0	>212.0	11/14/2017	18:15
10	I6447-05		1	>100.0	>212.0	11/14/2017	18:35
11	I6447-06		1	>100.0	>212.0	11/14/2017	18:48
12	I6447-07		1	>100.0	>212.0	11/14/2017	18:58
13	I6469-02		1	>100.0	>212.0	11/14/2017	19:05

Result = (Celsius * 1.8) + 32

WORKLIST(Hardcopy Internal Chain)

WorkList Name : FP111417-1

WorkList ID : 105463

Date : 11/14/2017 8:53:59 AM

LB91585

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
11/15/2017	Water	✓ 16436-04 C	Flash Point	Cool 4 deg C	PSEG01	H52	153RD-63RD-ST-DRUM	11/10/2017	1010A
11/15/2017	Water	✓ 16441-32 C	Flash Point	Cool 4 deg C	PSEG01	H52	TP-8	11/10/2017	1010A
11/15/2017	Water	✓ 16443-02 C	Flash Point	Cool 4 deg C	PSEG01	J52	FADW4870L	11/10/2017	1010A
11/15/2017	Water	✓ 16447-01	Flash Point	Cool 4 deg C	THEP04	J62	D-27	11/10/2017	1010A
11/15/2017	Water	✓ 16447-02	Flash Point	Cool 4 deg C	THEP04	J62	D-28	11/10/2017	1010A
11/15/2017	Water	✓ 16447-03	Flash Point	Cool 4 deg C	THEP04	J62	D-29	11/10/2017	1010A
11/15/2017	Water	✓ 16447-04	Flash Point	Cool 4 deg C	THEP04	J62	D-30	11/10/2017	1010A
11/15/2017	Water	✓ 16447-05	Flash Point	Cool 4 deg C	THEP04	J62	D-31	11/10/2017	1010A
11/15/2017	Water	✓ 16447-06	Flash Point	Cool 4 deg C	THEP04	J62	D-32	11/10/2017	1010A
11/15/2017	Water	✓ 16447-07	Flash Point	Cool 4 deg C	THEP04	J62	D-34	11/10/2017	1010A
11/16/2017	Water	✓ 16469-02 C	Flash Point	Cool 4 deg C	PSEG01	J52	SV30666L	11/13/2017	1010A

Date/Time 11/14/17 15:30

Received by: RS

Relinquished by: CP

Date/Time

Received by: CP

Relinquished by: RS

11/14/17 19:10