



Analytical Summary Report

Analysis Method: 9045C

Analyst By : Jignesh

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB91586

Temp.: 24.1 °C

BalanceID: WC SC-4

SLOPE.: 98.8

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2270
buffer solution pH 7 yellow	W2159
PH 10.01 BUFFER,COLOR CD 475ML	W2154
buffer solution pH 7 yellow	W2219
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.01 Control Limits[+/- 0.1].

True Value of CCV2 = 2.01 Control Limits[+/- 0.1].

True Value of CCV3 = 12.02 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			4.01	11/14/2017	16:30
2	CAL2	1	Water			7.01	11/14/2017	16:31
3	CAL3	1	Water			10.04	11/14/2017	16:33
4	ICV	1	Water			7.00	11/14/2017	16:35
5	CCV1	1	Water			2.01	11/14/2017	16:36
6	I6289-06	1	Solid	20.02	20	7.76	11/14/2017	16:37
7	I6289-06DUP	1	Solid	20.03	20	7.77	11/14/2017	16:38
8	I6476-02	1	Solid	20.02	20	5.17	11/14/2017	16:39
9	I6476-04	1	Solid	20.04	20	8.46	11/14/2017	16:40
10	I6476-06	1	Solid	20.04	20	8.53	11/14/2017	16:41
11	I6478-02	1	Solid	20.03	20	7.34	11/14/2017	16:42
12	I6478-04	1	Solid	20.04	20	6.99	11/14/2017	16:43
13	I6478-06	1	Solid	20.06	20	7.07	11/14/2017	16:45
14	I6478-08	1	Solid	20.03	20	6.97	11/14/2017	16:46
15	I6479-02	1	Solid	20.01	20	5.96	11/14/2017	16:47
16	CCV2	1	Water			2.01	11/14/2017	16:48
17	I6479-04	1	Solid	20.03	20	6.66	11/14/2017	16:49
18	I6481-03	1	Solid	20.01	20	5.85	11/14/2017	16:51
19	CCV3	1	Water			12.02	11/14/2017	16:55

WORKLIST(Hardcopy Internal Chain)

2391586

WorkList Name : CORROSIONITY I6481

WorkList ID : 105498

Date : 11/14/2017 3:51:51 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
11/17/2017	Solid	I6289-06	Corrosivity	Cool 4 deg C	PSEG05	J52	SU-01-111417	11/14/2017	9045C
11/17/2017	Solid	I6476-02	Corrosivity	Cool 4 deg C	PSEG01	J52	72-11934-290	11/14/2017	9045C
11/17/2017	Solid	I6476-04	Corrosivity	Cool 4 deg C	PSEG01	J52	SP-1-RIGHT	11/14/2017	9045C
11/17/2017	Solid	I6476-06	Corrosivity	Cool 4 deg C	PSEG01	J52	SP-1-LEFT	11/14/2017	9045C
11/21/2017	Solid	I6478-02	Corrosivity	Cool 4 deg C	PSEG01	J52	SP-1-A	11/14/2017	9045C
11/21/2017	Solid	I6478-04	Corrosivity	Cool 4 deg C	PSEG01	J52	SP-1-B	11/14/2017	9045C
11/21/2017	Solid	I6478-06	Corrosivity	Cool 4 deg C	PSEG01	J52	SP-1-C	11/14/2017	9045C
11/21/2017	Solid	I6478-08	Corrosivity	Cool 4 deg C	PSEG01	J52	SP-1-D	11/14/2017	9045C
11/17/2017	Solid	I6479-02	Corrosivity	Cool 4 deg C	PSEG01	J52	T-20	11/14/2017	9045C
11/17/2017	Solid	I6479-04	Corrosivity	Cool 4 deg C	PSEG01	J52	POT-HEAD	11/14/2017	9045C
11/17/2017	Solid	I6481-03	Corrosivity	Cool 4 deg C	SUSS01	J21	SCMUA-SLUDGE-CAKE	11/14/2017	9045C

Date/Time 11-14-17 4:00 PM
 Received by: JP
 Relinquished by: JP

Date/Time 11-14-17 6:30 PM
 Received by: JP
 Relinquished by: JP