



Analytical Summary Report

Analysis Method: SM 4500-PH B

Analyst By : JIGNESH

Parameter: pH

Supervisor Review By : apatel

Run Number: LB91588

Temp.: 20.1 °C

SLOPE.: 98.2

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2270
buffer solution pH 7 yellow	W2159
PH 10.01 BUFFER,COLOR CD 475ML	W2154
buffer solution pH 7 yellow	W2219
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.01 Control Limits[+/- 0.1].

True Value of CCV2 = 12.01 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			4.01	11/14/2017	18:15
2	CAL2	1	Water			6.99	11/14/2017	18:16
3	CAL3	1	Water			10.02	11/14/2017	18:19
4	ICV	1	Water			7.00	11/14/2017	18:20
5	CCV1	1	Water			2.01	11/14/2017	18:21
6	I6477-01	1	Water			11.84	11/14/2017	18:22
7	I6477-01DUP	1	Water			11.85	11/14/2017	18:24
8	CCV2	1	Water			12.01	11/14/2017	18:25

201588

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH I6477 WorkList ID : 105502 Date : 11/14/2017 6:05:33 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
11/14/2017	Water	I6477-01-2	pH	Cool 4 deg C	PSEG01	J52	TOTE-1-2	11/14/2017	SM 4500-PH B

Date/Time 11-14-17 6:00 PM
Received by: JP
Relinquished by: JP

Date/Time 11-14-17 7:00 PM
Received by: JP
Relinquished by: JP