



Analytical Summary Report

Analysis Method: 1010A

Reviewed By: HIRAL

Parameter: Flash Point

Supervisor Review By: apatel

Run Number: LB91599

Ambient Barometric Pressure (mmHg): 760

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2168

Seq	LabID	True Value °F	Dilution	Celsius °C	Result °F	Anal Date	Anal Time
1	ICV	81	1	28.0	82.4	11/15/2017	10:03
2	I6477-02		1	>100.0	>212.0	11/15/2017	08:45
3	I6477-02DUP		1	>100.0	>212.0	11/15/2017	09:20

Result = (Celsius * 1.8) + 32

WORKLIST(Hardcopy Internal Chain)

WorkList Name : FP111517

WorkList ID : 105508

Date : 11/15/2017 8:05:26 AM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
11/17/2017	Water	I6477-02	C Flash Point	Cool 4 deg C	PSEG01	J52	TOTE-1-2	11/14/2017	1010A

Date/Time

Received by:

Relinquished by:

11/15/17 8:15 AM
HP
CP

Date/Time

Received by:

Relinquished by:

11/15/17 09:25 AM
CP
HP

LB 91599