


7470A

LB91601
INSTRUMENT ID: CVJ

Linear

 μ Abs.:

34646



A= 0.0000e+000

B= 2.8317e-004

C= 4.0860e-002

Rho= 0.9995595

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	-0.007	-0.007	-169	0.000	-169				
0.2	0.200	0.192	-0.008	535	0.0 %	535				
2.5	2.500	2.480	-0.020	8614	0.0 %	8614				
5.0	5.000	4.964	-0.036	17384	0.0 %	17384				
7.5	7.500	7.719	0.219	27116	0.0 %	27116				
10.0	10.000	9.852	-0.148	34646	0.0 %	34646				

LB91601 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 15 Nov 2017 10:38:39

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	-169	-	0.0000	7470A	PPB	15 Nov 2017 11:11:43	S	Std
0.2 - 1	50.2	535	-	0.2000	7470A	PPB	15 Nov 2017 11:13:51	S	Std
2.5 - 1	52.5	8614	-	2.5000	7470A	PPB	15 Nov 2017 11:16:00	S	Std
5.0 - 1	55	17384	-	5.0000	7470A	PPB	15 Nov 2017 11:18:08	S	Std
7.5 - 1	57.5	27116	-	7.5000	7470A	PPB	15 Nov 2017 11:20:16	S	Std
10.0 - 1	510	34646	-	10.0000	7470A	PPB	15 Nov 2017 11:22:24	S	Std
ICV64 - 1	ICV64	13006	3.7238	-7470A	PPB	15 Nov 2017 11:27:32	U	SMPL	
ICB64 - 1	ICB64	-285	-0.0398	-7470A	PPB	15 Nov 2017 11:29:38	U	SMPL	
CCV08 - 1	CCV08	17931	5.1184	-7470A	PPB	15 Nov 2017 11:31:44	U	SMPL	
CCB08 - 1	CCB08	-315	-0.0483	-7470A	PPB	15 Nov 2017 11:33:51	U	SMPL	
CRA - 1	CRA	537	0.1929	-7470A	PPB	15 Nov 2017 11:35:59	U	SMPL	
HighStd - 1	HighStd	35312	10.0402	-7470A	PPB	15 Nov 2017 11:38:05	U	SMPL	
ChkStd - 1	ChkStd	25223	7.1833	-7470A	PPB	15 Nov 2017 11:40:12	U	SMPL	
PB104258BL - 1	PBW	-138	0.0018	-7470A	PPB	15 Nov 2017 11:42:21	U	SMPL	
PB104258BS - 1	LCSW	12887	3.6901	-7470A	PPB	15 Nov 2017 11:44:30	U	SMPL	
PB104258TB - 1	PB104230TB	25	0.0479	-7470A	PPB	15 Nov 2017 11:46:37	U	SMPL	
I6302-02 - 1	HR-04-111317-A	34	0.0505	-7470A	PPB	15 Nov 2017 11:48:45	U	SMPL	
I6302-04 - 1	HR-04-111317-B	15	0.0451	-7470A	PPB	15 Nov 2017 11:50:54	U	SMPL	
I6302-06 - 1	HR-04-111317-C	2	0.0414	-7470A	PPB	15 Nov 2017 11:53:03	U	SMPL	
I6305-04 - 1	PL-01-111317-A	47	0.0542	-7470A	PPB	15 Nov 2017 11:55:09	U	SMPL	
CCV09 - 1	CCV09	17973	5.1303	-7470A	PPB	15 Nov 2017 11:57:16	U	SMPL	
CCB09 - 1	CCB09	-285	-0.0398	-7470A	PPB	15 Nov 2017 11:59:22	U	SMPL	
I6305-06 - 1	PL-01-111317-B	25	0.0479	-7470A	PPB	15 Nov 2017 12:01:30	U	SMPL	
I6305-08 - 1	PL-01-111317-C	52	0.0556	-7470A	PPB	15 Nov 2017 12:03:36	U	SMPL	
I6305-10 - 1	PL-01-111317-D	25	0.0479	-7470A	PPB	15 Nov 2017 12:05:43	U	SMPL	
I6305-10DUP - 1	PL-01-111317-DDUP	27	0.0485	-7470A	PPB	15 Nov 2017 12:07:50	U	SMPL	
I6305-10MS - 1	PL-01-111317-DMS	13080	3.7448	-7470A	PPB	15 Nov 2017 12:09:57	U	SMPL	
I6305-10MSD - 1	PL-01-111317-DMSD	13050	3.7363	-7470A	PPB	15 Nov 2017 12:12:05	U	SMPL	
I6469-02 - 1	SV30666L	-103	0.0117	-7470A	PPB	15 Nov 2017 12:14:13	U	SMPL	
PB104261BL - 1	PBW	-19	0.0355	-7470A	PPB	15 Nov 2017 12:16:21	U	SMPL	
PB104261BS - 1	LCSW	12732	3.6462	-7470A	PPB	15 Nov 2017 12:18:30	U	SMPL	
I6469-01 - 1	SV30666L	233647	66.2032	-7470A	PPB	15 Nov 2017 12:20:37	U	SMPL	
CCV10 - 1	CCV10	16462	4.7024	-7470A	PPB	15 Nov 2017 12:22:44	U	SMPL	
CCB10 - 1	CCB10	-424	-0.0792	-7470A	PPB	15 Nov 2017 12:25:09	U	SMPL	
I6469-01DUP - 1	SV30666LDUP	235423	66.7061	-7470A	PPB	15 Nov 2017 12:27:18	U	SMPL	
I6469-01MS - 1	SV30666LMS	235883	66.8364	-7470A	PPB	15 Nov 2017 12:34:11	U	SMPL	
I6469-01MSD - 1	SV30666LMSD	240095	68.0291	-7470A	PPB	15 Nov 2017 12:36:18	U	SMPL	
I6305-10LX5 - 1		-269	-0.0353	-7470A	PPB	15 Nov 2017 12:41:29	U	SMPL	
I6305-10A - 1		13783	3.9438	-7470A	PPB	15 Nov 2017 12:43:36	U	SMPL	
I6469-01LX5 - 1		44331	12.5942	-7470A	PPB	15 Nov 2017 12:45:44	U	SMPL	
I6469-01A - 1		233153	66.0633	-7470A	PPB	15 Nov 2017 12:47:52	U	SMPL	
CCV11 - 1	CCV11	17009	4.8573	-7470A	PPB	15 Nov 2017 12:50:02	U	SMPL	
CCB11 - 1	CCB11	-396	-0.0713	-7470A	PPB	15 Nov 2017 12:52:31	U	SMPL	
I6469-01DLX10 - 1		22467	6.4029	-7470A	PPB	15 Nov 2017 12:54:39	U	SMPL	
I6469-01DUPDLX10 - 1		22833	6.5065	-7470A	PPB	15 Nov 2017 12:56:46	U	SMPL	
I6469-01MSDLX10 - 1		22890	6.5227	-7470A	PPB	15 Nov 2017 12:58:55	U	SMPL	
I6469-01MSDDLX10 - 1		23842	6.7923	-7470A	PPB	15 Nov 2017 13:01:04	U	SMPL	
I6469-01LDLX50 - 1		4070	1.1934	-7470A	PPB	15 Nov 2017 13:03:12	U	SMPL	
I6469-01ADLX10 - 1		24836	7.0737	-7470A	PPB	15 Nov 2017 13:08:27	U	SMPL	
CCV12 - 1	CCV12	17823	5.0878	-7470A	PPB	15 Nov 2017 13:10:33	U	SMPL	
CCB12 - 1	CCB12	-305	-0.0455	-7470A	PPB	15 Nov 2017 13:12:43	U	SMPL	