



Analytical Summary Report

Analysis Method: SM 4500-PH B

Analyst By : JIGNESH

Parameter: pH

Supervisor Review By : apatel

Run Number: LB91615

Temp.: 16.8 °C

SLOPE.: 97.6

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2270
buffer solution pH 7 yellow	W2159
PH 10.01 BUFFER,COLOR CD 475ML	W2154
buffer solution pH 7 yellow	W2219
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.01 Control Limits[+/- 0.1].

True Value of CCV2 = 12.01 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			4.01	11/15/2017	12:35
2	CAL2	1	Water			6.99	11/15/2017	12:36
3	CAL3	1	Water			10.01	11/15/2017	12:37
4	ICV	1	Water			7.00	11/15/2017	12:38
5	CCV1	1	Water			2.01	11/15/2017	12:39
6	I6493-01	1	Water			5.77	11/15/2017	12:40
7	I6493-01DUP	1	Water			5.78	11/15/2017	12:41
8	CCV2	1	Water			12.01	11/15/2017	12:42

1891615

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH 16493

WorkList ID : 105534

Date : 11/15/2017 12:21:50 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
11/14/2017	Water	I6493-01-C	pH	Cool 4 deg C	STER01	J52	POOLWATER	11/14/2017	SM 4500-PH B

Date/Time 11-15-17 12:30 PM
Received by: JP
Relinquished by: CP

Date/Time 11-15-17 2:20 PM
Received by: CP
Relinquished by: JP