

245.1


LB 91622

INSTRUMENT ID: CV1

Linear

 μ Abs.:

33255



A= 0.0000e+000

B= 2.9730e-004

C= 1.8546e-002

Rho= 0.9998291

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	-0.039	-0.039	-194	0.000	-194				
0.2	0.200	0.175	-0.025	527	0.0 %	527				
2.5	2.500	2.492	-0.008	8321	0.0 %	8321				
5.0	5.000	5.121	0.121	17164	0.0 %	17164				
7.5	7.500	7.545	0.045	25315	0.0 %	25315				
10.0	10.000	9.905	-0.095	33255	0.0 %	33255				

LB91622 INSTRUMENT ID : CV1

Method: 245.1

Operator: Admin

Date of Analysis: 15 Nov 2017 17:53:46

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	-194	-	0.0000	245.1	PPB	15 Nov 2017 17:55:17	S	Std
0.2 - 1	50.2	527	-	0.2000	245.1	PPB	15 Nov 2017 17:57:25	S	Std
2.5 - 1	52.2	8321	-	2.5000	245.1	PPB	15 Nov 2017 17:59:33	S	Std
5.0 - 1	55	17164	-	5.0000	245.1	PPB	15 Nov 2017 18:01:41	S	Std
7.5 - 1	57.5	25315	-	7.5000	245.1	PPB	15 Nov 2017 18:03:49	S	Std
10.0 - 1	510	33255	-	10.0000	245.1	PPB	15 Nov 2017 18:05:57	S	Std
ICV65 - 1	ICV65	13118	3.9186	-245.1		PPB	15 Nov 2017 18:08:38	U	SMPL
ICB65 - 1	ICB65	-311	-0.0739	-245.1		PPB	15 Nov 2017 18:10:45	U	SMPL
CCV17 - 1	CCV17	17200	5.1321	-245.1		PPB	15 Nov 2017 18:12:52	U	SMPL
CCB17 - 1	CCB17	-322	-0.0772	-245.1		PPB	15 Nov 2017 18:14:58	U	SMPL
CRA - 1	CRA	489	0.1639	-245.1		PPB	15 Nov 2017 18:17:06	U	SMPL
HighStd - 1	HighStd	33657	10.0248	-245.1		PPB	15 Nov 2017 18:19:12	U	SMPL
ChkStd - 1	ChkStd	22685	6.7628	-245.1		PPB	15 Nov 2017 18:21:19	U	SMPL
PB104263BL - 1	PBW	-318	-0.0760	-245.1		PPB	15 Nov 2017 18:23:27	U	SMPL
PB104263BS - 1	LCSW	13022	3.8900	-245.1		PPB	15 Nov 2017 18:25:36	U	SMPL
I6465-01 - 1	111017-TDO-E-D-M	-12	0.0150	-245.1		PPB	15 Nov 2017 18:27:44	U	SMPL
I6465-02 - 1	111017-TDO-E-U-M	50	0.0334	-245.1		PPB	15 Nov 2017 18:29:52	U	SMPL
I6465-02DUP - 1	111017-TDO-E-U-MDUP	-133	-0.0210	-245.1		PPB	15 Nov 2017 18:32:01	U	SMPL
I6465-02MS - 1	111017-TDO-E-U-MMS	15462	4.6154	-245.1		PPB	15 Nov 2017 18:34:09	U	SMPL
I6465-02MSD - 1	111017-TDO-E-U-MMSD	15774	4.7082	-245.1		PPB	15 Nov 2017 18:36:16	U	SMPL
CCV18 - 1	CCV18	17130	5.1113	-245.1		PPB	15 Nov 2017 18:38:22	U	SMPL
CCB18 - 1	CCB18	-283	-0.0656	-245.1		PPB	15 Nov 2017 18:40:29	U	SMPL
I6465-02LX5 - 1		-120	-0.0171	-245.1		PPB	15 Nov 2017 18:42:37	U	SMPL
I6465-02A - 1		14607	4.3612	-245.1		PPB	15 Nov 2017 18:44:44	U	SMPL
CCV19 - 1	CCV19	17197	5.1312	-245.1		PPB	15 Nov 2017 18:46:50	U	SMPL
CCB19 - 1	CCB19	-281	-0.0650	-245.1		PPB	15 Nov 2017 18:48:57	U	SMPL