


7470A

L391626
INSTRUMENT ID: CV2

Linear

 μ Abs.:

35025



A= 0.0000e+000

B= 2.8344e-004

C= 6.4776e-002

Rho= 0.9999237

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	-0.011	-0.011	-267	0.000	-267				
0.2	0.200	0.178	-0.022	400	0.0 %	400				
2.5	2.500	2.499	-0.001	8590	0.0 %	8590				
5.0	5.000	5.095	0.095	17746	0.0 %	17746				
7.5	7.500	7.447	-0.053	26044	0.0 %	26044				
10.0	10.000	9.992	-0.008	35025	0.0 %	35025				

LB91626 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 16 Nov 2017 11:11:50

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	-267	-	0.0000	7470A	PPB	16 Nov 2017 11:32:36	S	Std
0.2 - 1	50.2	400	-	0.2000	7470A	PPB	16 Nov 2017 11:34:45	S	Std
2.5 - 1	52.5	8590	-	2.5000	7470A	PPB	16 Nov 2017 11:36:54	S	Std
5.0 - 1	55	17746	-	5.0000	7470A	PPB	16 Nov 2017 11:39:02	S	Std
7.5 - 1	57.5	26044	-	7.5000	7470A	PPB	16 Nov 2017 11:41:09	S	Std
10.0 - 1	510	35025	-	10.0000	7470A	PPB	16 Nov 2017 11:43:17	S	Std
ICV66 - 1	ICV66	13223	3.8126	-7470A	PPB	16 Nov 2017 11:45:53	U	SMPL	
ICB66 - 1	ICB66	-411	-0.0517	-7470A	PPB	16 Nov 2017 11:48:00	U	SMPL	
CCV20 - 1	CCV20	18097	5.1941	-7470A	PPB	16 Nov 2017 11:50:06	U	SMPL	
CCB20 - 1	CCB20	-423	-0.0551	-7470A	PPB	16 Nov 2017 11:52:13	U	SMPL	
CRA - 1	CRA	386	0.1742	-7470A	PPB	16 Nov 2017 11:54:20	U	SMPL	
HighStd - 1	HighStd	35654	10.1704	-7470A	PPB	16 Nov 2017 11:56:27	U	SMPL	
ChkStd - 1	ChkStd	24742	7.0775	-7470A	PPB	16 Nov 2017 11:58:34	U	SMPL	
PB104304BL - 1	PBW	-445	-0.0614	-7470A	PPB	16 Nov 2017 12:00:42	U	SMPL	
PB104304BS - 1	LCSW	13126	3.7851	-7470A	PPB	16 Nov 2017 12:02:51	U	SMPL	
PB104304TB - 1	PB104274TB	-373	-0.0409	-7470A	PPB	16 Nov 2017 12:04:59	U	SMPL	
I6289-06 - 1	SU-01-111417	-315	-0.0245	-7470A	PPB	16 Nov 2017 12:07:07	U	SMPL	
I6289-06DUP - 1	SU-01-111417DUP	-291	-0.0177	-7470A	PPB	16 Nov 2017 12:09:15	U	SMPL	
I6289-06MS - 1	SU-01-111417MS	12233	3.5320	-7470A	PPB	16 Nov 2017 12:11:23	U	SMPL	
I6289-06MSD - 1	SU-01-111417MSD	12949	3.7350	-7470A	PPB	16 Nov 2017 12:13:30	U	SMPL	
CCV21 - 1	CCV21	18021	5.1726	-7470A	PPB	16 Nov 2017 12:15:37	U	SMPL	
CCB21 - 1	CCB21	-427	-0.0563	-7470A	PPB	16 Nov 2017 12:17:43	U	SMPL	
I6476-02 - 1	72-11934-290	-288	-0.0169	-7470A	PPB	16 Nov 2017 12:19:51	U	SMPL	
I6476-04 - 1	SP-1-RIGHT	-281	-0.0149	-7470A	PPB	16 Nov 2017 12:21:58	U	SMPL	
I6476-06 - 1	SP-1-LEFT	-280	-0.0146	-7470A	PPB	16 Nov 2017 12:24:05	U	SMPL	
I6477-02 - 1	TOTE-1-2	-275	-0.0132	-7470A	PPB	16 Nov 2017 12:26:12	U	SMPL	
I6478-02 - 1	SP-1-A	-227	0.0004	-7470A	PPB	16 Nov 2017 12:28:19	U	SMPL	
I6478-04 - 1	SP-1-B	-274	-0.0129	-7470A	PPB	16 Nov 2017 12:30:27	U	SMPL	
I6478-06 - 1	SP-1-C	-280	-0.0146	-7470A	PPB	16 Nov 2017 12:32:35	U	SMPL	
I6478-08 - 1	SP-1-D	-285	-0.0160	-7470A	PPB	16 Nov 2017 12:34:43	U	SMPL	
I6479-02 - 1	T-20	-268	-0.0112	-7470A	PPB	16 Nov 2017 12:36:52	U	SMPL	
I6479-04 - 1	POT-HEAD	-258	-0.0084	-7470A	PPB	16 Nov 2017 12:38:59	U	SMPL	
CCV22 - 1	CCV22	17956	5.1541	-7470A	PPB	16 Nov 2017 12:41:06	U	SMPL	
CCB22 - 1	CCB22	-412	-0.0520	-7470A	PPB	16 Nov 2017 12:43:13	U	SMPL	
I6481-03 - 1	SCMUA-SLUDGE-CAKE	-123	0.0299	-7470A	PPB	16 Nov 2017 12:45:20	U	SMPL	
PB104305BL - 1	PBW	-300	-0.0203	-7470A	PPB	16 Nov 2017 12:47:28	U	SMPL	
PB104305BS - 1	LCSW	12632	3.6451	-7470A	PPB	16 Nov 2017 12:49:35	U	SMPL	
I6477-01 - 1	TOTE-1-2	-369	-0.0398	-7470A	PPB	16 Nov 2017 12:51:42	U	SMPL	
I6477-01DUP - 1	TOTE-1-2DUP	-281	-0.0149	-7470A	PPB	16 Nov 2017 12:53:49	U	SMPL	
I6477-01MS - 1	TOTE-1-2MS	12396	3.5782	-7470A	PPB	16 Nov 2017 12:55:56	U	SMPL	
I6477-01MSD - 1	TOTE-1-2MSD	12501	3.6080	-7470A	PPB	16 Nov 2017 12:58:04	U	SMPL	
CCV23 - 1	CCV23	17982	5.1615	-7470A	PPB	16 Nov 2017 13:00:13	U	SMPL	
CCB23 - 1	CCB23	-430	-0.0571	-7470A	PPB	16 Nov 2017 13:02:22	U	SMPL	
I6289-06LX5 - 1		-292	-0.0180	-7470A	PPB	16 Nov 2017 13:04:29	U	SMPL	
I6289-06A - 1		13332	3.8435	-7470A	PPB	16 Nov 2017 13:06:36	U	SMPL	
I6477-01LX5 - 1		-374	-0.0412	-7470A	PPB	16 Nov 2017 13:08:43	U	SMPL	
I6477-01A - 1		12938	3.7319	-7470A	PPB	16 Nov 2017 13:10:51	U	SMPL	
CCV24 - 1	CCV24	17996	5.1655	-7470A	PPB	16 Nov 2017 13:12:58	U	SMPL	
CCB24 - 1	CCB24	-417	-0.0534	-7470A	PPB	16 Nov 2017 13:15:06	U	SMPL	