

7471A


LB91638

INSTRUMENT ID: CV2

Linear

 μ Abs.:

44711



A= 0.0000e+000

B= 2.2328e-004

C= 6.4181e-002

Rho= 0.9999565

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.008	0.008	-253	0.000	-253				
0.2	0.200	0.207	0.007	641	0.0 %	641				
2.5	2.500	2.500	-0.000	10907	0.0 %	10907				
5.0	5.000	5.007	0.007	22136	0.0 %	22136				
7.5	7.500	7.431	-0.069	32995	0.0 %	32995	0			
10.0	10.000	10.047	0.047	44711	0.0 %	44711				

LB91638 INSTRUMENT ID : CV1

Method: 7471A

Operator: Admin

Date of Analysis: 16 Nov 2017 15:08:07

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	-253	-	0.00007471A	PPB		16 Nov 2017 15:11:01	S	Std
0.2 - 1	50.2	641	-	0.20007471A	PPB		16 Nov 2017 15:13:18	S	Std
2.5 - 1	52.5	10907	-	2.50007471A	PPB		16 Nov 2017 15:15:36	S	Std
5.0 - 1	55	22136	-	5.00007471A	PPB		16 Nov 2017 15:17:53	S	Std
7.5 - 1	57.5	32995	-	7.50007471A	PPB		16 Nov 2017 15:20:11	S	Std
10.0 - 1	510	44711	-	10.00007471A	PPB		16 Nov 2017 15:25:03	S	Std
ICV67 - 1	ICV67	16772	3.8091	-7471A	PPB		16 Nov 2017 15:28:15	U	SMPL
ICB67 - 1	ICB67	-409	-0.0271	-7471A	PPB		16 Nov 2017 15:30:31	U	SMPL
CCV25 - 1	CCV25	20893	4.7292	-7471A	PPB		16 Nov 2017 15:32:46	U	SMPL
CCB25 - 1	CCB25	-409	-0.0271	-7471A	PPB		16 Nov 2017 15:35:02	U	SMPL
CRA - 1	CRA	600	0.1982	-7471A	PPB		16 Nov 2017 15:37:18	U	SMPL
HighStd - 1	HighStd	45202	10.1570	-7471A	PPB		16 Nov 2017 15:39:34	U	SMPL
ChkStd - 1	ChkStd	29741	6.7048	-7471A	PPB		16 Nov 2017 15:41:50	U	SMPL
PB104296BL - 1	PBS	-396	-0.0242	-7471A	PPB		16 Nov 2017 15:44:09	U	SMPL
PB104296BS - 1	LCSS	14919	3.3953	-7471A	PPB		16 Nov 2017 15:46:27	U	SMPL
I6289-05 - 1	SU-01-111417	1517	0.4029	-7471A	PPB		16 Nov 2017 15:48:44	U	SMPL
I6466-01 - 1	CHAP-5-E(0-1)	3012	0.7367	-7471A	PPB		16 Nov 2017 15:51:01	U	SMPL
I6466-02 - 1	CHAP-5-E(4-5)	4257	1.0147	-7471A	PPB		16 Nov 2017 15:53:19	U	SMPL
I6466-03 - 1	CHAP-5-E(7-8)	-57	0.0515	-7471A	PPB		16 Nov 2017 15:55:36	U	SMPL
I6466-04 - 1	CHAP-4-E(0-1)	1910	0.4906	-7471A	PPB		16 Nov 2017 15:57:52	U	SMPL
CCV26 - 1	CCV26	21682	4.9054	-7471A	PPB		16 Nov 2017 16:00:08	U	SMPL
CCB26 - 1	CCB26	-424	-0.0305	-7471A	PPB		16 Nov 2017 16:02:24	U	SMPL
I6466-05 - 1	CHAP-4-E(3-4)	32006	7.2106	-7471A	PPB		16 Nov 2017 16:04:40	U	SMPL
I6466-06 - 1	CHAP-4-E(7-8)	-261	0.0059	-7471A	PPB		16 Nov 2017 16:06:56	U	SMPL
I6466-07 - 1	CHAP-2-E(0-1)	4597	1.0906	-7471A	PPB		16 Nov 2017 16:09:14	U	SMPL
I6466-08 - 1	CHAP-2-E(4-5)	5727	1.3429	-7471A	PPB		16 Nov 2017 16:11:30	U	SMPL
I6466-09 - 1	CHAP-2-E(6-7)	2132	0.5402	-7471A	PPB		16 Nov 2017 16:13:47	U	SMPL
I6476-01 - 1	72-11934-290	2806	0.6907	-7471A	PPB		16 Nov 2017 16:16:04	U	SMPL
I6476-03 - 1	SP-1-RIGHT	-65	0.0497	-7471A	PPB		16 Nov 2017 16:18:21	U	SMPL
I6476-05 - 1	SP-1-LEFT	148	0.0972	-7471A	PPB		16 Nov 2017 16:20:39	U	SMPL
I6476-05DUP - 1	SP-1-LEFTDUP	142	0.0959	-7471A	PPB		16 Nov 2017 16:25:31	U	SMPL
I6476-05MS - 1	SP-1-LEFTMS	16068	3.6519	-7471A	PPB		16 Nov 2017 16:27:47	U	SMPL
CCV27 - 1	CCV27	21570	4.8804	-7471A	PPB		16 Nov 2017 16:30:04	U	SMPL
CCB27 - 1	CCB27	-427	-0.0312	-7471A	PPB		16 Nov 2017 16:32:20	U	SMPL
I6476-05MSD - 1	SP-1-LEFTMSD	15964	3.6287	-7471A	PPB		16 Nov 2017 16:34:37	U	SMPL
I6478-01 - 1	SP-1-A	1883	0.4846	-7471A	PPB		16 Nov 2017 16:36:54	U	SMPL
I6478-03 - 1	SP-1-B	1007	0.2890	-7471A	PPB		16 Nov 2017 16:39:10	U	SMPL
I6478-05 - 1	SP-1-C	774	0.2370	-7471A	PPB		16 Nov 2017 16:41:27	U	SMPL
I6478-07 - 1	SP-1-D	1050	0.2986	-7471A	PPB		16 Nov 2017 16:43:44	U	SMPL
I6479-01 - 1	T-20	347	0.1417	-7471A	PPB		16 Nov 2017 16:46:01	U	SMPL
I6479-03 - 1	POT-HEAD	698	0.2200	-7471A	PPB		16 Nov 2017 16:48:18	U	SMPL
PB104297BL - 1	PBS	-268	0.0043	-7471A	PPB		16 Nov 2017 16:50:35	U	SMPL
PB104297BS - 1	LCSS	15303	3.4811	-7471A	PPB		16 Nov 2017 16:52:53	U	SMPL
I6309-01 - 1	BU-02-111517	2287	0.5748	-7471A	PPB		16 Nov 2017 16:55:09	U	SMPL
CCV28 - 1	CCV28	21609	4.8891	-7471A	PPB		16 Nov 2017 16:57:26	U	SMPL
CCB28 - 1	CCB28	-402	-0.0256	-7471A	PPB		16 Nov 2017 16:59:42	U	SMPL
I6481-01 - 1	SCMUA-SLUDGE-CAKE	3610	0.8702	-7471A	PPB		16 Nov 2017 17:01:59	U	SMPL
I6486-01 - 1	EPE-125-10(0-5)	559	0.1890	-7471A	PPB		16 Nov 2017 17:04:15	U	SMPL
I6486-02 - 1	EPE-125-10(5-10)	1708	0.4455	-7471A	PPB		16 Nov 2017 17:06:32	U	SMPL
I6486-03 - 1	EPE-L-9(0-5)	22049	4.9873	-7471A	PPB		16 Nov 2017 17:08:48	U	SMPL
I6486-03DUP - 1	EPE-L-9(0-5)DUP	22653	5.1222	-7471A	PPB		16 Nov 2017 17:13:41	U	SMPL
I6486-03MS - 1	EPE-L-9(0-5)MS	41486	9.3273	-7471A	PPB		16 Nov 2017 17:18:36	U	SMPL
I6486-03MSD - 1	EPE-L-9(0-5)MSD	39547	8.8943	-7471A	PPB		16 Nov 2017 17:20:53	U	SMPL
I6486-04 - 1	EPE-L-9(5-10)	6246	1.4588	-7471A	PPB		16 Nov 2017 17:23:13	U	SMPL
I6499-01 - 1	PG1-2-E(0-1)	90133	20.1893	-7471A	PPB		16 Nov 2017 17:25:33	U	SMPL
I6499-02 - 1	PG1-2-E(4-5)	148	0.0972	-7471A	PPB		16 Nov 2017 17:30:29	U	SMPL
CCV29 - 1	CCV29	21538	4.8732	-7471A	PPB		16 Nov 2017 17:32:45	U	SMPL
CCB29 - 1	CCB29	-408	-0.0269	-7471A	PPB		16 Nov 2017 17:35:02	U	SMPL
I6499-03 - 1	PG1-2-E(7-8)	1349	0.3654	-7471A	PPB		16 Nov 2017 17:37:18	U	SMPL
I6499-04 - 1	PG1-1-E(0-1)	41045	9.2288	-7471A	PPB		16 Nov 2017 17:39:35	U	SMPL
I6499-05 - 1	PG1-1-E(7-8)	583	0.1944	-7471A	PPB		16 Nov 2017 17:41:51	U	SMPL
I6499-06 - 1	PG1-4-E(0-1)	22499	5.0878	-7471A	PPB		16 Nov 2017 17:44:09	U	SMPL
I6499-07 - 1	PG1-4-E(7-8)	185366	41.4531	-7471A	PPB		16 Nov 2017 17:46:26	U	SMPL
I6499-08 - 1	PG1-3-E(4-5)	36086	8.1215	-7471A	PPB		16 Nov 2017 17:48:44	U	SMPL
I6499-09 - 1	PG1-3-E(7-8)	23480	5.3068	-7471A	PPB		16 Nov 2017 17:51:13	U	SMPL
I6503-01 - 1	EPE-L-9(21.3-23.3)	-68	0.0490	-7471A	PPB		16 Nov 2017 17:53:32	U	SMPL
I6503-02 - 1	EPE-109-3(0-5)	12452	2.8445	-7471A	PPB		16 Nov 2017 17:55:51	U	SMPL
CCV30 - 1	CCV30	21783	4.9279	-7471A	PPB		16 Nov 2017 17:58:07	U	SMPL
CCB30 - 1	CCB30	-415	-0.0285	-7471A	PPB		16 Nov 2017 18:00:24	U	SMPL
I6476-05LX5 - 1		-168	0.0267	-7471A	PPB		16 Nov 2017 18:02:42	U	SMPL
I6476-05A - 1		16637	3.7789	-7471A	PPB		16 Nov 2017 18:08:41	U	SMPL
I6486-03LX5 - 1		4034	0.9649	-7471A	PPB		16 Nov 2017 18:10:57	U	SMPL
I6486-03A - 1		44482	9.9962	-7471A	PPB		16 Nov 2017 18:13:14	U	SMPL
I6499-01DLX5 - 1		16658	3.7836	-7471A	PPB		16 Nov 2017 18:15:30	U	SMPL
I6499-07DLX10 - 1		18266	4.1427	-7471A	PPB		16 Nov 2017 18:17:50	U	SMPL
CCV31 - 1	CCV31	21518	4.8688	-7471A	PPB		16 Nov 2017 18:20:07	U	SMPL
CCB31 - 1	CCB31	-447	-0.0356	-7471A	PPB		16 Nov 2017 18:22:25	U	SMPL